

**LOT** PX 031

**CONTROL**

(Exp.) 2011-05-05  
 (YYYY - MM - DD)

Rev 1

		ABX Lyse																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA					MS60		PENTRA					MS60		PENTRA					MS60	
		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120				60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120				60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.5	2.5		± 0.4	7.6	7.5	7.9	7.9		± 1.0	17.4	17.5	18.3	18.3		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.45	2.40	2.45	2.45		± 0.12	4.70	4.70	4.80	4.80		± 0.15	5.18	5.20	5.30	5.30		± 0.20			
HB HGB	g/dl	6.3	6.2	6.3	6.3		± 0.4	14.0	13.9	14.0	14.0		± 0.5	17.0	16.9	16.9	16.9		± 0.6			
	g/l	63	62	63	63		± 4	140	139	140	140		± 5	170	169	169	169		± 6			
	mmol/l	3.91	3.85	3.91	3.91		± 0.25	8.69	8.63	8.69	8.69		± 0.31	10.56	10.49	10.49	10.49		± 0.37			
HT HCT	%	18.9	19.2	19.6	19.6		± 1.5	40.4	40.9	42.2	42.2		± 2.0	48.7	48.9	49.8	49.8		± 2.5			
	l/l	0.189	0.192	0.196	0.196		± 0.015	0.404	0.409	0.422	0.422		± 0.020	0.487	0.489	0.498	0.498		± 0.025			
VGM MCV	µm <sup>3</sup> /fl	77	80	80	80		± 4	86	87	88	88		± 4	94	94	94	94		± 4			
TGMH MCH	pg	25.7	25.8	25.7	25.7		± 2.0	29.8	29.6	29.2	29.2		± 2.0	32.8	32.5	31.9	31.9		± 2.5			
	fmol	1.60	1.60	1.60	1.60		± 0.12	1.85	1.84	1.81	1.81		± 0.12	2.04	2.02	1.98	1.98		± 0.16			
CCMH MCHC	g/dl	33.4	32.3	32.1	32.1		± 3.0	34.6	34.0	33.1	33.1		± 3.0	34.9	34.6	33.9	33.9		± 3.0			
	g/l	334	323	321	321		± 30	346	340	331	331		± 30	349	346	339	339		± 30			
	mmol/l	20.74	20.05	19.96	19.96		± 1.86	21.51	21.11	20.58	20.58		± 1.86	21.68	21.47	21.07	21.07		± 1.86			
IDR RDW	%	14.2	13.8	16.0	17.0		± 4.0	12.0	12.5	14.0	14.4		± 4.0	11.3	11.6	13.6	14.2		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	68	65	65	65		± 20	262	265	265	265		± 30	490	492	497	497		± 50			
VPM MPV	µm <sup>3</sup> /fl	9.1	9.3	8.9	8.9		± 2.0	8.3	8.6	8.5	8.5		± 2.0	8.0	8.2	8.2	8.2		± 2.0			
NEUT	#	1.39	1.39	1.45	1.48		± 0.30	4.64	4.50	4.94	4.90		± 0.80	13.17	13.13	14.40	14.10		± 1.90			
	%	57.8	57.8	57.8	59.0		± 10.0	61.0	60.0	62.5	62.0		± 10.0	75.7	75.0	78.5	77.0		± 10.0			
LYMPHO	#	0.74	0.74	0.76	0.73		± 0.20	2.32	2.36	2.33	2.29		± 0.70	2.44	2.54	2.47	2.47		± 1.50			
	%	30.7	31.0	30.5	29.0		± 8.0	30.5	31.5	29.5	29.0		± 8.0	14.0	14.5	13.5	13.5		± 8.0			
MONO	#	0.04	0.04	0.04	0.04		± 0.04	0.12	0.12	0.12	0.12		± 0.12	0.37	0.37	0.37	0.37		± 0.37			
	%	1.5	1.5	1.5	1.5		± 1.5	1.5	1.5	1.5	1.5		± 1.5	2.0	2.0	2.0	2.0		± 2.0			
EOS	#	0.16	0.16	0.18	0.19		± 0.16	0.30	0.30	0.28	0.36		± 0.28	0.73	0.79	0.73	0.92		± 0.73			
	%	6.8	6.5	7.2	7.5		± 6.0	4.0	4.0	3.5	4.5		± 3.5	4.0	4.5	4.0	5.0		± 4.0			
BASO	#	0.08	0.08	0.08	0.08		± 0.08	0.24	0.24	0.24	0.24		± 0.24	0.75	0.70	0.37	0.46		± 0.37			
	%	3.2	3.2	3.0	3.0		± 3.0	3.0	3.0	3.0	3.0		± 3.0	4.3	4.0	2.0	2.5		± 2.0			

Ref: FORM-0821 Rev.11 FRONT / RECTO 9930062-A