

**HORIBA ABX SAS**  
 Parc Euromédecine  
 34184 Montpellier cedex 4 - France  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** PX 011  
 Rev 2

**CONTROL**

(Exp.) 2011-03-05  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse																	
			CONTROL				L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA				MS60	TOLERANCES TOLERANCE	PENTRA				MS60	TOLERANCES TOLERANCE	PENTRA				MS60	
			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120		
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.3	2.3	2.4	2.4		± 0.4	7.7	7.8	8.1	8.1		± 1.0	17.7	17.7	18.6	18.6		± 2.2
GR	RBC	10 <sup>12</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.47	2.42	2.48	2.48		± 0.12	4.60	4.60	4.67	4.67		± 0.15	5.12	5.15	5.26	5.26		± 0.20
HB	HGB	g/dl	6.3	6.3	6.3	6.3		± 0.4	13.8	13.8	13.8	13.8		± 0.5	16.8	16.8	16.7	16.7		± 0.6
		g/l	63	63	63	63		± 4	138	138	138	138		± 5	168	168	167	167		± 6
		mmol/l	3.91	3.91	3.91	3.91		± 0.25	8.57	8.57	8.57	8.57		± 0.31	10.43	10.43	10.37	10.37		± 0.37
HT	HCT	%	18.5	18.9	19.3	19.3		± 1.5	40.0	40.5	41.1	41.1		± 2.0	48.1	48.4	48.9	48.9		± 2.5
		l/l	0.185	0.189	0.193	0.193		± 0.015	0.400	0.405	0.411	0.411		± 0.020	0.481	0.484	0.489	0.489		± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	75	78	78	78		± 4	87	88	88	88		± 4	94	94	93	93		± 4
TGMH	MCH	pg	25.5	26.0	25.4	25.4		± 2.0	30.0	30.0	29.5	29.5		± 2.0	32.8	32.6	31.7	31.7		± 2.5
		fmol	1.58	1.62	1.58	1.58		± 0.12	1.86	1.86	1.83	1.83		± 0.12	2.04	2.03	1.97	1.97		± 0.16
CCMH	MCHC	g/dl	34.0	33.4	32.6	32.6		± 3.0	34.5	34.1	33.6	33.6		± 3.0	34.9	34.7	34.1	34.1		± 3.0
		g/l	340	334	326	326		± 30	345	341	336	336		± 30	349	347	341	341		± 30
		mmol/l	21.12	20.73	20.22	20.22		± 1.86	21.41	21.17	20.85	20.85		± 1.86	21.68	21.55	21.20	21.20		± 1.86
IDR	RDW	%	13.8	13.8	16.0	16.0		± 4.0	11.9	12.2	13.9	13.9		± 4.0	11.3	11.7	13.5	13.5		± 4.0
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	70	70	70		± 20	250	248	252	252		± 30	480	475	485	485		± 50
VPM	MPV	µm <sup>3</sup> ; fl	9.2	9.3	9.1	9.1		± 2.0	9.0	9.4	9.6	9.6		± 2.0	8.5	8.8	8.9	8.9		± 2.0
NEUT	#		1.35	1.34	1.40	1.40		± 0.30	4.81	4.91	5.22	5.22		± 0.80	13.36	13.29	14.70	14.70		± 1.80
		%	58.5	58.4	58.2	58.2		± 10.0	62.5	63.0	64.5	64.5		± 10.0	75.5	75.1	79.0	79.0		± 10.0
LYMPHO	#		0.67	0.66	0.71	0.71		± 0.30	2.19	2.18	2.19	2.19		± 0.70	2.21	2.30	2.23	2.23		± 1.50
		%	29.0	28.5	29.5	29.5		± 8.0	28.5	28.0	27.0	27.0		± 8.0	12.5	13.0	12.0	12.0		± 8.0
MONO	#		0.04	0.04	0.04	0.04		± 0.04	0.16	0.16	0.16	0.16		± 0.16	0.37	0.37	0.37	0.37		± 0.37
		%	1.8	1.8	1.8	1.8		± 1.8	2.0	2.0	2.0	2.0		± 2.0	2.0	2.0	2.0	2.0		± 2.0
EOS	#		0.18	0.18	0.18	0.18		± 0.17	0.28	0.28	0.28	0.28		± 0.28	0.97	0.97	0.84	0.84		± 0.84
		%	7.5	8.0	7.5	7.5		± 7.0	3.5	3.5	3.5	3.5		± 3.5	5.5	5.5	4.5	4.5		± 4.5
BASO	#		0.07	0.08	0.07	0.07		± 0.07	0.27	0.27	0.24	0.24		± 0.24	0.80	0.78	0.47	0.47		± 0.47
		%	3.2	3.3	3.0	3.0		± 3.0	3.5	3.5	3.0	3.0		± 3.0	4.5	4.4	2.5	2.5		± 2.5

Ref: FORM-0821 Rev.11 FRONT / RECTO 9930052-A