

**LOT** PX 110

**CONTROL**

(Exp.) 2011-01-05  
 (YYYY - MM - DD)

Rev 1

		ABX Lyse																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA							PENTRA							PENTRA						
		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4	2.4		± 0.4	7.5	7.4	7.7	7.7		± 1.0	17.7	17.6	18.2	18.2		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.33	2.39	2.39		± 0.12	4.66	4.64	4.72	4.72		± 0.15	5.14	5.14	5.25	5.25		± 0.20			
HB HGB	g/dl	6.2	6.2	6.2	6.2		± 0.4	13.9	13.9	13.9	13.9		± 0.5	16.7	16.6	16.6	16.6		± 0.6			
	g/l	62	62	62	62		± 4	139	139	139	139		± 5	167	166	166	166		± 6			
	mmol/l	3.85	3.85	3.85	3.85		± 0.25	8.63	8.63	8.63	8.63		± 0.31	10.37	10.31	10.31	10.31		± 0.37			
HT HCT	%	18.3	18.6	19.4	19.4		± 1.5	40.1	40.4	41.5	41.5		± 2.0	47.8	47.8	48.8	48.8		± 2.5			
	l/l	0.183	0.186	0.194	0.194		± 0.015	0.401	0.404	0.415	0.415		± 0.020	0.478	0.478	0.488	0.488		± 0.025			
VGM MCV	µm <sup>3</sup> /fl	77	80	81	81		± 4	86	87	88	88		± 4	93	93	93	93		± 4			
TGMH MCH	pg	26.1	26.6	25.9	25.9		± 2.0	29.8	30.0	29.4	29.4		± 2.0	32.5	32.3	31.6	31.6		± 2.5			
	fmol	1.62	1.65	1.61	1.61		± 0.12	1.85	1.86	1.83	1.83		± 0.12	2.02	2.01	1.96	1.96		± 0.16			
CCMH MCHC	g/dl	33.8	33.3	32.0	32.0		± 3.0	34.7	34.4	33.5	33.5		± 3.0	34.9	34.7	34.0	34.0		± 3.0			
	g/l	338	333	320	320		± 30	347	344	335	335		± 30	349	347	340	340		± 30			
	mmol/l	21.01	20.66	19.89	19.89		± 1.86	21.54	21.38	20.78	20.78		± 1.86	21.70	21.57	21.11	21.11		± 1.86			
IDR RDW	%	14.3	14.1	15.4	15.4		± 4.0	12.4	12.6	13.4	13.4		± 4.0	12.0	12.4	13.2	13.2		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	68	70	70		± 20	245	240	250	250		± 30	485	485	485	485		± 50			
VPM MPV	µm <sup>3</sup> /fl	8.9	8.9	8.8	8.8		± 2.0	8.9	9.1	9.1	9.1		± 2.0	8.3	8.6	8.6	8.6		± 2.0			
NEUT	#	1.41	1.44	1.43	1.43		± 0.30	4.61	4.55	4.81	4.81		± 0.80	13.51	13.46	14.40	14.40		± 1.90			
	%	58.6	60.2	59.4	59.4		± 10.0	61.5	61.5	62.5	62.5		± 10.0	76.3	76.5	79.0	79.0		± 10.0			
LYMPHO	#	0.69	0.65	0.67	0.67		± 0.20	2.21	2.22	2.27	2.27		± 0.60	2.30	2.29	2.18	2.18		± 1.50			
	%	28.8	27.0	28.0	28.0		± 8.0	29.5	30.0	29.5	29.5		± 8.0	13.0	13.0	12.0	12.0		± 8.0			
MONO	#	0.04	0.04	0.04	0.04		± 0.04	0.11	0.11	0.11	0.11		± 0.11	0.27	0.27	0.27	0.27		± 0.27			
	%	1.8	1.8	1.8	1.8		± 1.8	1.5	1.5	1.5	1.5		± 1.5	1.5	1.5	1.5	1.5		± 1.5			
EOS	#	0.18	0.18	0.19	0.19		± 0.18	0.30	0.27	0.27	0.27		± 0.27	0.89	0.88	0.82	0.82		± 0.82			
	%	7.5	7.5	7.8	7.8		± 7.0	4.0	3.5	3.5	3.5		± 3.5	5.0	5.0	4.5	4.5		± 4.5			
BASO	#	0.08	0.08	0.07	0.07		± 0.07	0.26	0.26	0.23	0.23		± 0.23	0.74	0.70	0.55	0.55		± 0.55			
	%	3.3	3.5	3.0	3.0		± 3.0	3.5	3.5	3.0	3.0		± 3.0	4.2	4.0	3.0	3.0		± 3.0			

Ref: FORM-0821 Rev.11 FRONT / RECTO 9930062-A