


LOT PX 412
Rev 1


CONTROL

 (Exp.) 2018-09-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.3	7.3			± 1.0	15.9	15.9			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.27	2.27			± 0.16	4.60	4.60			± 0.20	5.04	5.04			± 0.25			
HB HGB	g/dl	6.6	6.6			± 0.4	13.9	13.9			± 0.5	15.8	15.8			± 0.6			
	g/l	66	66			± 4	139	139			± 5	158	158			± 6			
	mmol/l	4.10	4.10			± 0.25	8.63	8.63			± 0.31	9.81	9.81			± 0.37			
HT HCT	%	18.8	18.8			± 1.5	38.2	38.2			± 2.0	43.8	43.3			± 2.5			
	l/l	0.188	0.188			± 0.015	0.382	0.382			± 0.020	0.438	0.433			± 0.025			
VGM MCV	µm ³ ; fl	83	83			± 5	83	83			± 5	87	86			± 5			
TGMH MCH	pg	29.1	29.1			± 2.0	30.2	30.2			± 2.0	31.3	31.3			± 2.5			
	fmol	1.81	1.81			± 0.12	1.88	1.88			± 0.12	1.95	1.95			± 0.16			
CCMH MCHC	g/dl	35.0	35.0			± 3.0	36.4	36.4			± 3.0	36.0	36.5			± 3.0			
	g/l	350	350			± 30	364	364			± 30	360	365			± 30			
	mmol/l	21.75	21.75			± 1.86	22.61	22.61			± 1.86	22.38	22.64			± 1.86			
IDR RDW	%	12.8	11.7			± 4.0	13.8	12.5			± 4.0	13.8	12.7			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	66	66			± 20	243	243			± 30	480	480			± 50			
VPM MPV	µm ³ ; fl	9.7	9.7			± 2.0	9.7	10.0			± 2.0	9.4	9.7			± 2.0			
NEUT	#	1.42	1.42			± 0.35	4.15	4.15			± 0.90	10.80	10.80			± 1.90			
	%	56.8	56.8			± 10.0	56.9	56.9			± 10.0	67.9	67.9			± 10.0			
LYMPHO	#	0.75	0.75			± 0.33	2.39	2.39			± 0.70	2.91	2.91			± 1.50			
	%	30.1	30.1			± 12.0	32.7	32.7			± 8.0	18.3	18.3			± 8.0			
MONO	#	0.09	0.09			± 0.09	0.26	0.26			± 0.26	0.59	0.59			± 0.59			
	%	3.4	3.4			± 3.4	3.5	3.5			± 3.5	3.7	3.7			± 3.7			
EOS	#	0.16	0.16			± 0.16	0.24	0.24			± 0.24	0.92	0.92			± 0.92			
	%	6.4	6.4			± 6.4	3.3	3.3			± 3.3	5.8	5.8			± 5.8			
BASO	#	0.08	0.08			± 0.08	0.26	0.26			± 0.26	0.68	0.68			± 0.68			
	%	3.3	3.3			± 3.3	3.6	3.6			± 3.6	4.3	4.3			± 4.3			

LOT PX 412
Rev 1

CONTROL

 (Exp.) 2018-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE						
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5			± 0.4	7.3	7.3	7.4			± 1.0	16.0	15.9	16.4			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.30	2.30			± 0.16	4.60	4.60	4.60			± 0.20	5.04	5.02	5.06			± 0.25
		g/dl	6.7	6.6	6.7			± 0.4	14.0	13.8	13.9			± 0.5	15.8	15.7	15.8			± 0.6
HB	HGB	g/l	67	66	67			± 4	140	138	139			± 5	158	157	158			± 6
		mmol/l	4.16	4.10	4.16			± 0.25	8.69	8.57	8.63			± 0.31	9.81	9.75	9.81			± 0.37
HT	HCT	%	18.9	18.4	18.6			± 1.5	38.2	37.7	37.7			± 2.0	43.8	43.2	43.0			± 2.5
		l/l	0.189	0.184	0.186			± 0.015	0.382	0.377	0.377			± 0.020	0.438	0.432	0.430			± 0.025
VGM	MCV	µm ³ ; fl	81	80	81.0			± 5	83	82	82.0			± 5	87	86	85.0			± 5
TGMH	MCH	pg	28.8	28.7	29.1			± 2.0	30.4	30.0	30.2			± 2.0	31.3	31.3	31.2			± 2.5
		fmol	1.79	1.78	1.81			± 0.12	1.89	1.86	1.88			± 0.12	1.95	1.94	1.94			± 0.16
		g/dl	35.5	35.9	36.0			± 3.0	36.7	36.6	36.9			± 3.0	36.0	36.4	36.7			± 3.0
CCMH	MCHC	g/l	355	359	360			± 30	367	366	369			± 30	360	364	367			± 30
		mmol/l	22.05	22.28	22.33			± 1.86	22.77	22.72	22.88			± 1.86	22.38	22.58	22.81			± 1.86
IDR	RDW	%	12.5	12.2	12.4			± 4.0	13.0	13.0	12.0			± 4.0	13.0	12.5	12.0			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70	68			± 20	250	250	240			± 30	490	490	475			± 50
VPM	MPV	µm ³ ; fl	9.5	9.4	9.2			± 2.0	9.4	9.4	9.1			± 2.0	9.2	9.2	8.8			± 2.0
		#	1.39	1.37	1.47			± 0.35	4.20	4.13	4.29			± 0.90	11.09	10.84	11.48			± 1.90
		%	58.0	57.2	58.6			± 10.0	57.5	56.6	58.0			± 10.0	69.3	68.2	70.0			± 10.0
		#	0.70	0.71	0.73			± 0.33	2.35	2.37	2.41			± 0.70	2.74	2.81	2.90			± 1.50
		%	29.1	29.6	29.2			± 12.0	32.2	32.5	32.5			± 8.0	17.1	17.7	17.7			± 8.0
		#	0.08	0.09	0.08			± 0.08	0.23	0.28	0.22			± 0.22	0.56	0.65	0.49			± 0.49
		%	3.2	3.7	3.0			± 3.0	3.2	3.9	3.0			± 3.0	3.5	4.1	3.0			± 3.0
		#	0.15	0.15	0.15			± 0.15	0.26	0.26	0.22			± 0.22	0.93	0.92	0.87			± 0.87
		%	6.2	6.2	5.8			± 5.8	3.5	3.5	3.0			± 3.0	5.8	5.8	5.3			± 5.3
		#	0.08	0.08	0.09			± 0.08	0.26	0.26	0.26			± 0.26	0.69	0.67	0.66			± 0.66
		%	3.5	3.3	3.4			± 3.3	3.6	3.5	3.5			± 3.5	4.3	4.2	4.0			± 4.0

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.42