


**LOT** PX 402  
Rev 2


**CONTROL**

 (Exp.) 2017-01-05  
(YYYY - MM - DD)

		<b>ABX Lysebio</b>																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA							PENTRA							PENTRA						
		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4	2.4	2.5	± 0.4	7.6	7.6	7.6	7.6	7.6	± 1.0	18.0	17.7	17.9	17.7	18.2	± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.32	2.35	2.32	2.35	± 0.16	4.54	4.54	4.54	4.54	4.52	± 0.20	5.05	5.12	5.10	5.12	5.10	± 0.25			
HB HGB	g/dl	6.8	6.8	6.8	6.8	6.8	± 0.4	13.5	13.6	13.5	13.6	13.4	± 0.5	16.4	16.5	16.3	16.5	16.3	± 0.6			
	g/l	68	68	68	68	68	± 4	135	136	135	136	134	± 5	164	165	163	165	163	± 6			
	mmol/l	4.22	4.22	4.22	4.22	4.22	± 0.25	8.38	8.45	8.38	8.45	8.32	± 0.31	10.18	10.25	10.12	10.25	10.12	± 0.37			
HT HCT	%	19.0	19.3	19.0	19.3	18.9	± 1.5	36.8	36.8	36.3	36.8	36.6	± 2.0	44.4	45.1	43.9	45.1	44.6	± 2.5			
	l/l	0.190	0.193	0.190	0.193	0.189	± 0.015	0.368	0.368	0.363	0.368	0.366	± 0.020	0.444	0.451	0.439	0.451	0.446	± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	81	83	81	83	80.5	± 5	81	81	80	81	81.0	± 5	88	88	86	88	87.5	± 5			
TGMH MCH	pg	28.9	29.3	28.9	29.3	28.9	± 2.0	29.7	30.0	29.7	30.0	29.6	± 2.0	32.5	32.2	32.0	32.2	32.0	± 2.5			
	fmol	1.80	1.82	1.80	1.82	1.80	± 0.12	1.85	1.86	1.85	1.86	1.84	± 0.12	2.02	2.00	1.98	2.00	1.98	± 0.16			
CCMH MCHC	g/dl	35.7	35.3	35.7	35.3	35.9	± 3.0	36.7	37.0	37.2	37.0	36.6	± 3.0	36.9	36.6	37.2	36.6	36.5	± 3.0			
	g/l	357	353	357	353	359	± 30	367	370	372	370	366	± 30	369	366	372	366	365	± 30			
	mmol/l	22.18	21.93	22.18	21.93	22.32	± 1.86	22.80	22.97	23.08	22.97	22.73	± 1.86	22.92	22.74	23.08	22.74	22.68	± 1.86			
IDR RDW	%	13.4	13.2	12.7	13.2	12.0	± 4.0	13.4	13.8	12.6	13.8	13.0	± 4.0	12.7	13.3	12.0	13.3	12.4	± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	68	68	68	70	± 20	248	243	248	243	241	± 30	490	495	510	495	485	± 50			
VPM MPV	µm <sup>3</sup> ; fl	8.9	9.2	9.2	9.2	8.7	± 2.0	9.2	9.4	9.2	9.4	8.8	± 2.0	8.8	9.0	8.9	9.0	8.4	± 2.0			
NEUT	#	1.38	1.41	1.38	1.41	1.47	± 0.35	4.40	4.45	4.37	4.45	4.45	± 0.90	12.96	12.96	13.07	12.96	13.41	± 1.90			
	%	57.5	58.7	57.5	58.7	58.9	± 10.0	57.9	58.5	57.5	58.5	58.6	± 10.0	72.0	73.2	73.0	73.2	73.7	± 10.0			
LYMPHO	#	0.69	0.67	0.70	0.67	0.72	± 0.33	2.28	2.26	2.39	2.26	2.33	± 0.70	2.52	2.37	2.51	2.37	2.60	± 1.50			
	%	28.7	27.8	29.0	27.8	28.6	± 12.0	30.0	29.7	31.5	29.7	30.7	± 8.0	14.0	13.4	14.0	13.4	14.3	± 8.0			
MONO	#	0.11	0.09	0.10	0.09	0.11	± 0.09	0.34	0.30	0.27	0.30	0.29	± 0.27	0.81	0.80	0.64	0.80	0.66	± 0.64			
	%	4.5	3.7	4.2	3.7	4.2	± 3.7	4.5	4.0	3.5	4.0	3.8	± 3.5	4.5	4.5	3.6	4.5	3.6	± 3.6			
EOS	#	0.14	0.16	0.14	0.16	0.12	± 0.12	0.29	0.32	0.30	0.32	0.27	± 0.27	0.90	0.80	0.90	0.80	0.76	± 0.76			
	%	6.0	6.5	6.0	6.5	4.9	± 4.9	3.8	4.2	4.0	4.2	3.5	± 3.5	5.0	4.5	5.0	4.5	4.2	± 4.2			
BASO	#	0.08	0.08	0.08	0.08	0.09	± 0.08	0.29	0.27	0.27	0.27	0.26	± 0.26	0.81	0.78	0.79	0.78	0.76	± 0.76			
	%	3.3	3.3	3.3	3.3	3.4	± 3.3	3.8	3.6	3.5	3.6	3.4	± 3.4	4.5	4.4	4.4	4.4	4.2	± 4.2			

**LOT** PX 402  
Rev 2

**CONTROL**

 (Exp.) 2017-01-05  
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4			± 0.4	7.8	7.8	7.8			± 1.0	18.6	18.6	18.6			± 2.2	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.35	2.35			± 0.16	4.54	4.54	4.54			± 0.20	5.16	5.16	5.16			± 0.25	
	g/dl	6.9	6.9	6.9			± 0.4	13.5	13.5	13.5			± 0.5	16.3	16.3	16.3			± 0.6	
HB HGB	g/l	69	69	69			± 4	135	135	135			± 5	163	163	163			± 6	
	mmol/l	4.28	4.28	4.28			± 0.25	8.38	8.38	8.38			± 0.31	10.12	10.12	10.12			± 0.37	
HT HCT	%	19.5	19.5	19.5			± 1.5	37.2	37.2	37.2			± 2.0	45.4	45.4	45.4			± 2.5	
	l/l	0.195	0.195	0.195			± 0.015	0.372	0.372	0.372			± 0.020	0.454	0.454	0.454			± 0.025	
VGM MCV	µm <sup>3</sup> ·fl	83	83	83			± 5	82	82	82			± 5	88	88	88			± 5	
TGMH MCH	pg	29.4	29.4	29.4			± 2.0	29.7	29.7	29.7			± 2.0	31.6	31.6	31.6			± 2.5	
	fmol	1.82	1.82	1.82			± 0.12	1.85	1.85	1.85			± 0.12	1.96	1.96	1.96			± 0.16	
	g/dl	35.4	35.4	35.4			± 3.0	36.3	36.3	36.3			± 3.0	35.9	35.9	35.9			± 3.0	
CCMH MCHC	g/l	354	354	354			± 30	363	363	363			± 30	359	359	359			± 30	
	mmol/l	21.97	21.97	21.97			± 1.86	22.52	22.52	22.52			± 1.86	22.29	22.29	22.29			± 1.86	
IDR RDW	%	15.0	15.0	15.0			± 4.0	16.8	16.8	16.8			± 4.0	14.7	14.7	14.7			± 4.0	
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	72			± 20	250	250	250			± 30	490	490	490			± 50	
VPM MPV	µm <sup>3</sup> ·fl	9.5	9.5	9.5			± 2.0	9.6	9.6	9.6			± 2.0	9.2	9.2	9.2			± 2.0	
NEUT	#	1.41	1.46	1.46			± 0.35	4.56	4.68	4.68			± 0.90	13.90	14.00	14.00			± 1.90	
	%	58.8	60.9	60.9			± 10.0	58.5	60.0	60.0			± 10.0	74.5	75.4	75.4			± 10.0	
LYMPHO	#	0.68	0.62	0.62			± 0.33	2.33	2.18	2.18			± 0.70	2.38	2.21	2.21			± 1.50	
	%	28.3	25.7	25.7			± 12.0	29.9	27.9	27.9			± 8.0	12.8	11.9	11.9			± 8.0	
MONO	#	0.10	0.10	0.10			± 0.10	0.35	0.37	0.37			± 0.35	0.93	0.93	0.93			± 0.93	
	%	4.1	4.2	4.2			± 4.1	4.5	4.7	4.7			± 4.5	5.0	5.0	5.0			± 5.0	
EOS	#	0.15	0.16	0.16			± 0.15	0.32	0.34	0.34			± 0.32	0.97	0.97	0.97			± 0.97	
	%	6.3	6.7	6.7			± 6.3	4.1	4.4	4.4			± 4.1	5.2	5.2	5.2			± 5.2	
BASO	#	0.06	0.06	0.06			± 0.06	0.23	0.23	0.23			± 0.23	0.47	0.47	0.47			± 0.47	
	%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5	

1300027264-A  
FRONT / RECTO  
Ref: TEMP-0821 Rev.39