


LOT PX 411
Rev 1


CONTROL

 (Exp.) 2018-07-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.4	7.4			± 1.0	17.8	17.8			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.32	2.32			± 0.16	4.63	4.63			± 0.20	5.18	5.18			± 0.25			
HB HGB	g/dl	6.8	6.8			± 0.4	13.9	13.9			± 0.5	16.3	16.3			± 0.6			
	g/l	68	68			± 4	139	139			± 5	163	163			± 6			
	mmol/l	4.22	4.22			± 0.25	8.63	8.63			± 0.31	10.12	10.12			± 0.37			
HT HCT	%	19.5	19.5			± 1.5	38.4	38.4			± 2.0	45.1	45.1			± 2.5			
	l/l	0.195	0.195			± 0.015	0.384	0.384			± 0.020	0.451	0.451			± 0.025			
VGM MCV	µm ³ fl	84	84			± 5	83	83			± 5	87	87			± 5			
TGMH MCH	pg	29.3	29.3			± 2.0	30.0	30.0			± 2.0	31.5	31.5			± 2.5			
	fmol	1.82	1.82			± 0.12	1.86	1.86			± 0.12	1.95	1.95			± 0.16			
CCMH MCHC	g/dl	34.9	34.9			± 3.0	36.2	36.2			± 3.0	36.2	36.2			± 3.0			
	g/l	349	349			± 30	362	362			± 30	362	362			± 30			
	mmol/l	21.67	21.67			± 1.86	22.46	22.46			± 1.86	22.46	22.46			± 1.86			
IDR RDW	%	13.5	12.5			± 4.0	13.8	12.5			± 4.0	12.7	11.5			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	64	64			± 20	245	245			± 30	485	485			± 50			
VPM MPV	µm ³ fl	10.7	10.7			± 2.0	10.0	10.0			± 2.0	10.0	10.0			± 2.0			
NEUT	#	1.38	1.38			± 0.35	4.23	4.23			± 0.90	12.39	12.39			± 1.90			
	%	55.0	55.0			± 10.0	57.2	57.2			± 10.0	69.6	69.6			± 10.0			
LYMPHO	#	0.76	0.76			± 0.33	2.29	2.29			± 0.70	2.74	2.74			± 1.50			
	%	30.3	30.3			± 12.0	30.9	30.9			± 8.0	15.4	15.4			± 8.0			
MONO	#	0.13	0.13			± 0.13	0.34	0.34			± 0.34	0.85	0.85			± 0.85			
	%	5.1	5.1			± 5.1	4.6	4.6			± 4.6	4.8	4.8			± 4.8			
EOS	#	0.16	0.16			± 0.16	0.27	0.27			± 0.27	1.01	1.01			± 1.01			
	%	6.3	6.3			± 6.3	3.7	3.7			± 3.7	5.7	5.7			± 5.7			
BASO	#	0.08	0.08			± 0.08	0.27	0.27			± 0.27	0.80	0.80			± 0.80			
	%	3.3	3.3			± 3.3	3.6	3.6			± 3.6	4.5	4.5			± 4.5			

LOT PX 411
Rev 1

CONTROL

 (Exp.) 2018-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCRIP	60 60C+ ES60	MS60	MSCRIP	60 60C+ ES60	MS60	MSCRIP	60 60C+ ES60	MS60	MSCRIP						
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5			± 0.4	7.4	7.4	7.5			± 1.0	18.0	17.7	18.2			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.34	2.34			± 0.16	4.64	4.64	4.65			± 0.20	5.14	5.15	5.18			± 0.25
		g/dl	6.8	6.7	6.8			± 0.4	13.9	13.8	13.9			± 0.5	16.3	16.2	16.2			± 0.6
HB	HGB	g/l	68	67	68			± 4	139	138	139			± 5	163	162	162			± 6
		mmol/l	4.22	4.16	4.22			± 0.25	8.63	8.57	8.63			± 0.31	10.12	10.06	10.06			± 0.37
HT	HCT	%	19.2	19.0	19.0			± 1.5	38.0	37.6	38.1			± 2.0	44.7	44.3	44.5			± 2.5
		l/l	0.192	0.190	0.190			± 0.015	0.380	0.376	0.381			± 0.020	0.447	0.443	0.445			± 0.025
VGM	MCV	µm ³ ; fl	81	81	81.0			± 5	82	81	82.0			± 5	87	86	86.0			± 5
TGMH	MCH	pg	28.7	28.6	29.1			± 2.0	30.0	29.7	29.9			± 2.0	31.7	31.5	31.3			± 2.5
		fmol	1.78	1.78	1.80			± 0.12	1.86	1.85	1.86			± 0.12	1.97	1.95	1.94			± 0.16
		g/dl	35.4	35.3	35.9			± 3.0	36.5	36.7	36.5			± 3.0	36.5	36.6	36.4			± 3.0
CCMH	MCHC	g/l	354	353	359			± 30	365	367	365			± 30	365	366	364			± 30
		mmol/l	22.00	21.95	22.28			± 1.86	22.69	22.80	22.64			± 1.86	22.64	22.71	22.58			± 1.86
IDR	RDW	%	13.0	13.0	12.5			± 4.0	13.5	13.0	12.5			± 4.0	12.5	12.0	12.0			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	65	65			± 20	245	245	240			± 30	490	490	475			± 50
VPM	MPV	µm ³ ; fl	10.5	10.5	9.5			± 2.0	9.8	9.8	9.2			± 2.0	9.7	9.8	9.0			± 2.0
		#	1.32	1.32	1.39			± 0.35	4.22	4.20	4.35			± 0.90	12.60	12.46	13.01			± 1.90
		%	55.0	54.9	55.5			± 10.0	57.0	56.7	58.0			± 10.0	70.0	70.4	71.5			± 10.0
		#	0.73	0.73	0.78			± 0.33	2.29	2.32	2.33			± 0.70	2.74	2.67	2.82			± 1.50
		%	30.6	30.5	31.0			± 12.0	30.9	31.3	31.0			± 8.0	15.2	15.1	15.5			± 8.0
		#	0.12	0.13	0.11			± 0.11	0.36	0.36	0.30			± 0.30	0.90	0.87	0.67			± 0.67
		%	5.1	5.4	4.5			± 4.5	4.9	4.8	4.0			± 4.0	5.0	4.9	3.7			± 3.7
		#	0.14	0.14	0.14			± 0.14	0.27	0.27	0.26			± 0.26	0.95	0.92	0.91			± 0.91
		%	6.0	5.9	5.5			± 5.5	3.6	3.7	3.5			± 3.5	5.3	5.2	5.0			± 5.0
		#	0.08	0.08	0.09			± 0.08	0.27	0.26	0.26			± 0.26	0.81	0.78	0.78			± 0.78
		%	3.3	3.3	3.5			± 3.3	3.6	3.5	3.5			± 3.5	4.5	4.4	4.3			± 4.3

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.42