


LOT PX 420
Rev 2


CONTROL

 (Exp.) 2020-01-05
(YYYY - MM - DD)

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7				± 0.4	7.8	7.8				± 1.0	18.6	18.6				± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.44	2.44				± 0.16	4.48	4.48				± 0.20	5.10	5.10				± 0.25	
HB HGB	g/dl	7.3	7.3				± 0.4	13.5	13.5				± 0.5	16.2	16.2				± 0.6	
	g/l	73	73				± 4	135	135				± 5	162	162				± 6	
	mmol/l	4.53	4.53				± 0.25	8.38	8.38				± 0.31	10.06	10.06				± 0.37	
HT HCT	%	21.0	21.0				± 1.5	37.6	37.6				± 2.0	44.4	44.4				± 2.5	
	l/l	0.210	0.210				± 0.015	0.376	0.376				± 0.020	0.444	0.444				± 0.025	
VGM MCV	µm ³ fl	86	86				± 5	84	84				± 5	87	87				± 5	
TGMH MCH	pg	29.9	29.9				± 2.0	30.1	30.1				± 2.0	31.8	31.8				± 2.5	
	fmol	1.86	1.86				± 0.12	1.87	1.87				± 0.12	1.97	1.97				± 0.16	
CCMH MCHC	g/dl	34.8	34.8				± 3.0	35.9	35.9				± 3.0	36.5	36.5				± 3.0	
	g/l	348	348				± 30	359	359				± 30	365	365				± 30	
	mmol/l	21.60	21.60				± 1.86	22.28	22.28				± 1.86	22.67	22.67				± 1.86	
IDR RDW	%	13.5	12.5				± 4.0	13.8	13.0				± 4.0	13.0	12.0				± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	65	65				± 20	241	241				± 30	500	500				± 50	
VPM MPV	µm ³ fl	9.3	9.4				± 2.0	9.1	9.4				± 2.0	10.0	10.5				± 2.0	
NEUT	#	1.49	1.49				± 0.35	4.27	4.27				± 0.90	12.72	12.72				± 1.90	
	%	55.2	55.2				± 10.0	54.7	54.7				± 10.0	68.4	68.4				± 10.0	
LYMPHO	#	0.73	0.73				± 0.33	2.29	2.29				± 0.70	2.66	2.66				± 1.50	
	%	26.9	26.9				± 12.0	29.4	29.4				± 8.0	14.3	14.3				± 8.0	
MONO	#	0.19	0.19				± 0.19	0.53	0.53				± 0.53	1.21	1.21				± 1.21	
	%	7.1	7.1				± 7.1	6.8	6.8				± 6.8	6.5	6.5				± 6.5	
EOS	#	0.20	0.20				± 0.20	0.44	0.44				± 0.44	1.19	1.19				± 1.19	
	%	7.4	7.4				± 7.4	5.7	5.7				± 5.7	6.4	6.4				± 6.4	
BASO	#	0.09	0.09				± 0.09	0.27	0.27				± 0.27	0.82	0.82				± 0.82	
	%	3.4	3.4				± 3.4	3.4	3.4				± 3.4	4.4	4.4				± 4.4	

LOT PX 420
Rev 2

CONTROL

 (Exp.) 2020-01-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.7		± 0.4	7.9	7.8	7.9		± 1.0	18.9	18.7	19.0		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.50	2.45	2.45		± 0.16	4.48	4.45	4.45		± 0.20	5.08	5.05	5.08		± 0.25			
	g/dl	7.3	7.2	7.1		± 0.4	13.5	13.3	13.3		± 0.5	16.0	16.0	15.9		± 0.6			
HB HGB	g/l	73	72	71		± 4	135	133	133		± 5	160	160	159		± 6			
	mmol/l	4.53	4.47	4.41		± 0.25	8.38	8.26	8.26		± 0.31	9.94	9.94	9.87		± 0.37			
HT HCT	%	20.5	20.1	19.8		± 1.5	37.2	36.5	36.3		± 2.0	44.2	43.4	42.7		± 2.5			
	l/l	0.205	0.201	0.198		± 0.015	0.372	0.365	0.363		± 0.020	0.442	0.434	0.427		± 0.025			
VGM MCV	µm ³ ·fl	82	82	81.0		± 5	83	82	81.5		± 5	87	86	84.0		± 5			
TGMH MCH	pg	29.2	29.4	29.0		± 2.0	30.1	29.9	29.9		± 2.0	31.5	31.7	31.3		± 2.5			
	fmol	1.81	1.82	1.80		± 0.12	1.87	1.86	1.86		± 0.12	1.96	1.97	1.94		± 0.16			
	g/dl	35.6	35.8	35.8		± 3.0	36.3	36.4	36.7		± 3.0	36.2	36.8	37.3		± 3.0			
CCMH MCHC	g/l	356	358	358		± 30	363	364	367		± 30	362	368	373		± 30			
	mmol/l	22.11	22.26	22.22		± 1.86	22.55	22.63	22.77		± 1.86	22.48	22.88	23.14		± 1.86			
IDR RDW	%	13.0	13.0	12.5		± 4.0	13.0	13.0	12.5		± 4.0	12.5	12.0	12.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	66	65		± 20	243	247	235		± 30	503	513	480		± 50			
VPM MPV	µm ³ ·fl	9.1	9.3	8.7		± 2.0	9.1	9.2	8.5		± 2.0	10.2	10.4	9.2		± 2.0			
NEUT	#	1.42	1.43	1.49		± 0.35	4.34	4.27	4.31		± 0.90	12.87	12.72	13.11		± 1.90			
	%	54.8	55.1	55.0		± 10.0	54.9	54.7	54.5		± 10.0	68.1	68.0	69.0		± 10.0			
LYMPHO	#	0.75	0.74	0.77		± 0.33	2.54	2.45	2.57		± 0.70	3.12	2.97	3.14		± 1.50			
	%	28.8	28.5	28.5		± 12.0	32.1	31.4	32.5		± 8.0	16.5	15.9	16.5		± 8.0			
MONO	#	0.14	0.16	0.16		± 0.14	0.40	0.45	0.40		± 0.40	0.93	1.03	0.86		± 0.86			
	%	5.5	6.0	6.0		± 5.5	5.0	5.8	5.0		± 5.0	4.9	5.5	4.5		± 4.5			
EOS	#	0.20	0.18	0.18		± 0.18	0.34	0.37	0.36		± 0.34	1.13	1.16	1.05		± 1.05			
	%	7.5	7.0	6.5		± 6.5	4.3	4.7	4.5		± 4.3	6.0	6.2	5.5		± 5.5			
BASO	#	0.09	0.09	0.11		± 0.09	0.29	0.27	0.28		± 0.27	0.85	0.82	0.86		± 0.82			
	%	3.4	3.4	4.0		± 3.4	3.7	3.4	3.5		± 3.4	4.5	4.4	4.5		± 4.4			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43