


**LOT** PX 410  
Rev 2


**CONTROL**

 (Exp.) 2018-05-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5			± 0.4	7.4	7.4			± 1.0	17.5	17.5			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.26	2.26			± 0.16	4.63	4.63			± 0.20	5.01	5.01			± 0.25			
HB HGB	g/dl	6.6	6.6			± 0.4	14.0	14.0			± 0.5	15.8	15.8			± 0.6			
	g/l	66	66			± 4	140	140			± 5	158	158			± 6			
	mmol/l	4.10	4.10			± 0.25	8.69	8.69			± 0.31	9.81	9.81			± 0.37			
HT HCT	%	19.0	19.0			± 1.5	38.9	38.9			± 2.0	43.6	43.6			± 2.5			
	l/l	0.190	0.190			± 0.015	0.389	0.389			± 0.020	0.436	0.436			± 0.025			
VGM MCV	µm <sup>3</sup> fl	84	84			± 5	84	84			± 5	87	87			± 5			
TGMH MCH	pg	29.2	29.2			± 2.0	30.2	30.2			± 2.0	31.5	31.5			± 2.5			
	fmol	1.81	1.81			± 0.12	1.88	1.88			± 0.12	1.96	1.96			± 0.16			
CCMH MCHC	g/dl	34.8	34.8			± 3.0	36.0	36.0			± 3.0	36.2	36.2			± 3.0			
	g/l	348	348			± 30	360	360			± 30	362	362			± 30			
	mmol/l	21.59	21.59			± 1.86	22.35	22.35			± 1.86	22.51	22.51			± 1.86			
IDR RDW	%	13.5	12.5			± 4.0	14.0	13.0			± 4.0	13.0	12.0			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	68	68			± 20	245	245			± 30	486	486			± 50			
VPM MPV	µm <sup>3</sup> fl	9.6	9.6			± 2.0	9.8	9.8			± 2.0	9.3	9.3			± 2.0			
NEUT	#	1.39	1.39			± 0.35	4.15	4.15			± 0.90	12.04	12.04			± 1.90			
	%	55.7	55.7			± 10.0	56.1	56.1			± 10.0	68.8	68.8			± 10.0			
LYMPHO	#	0.75	0.75			± 0.33	2.34	2.34			± 0.70	2.92	2.92			± 1.50			
	%	30.0	30.0			± 12.0	31.6	31.6			± 8.0	16.7	16.7			± 8.0			
MONO	#	0.10	0.10			± 0.10	0.36	0.36			± 0.36	0.86	0.86			± 0.86			
	%	4.0	4.0			± 4.0	4.8	4.8			± 4.8	4.9	4.9			± 4.9			
EOS	#	0.18	0.18			± 0.18	0.28	0.28			± 0.28	0.91	0.91			± 0.91			
	%	7.1	7.1			± 7.1	3.8	3.8			± 3.8	5.2	5.2			± 5.2			
BASO	#	0.08	0.08			± 0.08	0.27	0.27			± 0.27	0.77	0.77			± 0.77			
	%	3.2	3.2			± 3.2	3.7	3.7			± 3.7	4.4	4.4			± 4.4			

**LOT** PX 410  
Rev 2

**CONTROL**

 (Exp.) 2018-05-05  
(YYYY - MM - DD)

		ABX Lysebilo																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.5		± 0.4	7.5	7.5	7.8			± 1.0	17.8	17.7	18.5		± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.30	2.33		± 0.16	4.64	4.64	4.67			± 0.20	5.00	5.02	5.05		± 0.25		
	g/dl	6.7	6.6	6.7		± 0.4	14.1	14.0	14.1			± 0.5	15.8	15.8	15.8		± 0.6		
HB HGB	g/l	67	66	67		± 4	141	140	141			± 5	158	158	158		± 6		
	mmol/l	4.16	4.10	4.16		± 0.25	8.76	8.69	8.76			± 0.31	9.81	9.81	9.81		± 0.37		
HT HCT	%	19.0	18.6	18.6		± 1.5	38.5	38.5	38.8			± 2.0	43.5	43.2	43.4		± 2.5		
	l/l	0.190	0.186	0.186		± 0.015	0.385	0.385	0.388			± 0.020	0.435	0.432	0.434		± 0.025		
VGM MCV	µm <sup>3</sup> /fl	81	81	80.0		± 5	83	83	83.0			± 5	87	86	86.0		± 5		
TGMH MCH	pg	28.5	28.7	28.8		± 2.0	30.4	30.2	30.2			± 2.0	31.6	31.5	31.3		± 2.5		
	fmol	1.77	1.78	1.79		± 0.12	1.89	1.87	1.87			± 0.12	1.96	1.95	1.94		± 0.16		
	g/dl	35.2	35.4	35.9		± 3.0	36.6	36.4	36.4			± 3.0	36.3	36.6	36.4		± 3.0		
CCMH MCHC	g/l	352	354	359		± 30	366	364	364			± 30	363	366	364		± 30		
	mmol/l	21.86	22.00	22.32		± 1.86	22.74	22.57	22.59			± 1.86	22.56	22.73	22.59		± 1.86		
IDR RDW	%	14.2	12.5	13.0		± 4.0	14.5	13.0	13.0			± 4.0	13.6	12.0	12.5		± 4.0		
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	70	70		± 20	253	252	245			± 30	490	500	485		± 50		
VPM MPV	µm <sup>3</sup> /fl	9.4	9.5	8.8		± 2.0	9.4	9.4	9.0			± 2.0	9.0	9.1	8.6		± 2.0		
NEUT	#	1.34	1.33	1.41		± 0.35	4.22	4.18	4.45			± 0.90	12.48	12.32	13.14		± 1.90		
	%	55.8	55.4	56.5		± 10.0	56.3	55.7	57.0			± 10.0	70.1	69.6	71.0		± 10.0		
LYMPHO	#	0.72	0.72	0.75		± 0.33	2.39	2.35	2.46			± 0.70	2.88	2.81	2.96		± 1.50		
	%	30.2	30.0	30.0		± 12.0	31.8	31.3	31.5			± 8.0	16.2	15.9	16.0		± 8.0		
MONO	#	0.10	0.11	0.10		± 0.10	0.35	0.42	0.35			± 0.35	0.82	0.97	0.83		± 0.82		
	%	4.0	4.5	4.0		± 4.0	4.6	5.6	4.5			± 4.5	4.6	5.5	4.5		± 4.5		
EOS	#	0.16	0.17	0.15		± 0.15	0.28	0.29	0.27			± 0.27	0.84	0.81	0.83		± 0.81		
	%	6.8	6.9	6.0		± 6.0	3.7	3.8	3.5			± 3.5	4.7	4.6	4.5		± 4.5		
BASO	#	0.08	0.08	0.09		± 0.08	0.27	0.27	0.27			± 0.27	0.78	0.78	0.74		± 0.74		
	%	3.2	3.2	3.5		± 3.2	3.6	3.6	3.5			± 3.5	4.4	4.4	4.0		± 4.0		

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.42