


LOT PX 404
Rev 1


CONTROL

 (Exp.) 2017-05-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | | UNITES UNITS | ABX Lysebio | | | | | | | | | | | | | | | | | | |
|--------------------------|------|--|--------------------|------------|-------|-------|--------|---------|--------------------|------------|-------|-------|--------|---------|--------------------|------------|-------|-------|--------|---------|-------------------------|
| | | | CONTROL | | | | | L | CONTROL | | | | | N | CONTROL | | | | | H | TOLERANCES TOLERANCE |
| | | | PENTRA | | | | | | PENTRA | | | | | | PENTRA | | | | | | |
| | | | 60 60C+ ES60 | 80 XL80 | MS60 | XLR | MSCRIP | | 60 60C+ ES60 | 80 XL80 | MS60 | XLR | MSCRIP | | 60 60C+ ES60 | 80 XL80 | MS60 | XLR | MSCRIP | | |
| GB | WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.5 | 2.5 | 2.5 | 2.5 | 2.6 | ± 0.4 | 7.3 | 7.3 | 7.4 | 7.3 | 7.5 | ± 1.0 | 17.4 | 17.1 | 17.4 | 17.1 | 17.7 | ± 2.2 | |
| GR | RBC | 10 ⁹ /mm ³ ; 10 ¹² /l | 2.34 | 2.30 | 2.35 | 2.30 | 2.35 | ± 0.16 | 4.58 | 4.60 | 4.60 | 4.60 | 4.62 | ± 0.20 | 5.15 | 5.19 | 5.20 | 5.19 | 5.22 | ± 0.25 | |
| HB | HGB | g/dl | 6.6 | 6.6 | 6.6 | 6.6 | 6.6 | ± 0.4 | 13.5 | 13.6 | 13.6 | 13.6 | 13.5 | ± 0.5 | 16.4 | 16.5 | 16.5 | 16.5 | 16.4 | ± 0.6 | |
| | | g/l | 66 | 66 | 66 | 66 | 66 | ± 4 | 135 | 136 | 136 | 136 | 135 | ± 5 | 164 | 165 | 165 | 165 | 164 | ± 6 | |
| | | mmol/l | 4.10 | 4.10 | 4.10 | 4.10 | 4.10 | ± 0.25 | 8.38 | 8.45 | 8.45 | 8.45 | 8.38 | ± 0.31 | 10.18 | 10.25 | 10.25 | 10.25 | 10.18 | ± 0.37 | |
| HT | HCT | % | 18.7 | 19.1 | 18.8 | 19.1 | 18.7 | ± 1.5 | 37.1 | 37.7 | 36.8 | 37.3 | 37.4 | ± 2.0 | 44.8 | 45.2 | 44.7 | 44.6 | 44.9 | ± 2.5 | |
| | | l/l | 0.187 | 0.191 | 0.188 | 0.191 | 0.187 | ± 0.015 | 0.371 | 0.377 | 0.368 | 0.373 | 0.374 | ± 0.020 | 0.448 | 0.452 | 0.447 | 0.446 | 0.449 | ± 0.025 | |
| VGM | MCV | µm ³ ; fl | 80 | 83 | 80 | 83 | 79.5 | ± 5 | 81 | 82 | 80 | 81 | 81.0 | ± 5 | 87 | 87 | 86 | 86 | 86.0 | ± 5 | |
| TGMH | MCH | pg | 28.2 | 28.7 | 28.1 | 28.7 | 28.1 | ± 2.0 | 29.5 | 29.6 | 29.6 | 29.6 | 29.2 | ± 2.0 | 31.8 | 31.8 | 31.7 | 31.8 | 31.4 | ± 2.5 | |
| | | fmol | 1.75 | 1.78 | 1.74 | 1.78 | 1.74 | ± 0.12 | 1.83 | 1.84 | 1.84 | 1.84 | 1.81 | ± 0.12 | 1.98 | 1.97 | 1.97 | 1.97 | 1.95 | ± 0.16 | |
| CCMH | MCHC | g/dl | 35.3 | 34.6 | 35.1 | 34.6 | 35.3 | ± 3.0 | 36.4 | 36.1 | 37.0 | 36.5 | 36.1 | ± 3.0 | 36.6 | 36.5 | 36.9 | 37.0 | 36.5 | ± 3.0 | |
| | | g/l | 353 | 346 | 351 | 346 | 353 | ± 30 | 364 | 361 | 370 | 365 | 361 | ± 30 | 366 | 365 | 369 | 370 | 365 | ± 30 | |
| | | mmol/l | 21.89 | 21.47 | 21.80 | 21.47 | 21.94 | ± 1.86 | 22.60 | 22.39 | 22.95 | 22.67 | 22.40 | ± 1.86 | 22.73 | 22.69 | 22.91 | 22.96 | 22.69 | ± 1.86 | |
| IDR | RDW | % | 12.8 | 13.0 | 12.7 | 13.0 | 12.3 | ± 4.0 | 12.5 | 13.0 | 12.6 | 13.0 | 12.5 | ± 4.0 | 12.6 | 13.0 | 12.3 | 13.0 | 12.8 | ± 4.0 | |
| PLAQ. | PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 74 | 72 | 73 | 72 | 73 | ± 20 | 250 | 247 | 253 | 247 | 243 | ± 30 | 495 | 495 | 520 | 495 | 490 | ± 50 | |
| VPM | MPV | µm ³ ; fl | 8.8 | 9.0 | 8.9 | 9.0 | 8.3 | ± 2.0 | 9.5 | 9.8 | 9.6 | 9.8 | 9.1 | ± 2.0 | 9.1 | 9.3 | 9.1 | 9.3 | 8.7 | ± 2.0 | |
| NEUT | # | | 1.39 | 1.37 | 1.41 | 1.37 | 1.46 | ± 0.35 | 4.07 | 4.07 | 4.08 | 4.07 | 4.19 | ± 0.90 | 12.21 | 11.83 | 12.23 | 11.83 | 12.53 | ± 1.90 | |
| | | % | 55.5 | 54.9 | 56.4 | 54.9 | 56.0 | ± 10.0 | 55.8 | 55.7 | 55.2 | 55.7 | 55.8 | ± 10.0 | 70.2 | 69.2 | 70.3 | 69.2 | 70.8 | ± 10.0 | |
| LYMPHO | # | | 0.79 | 0.80 | 0.79 | 0.80 | 0.84 | ± 0.33 | 2.39 | 2.39 | 2.52 | 2.39 | 2.55 | ± 0.70 | 3.03 | 2.98 | 3.10 | 2.98 | 3.20 | ± 1.50 | |
| | | % | 31.5 | 31.9 | 31.5 | 31.9 | 32.2 | ± 12.0 | 32.8 | 32.7 | 34.0 | 32.7 | 34.0 | ± 8.0 | 17.4 | 17.4 | 17.8 | 17.4 | 18.1 | ± 8.0 | |
| MONO | # | | 0.10 | 0.11 | 0.09 | 0.11 | 0.10 | ± 0.09 | 0.30 | 0.31 | 0.26 | 0.31 | 0.27 | ± 0.26 | 0.70 | 0.80 | 0.61 | 0.80 | 0.60 | ± 0.60 | |
| | | % | 4.0 | 4.3 | 3.4 | 4.3 | 3.7 | ± 3.4 | 4.1 | 4.3 | 3.5 | 4.3 | 3.6 | ± 3.5 | 4.0 | 4.7 | 3.5 | 4.7 | 3.4 | ± 3.4 | |
| EOS | # | | 0.14 | 0.14 | 0.14 | 0.14 | 0.12 | ± 0.12 | 0.27 | 0.28 | 0.28 | 0.28 | 0.25 | ± 0.25 | 0.73 | 0.75 | 0.71 | 0.75 | 0.64 | ± 0.64 | |
| | | % | 5.6 | 5.7 | 5.4 | 5.7 | 4.7 | ± 4.7 | 3.7 | 3.8 | 3.8 | 3.8 | 3.3 | ± 3.3 | 4.2 | 4.4 | 4.1 | 4.4 | 3.6 | ± 3.6 | |
| BASO | # | | 0.09 | 0.08 | 0.08 | 0.08 | 0.09 | ± 0.08 | 0.26 | 0.26 | 0.26 | 0.26 | 0.25 | ± 0.25 | 0.73 | 0.74 | 0.75 | 0.74 | 0.73 | ± 0.73 | |
| | | % | 3.4 | 3.2 | 3.3 | 3.2 | 3.4 | ± 3.2 | 3.6 | 3.5 | 3.5 | 3.5 | 3.3 | ± 3.3 | 4.2 | 4.3 | 4.3 | 4.3 | 4.1 | ± 4.1 | |

LOT PX 404
Rev 1

CONTROL

 (Exp.) 2017-05-05
(YYYY - MM - DD)

| | | ABX Lysebio | | | | | | | | | | | | | | | | | | | | | | | |
|--------------------------|--|-------------|-------|----------|--|--|---------|-----------|-------|----------|--|--|---------|-----------|-------|----------|--|--|---------|-------------------------|-----------|-------|----------|--|--|
| PARAMETRES PARAMETERS | UNITES UNITS | CONTROL | | | | | L | CONTROL | | | | | N | CONTROL | | | | | H | TOLERANCES TOLERANCE | | | | | |
| | | PENTRA | | | | | | PENTRA | | | | | | PENTRA | | | | | | | | | | | |
| | | 120 | DX120 | DX NEXUS | | | | 120 | DX120 | DX NEXUS | | | | 120 | DX120 | DX NEXUS | | | | | 120 | DX120 | DX NEXUS | | |
| | | 120 RETIC | DF120 | DF NEXUS | | | | 120 RETIC | DF120 | DF NEXUS | | | | 120 RETIC | DF120 | DF NEXUS | | | | | 120 RETIC | DF120 | DF NEXUS | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.6 | 2.6 | 2.6 | | | ± 0.4 | 7.7 | 7.7 | 7.7 | | | ± 1.0 | 18.5 | 18.5 | 18.5 | | | ± 2.2 | | | | | | |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.35 | 2.35 | 2.35 | | | ± 0.16 | 4.63 | 4.63 | 4.63 | | | ± 0.20 | 5.28 | 5.28 | 5.28 | | | ± 0.25 | | | | | | |
| | g/dl | 6.7 | 6.7 | 6.7 | | | ± 0.4 | 13.5 | 13.5 | 13.5 | | | ± 0.5 | 16.3 | 16.3 | 16.3 | | | ± 0.6 | | | | | | |
| HB HGB | g/l | 67 | 67 | 67 | | | ± 4 | 135 | 135 | 135 | | | ± 5 | 163 | 163 | 163 | | | ± 6 | | | | | | |
| | mmol/l | 4.16 | 4.16 | 4.16 | | | ± 0.25 | 8.38 | 8.38 | 8.38 | | | ± 0.31 | 10.12 | 10.12 | 10.12 | | | ± 0.37 | | | | | | |
| HT HCT | % | 19.5 | 19.5 | 19.5 | | | ± 1.5 | 38.0 | 38.0 | 38.0 | | | ± 2.0 | 46.5 | 46.5 | 46.5 | | | ± 2.5 | | | | | | |
| | l/l | 0.195 | 0.195 | 0.195 | | | ± 0.015 | 0.380 | 0.380 | 0.380 | | | ± 0.020 | 0.465 | 0.465 | 0.465 | | | ± 0.025 | | | | | | |
| VGM MCV | µm ³ ; fl | 83 | 83 | 83 | | | ± 5 | 82 | 82 | 82 | | | ± 5 | 88 | 88 | 88 | | | ± 5 | | | | | | |
| TGMH MCH | pg | 28.5 | 28.5 | 28.5 | | | ± 2.0 | 29.2 | 29.2 | 29.2 | | | ± 2.0 | 30.9 | 30.9 | 30.9 | | | ± 2.5 | | | | | | |
| | fmol | 1.77 | 1.77 | 1.77 | | | ± 0.12 | 1.81 | 1.81 | 1.81 | | | ± 0.12 | 1.92 | 1.92 | 1.92 | | | ± 0.16 | | | | | | |
| | g/dl | 34.4 | 34.4 | 34.4 | | | ± 3.0 | 35.6 | 35.6 | 35.6 | | | ± 3.0 | 35.1 | 35.1 | 35.1 | | | ± 3.0 | | | | | | |
| CCMH MCHC | g/l | 344 | 344 | 344 | | | ± 30 | 356 | 356 | 356 | | | ± 30 | 351 | 351 | 351 | | | ± 30 | | | | | | |
| | mmol/l | 21.33 | 21.33 | 21.33 | | | ± 1.86 | 22.08 | 22.08 | 22.08 | | | ± 1.86 | 21.79 | 21.79 | 21.79 | | | ± 1.86 | | | | | | |
| IDR RDW | % | 14.8 | 14.8 | 14.8 | | | ± 4.0 | 16.2 | 16.2 | 16.2 | | | ± 4.0 | 15.4 | 15.4 | 15.4 | | | ± 4.0 | | | | | | |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 72 | 72 | 72 | | | ± 20 | 252 | 252 | 252 | | | ± 30 | 495 | 495 | 495 | | | ± 50 | | | | | | |
| VPM MPV | µm ³ ; fl | 9.1 | 9.1 | 9.1 | | | ± 2.0 | 9.7 | 9.7 | 9.7 | | | ± 2.0 | 9.1 | 9.1 | 9.1 | | | ± 2.0 | | | | | | |
| NEUT | # | 1.43 | 1.49 | 1.49 | | | ± 0.35 | 4.24 | 4.37 | 4.37 | | | ± 0.90 | 12.90 | 13.20 | 13.20 | | | ± 1.90 | | | | | | |
| | % | 54.9 | 57.3 | 57.3 | | | ± 10.0 | 55.1 | 56.7 | 56.7 | | | ± 10.0 | 69.6 | 71.0 | 71.0 | | | ± 10.0 | | | | | | |
| LYMPHO | # | 0.82 | 0.75 | 0.75 | | | ± 0.33 | 2.45 | 2.29 | 2.29 | | | ± 0.70 | 3.15 | 2.90 | 2.90 | | | ± 1.50 | | | | | | |
| | % | 31.6 | 28.7 | 28.7 | | | ± 12.0 | 31.8 | 29.7 | 29.7 | | | ± 8.0 | 17.0 | 15.7 | 15.7 | | | ± 8.0 | | | | | | |
| MONO | # | 0.14 | 0.14 | 0.14 | | | ± 0.14 | 0.47 | 0.49 | 0.49 | | | ± 0.47 | 1.17 | 1.15 | 1.15 | | | ± 1.15 | | | | | | |
| | % | 5.3 | 5.4 | 5.4 | | | ± 5.3 | 6.1 | 6.3 | 6.3 | | | ± 6.1 | 6.3 | 6.2 | 6.2 | | | ± 6.2 | | | | | | |
| EOS | # | 0.15 | 0.16 | 0.16 | | | ± 0.15 | 0.31 | 0.33 | 0.33 | | | ± 0.31 | 0.85 | 0.85 | 0.85 | | | ± 0.85 | | | | | | |
| | % | 5.7 | 6.1 | 6.1 | | | ± 5.7 | 4.0 | 4.3 | 4.3 | | | ± 4.0 | 4.6 | 4.6 | 4.6 | | | ± 4.6 | | | | | | |
| BASO | # | 0.07 | 0.07 | 0.07 | | | ± 0.07 | 0.23 | 0.23 | 0.23 | | | ± 0.23 | 0.46 | 0.46 | 0.46 | | | ± 0.46 | | | | | | |
| | % | 2.5 | 2.5 | 2.5 | | | ± 2.5 | 3.0 | 3.0 | 3.0 | | | ± 3.0 | 2.5 | 2.5 | 2.5 | | | ± 2.5 | | | | | | |

Ref: TEMP-0821 Rev.40 FRONT / RECTO 1300032435-A