


LOT PX 416
Rev 1


CONTROL

 (Exp.) 2019-05-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.6	7.6			± 1.0	17.3	17.3			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.38	2.38			± 0.16	4.58	4.58			± 0.20	5.22	5.22			± 0.25			
HB HGB	g/dl	6.8	6.8			± 0.4	13.9	13.9			± 0.5	16.8	16.8			± 0.6			
	g/l	68	68			± 4	139	139			± 5	168	168			± 6			
	mmol/l	4.22	4.22			± 0.25	8.63	8.63			± 0.31	10.43	10.43			± 0.37			
HT HCT	%	20.0	20.0			± 1.5	38.5	38.5			± 2.0	46.5	46.5			± 2.5			
	l/l	0.200	0.200			± 0.015	0.385	0.385			± 0.020	0.465	0.465			± 0.025			
VGM MCV	µm ³ ; fl	84	84			± 5	84	84			± 5	89	89			± 5			
TGMH MCH	pg	28.6	28.6			± 2.0	30.3	30.3			± 2.0	32.2	32.2			± 2.5			
	fmol	1.77	1.77			± 0.12	1.88	1.88			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	34.0	34.0			± 3.0	36.1	36.1			± 3.0	36.2	36.2			± 3.0			
	g/l	340	340			± 30	361	361			± 30	362	362			± 30			
	mmol/l	21.12	21.12			± 1.86	22.44	22.44			± 1.86	22.46	22.46			± 1.86			
IDR RDW	%	13.5	12.3			± 4.0	13.8	12.6			± 4.0	13.8	12.5			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70			± 20	251	251			± 30	495	495			± 50			
VPM MPV	µm ³ ; fl	10.3	10.3			± 2.0	9.6	9.6			± 2.0	9.9	9.9			± 2.0			
NEUT	#	1.43	1.43			± 0.35	4.20	4.20			± 0.90	11.90	11.90			± 1.90			
	%	57.2	57.2			± 10.0	55.2	55.2			± 10.0	68.8	68.8			± 10.0			
LYMPHO	#	0.71	0.71			± 0.33	2.32	2.32			± 0.70	2.54	2.54			± 1.50			
	%	28.3	28.3			± 12.0	30.5	30.5			± 8.0	14.7	14.7			± 8.0			
MONO	#	0.11	0.11			± 0.11	0.46	0.46			± 0.46	0.93	0.93			± 0.93			
	%	4.3	4.3			± 4.3	6.1	6.1			± 6.1	5.4	5.4			± 5.4			
EOS	#	0.17	0.17			± 0.17	0.33	0.33			± 0.33	1.11	1.11			± 1.11			
	%	6.7	6.7			± 6.7	4.4	4.4			± 4.4	6.4	6.4			± 6.4			
BASO	#	0.09	0.09			± 0.09	0.29	0.29			± 0.29	0.81	0.81			± 0.81			
	%	3.5	3.5			± 3.5	3.8	3.8			± 3.8	4.7	4.7			± 4.7			

LOT PX 416
Rev 1

CONTROL

 (Exp.) 2019-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR			60 60C+ ES60	MS60		MSCR	
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.3	2.5			± 0.4	7.6	7.5	7.9			± 1.0	17.4	16.9	17.7			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.40	2.38			± 0.16	4.58	4.58	4.58			± 0.20	5.20	5.20	5.24			± 0.25
		g/dl	6.8	6.8	6.8			± 0.4	13.8	13.8	13.8			± 0.5	16.7	16.6	16.7			± 0.6
HB	HGB	g/l	68	68	68			± 4	138	138	138			± 5	167	166	167			± 6
		mmol/l	4.22	4.22	4.22			± 0.25	8.57	8.57	8.57			± 0.31	10.37	10.31	10.37			± 0.37
HT	HCT	%	19.7	19.2	19.2			± 1.5	38.0	37.6	37.8			± 2.0	46.3	45.8	45.6			± 2.5
		l/l	0.197	0.192	0.192			± 0.015	0.380	0.376	0.378			± 0.020	0.463	0.458	0.456			± 0.025
VGM	MCV	µm ³ ·fl	81	80	80.5			± 5	83	82	82.5			± 5	89	88	87.0			± 5
TGMH	MCH	pg	28.0	28.3	28.6			± 2.0	30.1	30.1	30.1			± 2.0	32.1	31.9	31.9			± 2.5
		fmol	1.74	1.76	1.77			± 0.12	1.87	1.87	1.87			± 0.12	1.99	1.98	1.98			± 0.16
		g/dl	34.5	35.4	35.5			± 3.0	36.3	36.7	36.5			± 3.0	36.1	36.3	36.6			± 3.0
CCMH	MCHC	g/l	345	354	355			± 30	363	367	365			± 30	361	363	366			± 30
		mmol/l	21.45	21.99	22.04			± 1.86	22.54	22.82	22.68			± 1.86	22.41	22.53	22.75			± 1.86
IDR	RDW	%	13.0	12.5	12.5			± 4.0	13.0	13.0	12.0			± 4.0	12.5	13.0	12.5			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	73	73	70			± 20	250	251	245			± 30	490	500	480			± 50
VPM	MPV	µm ³ ·fl	10.0	10.1	9.4			± 2.0	9.3	9.4	8.7			± 2.0	9.6	9.7	9.0			± 2.0
		#	1.40	1.31	1.41			± 0.35	4.25	4.13	4.42			± 0.90	12.18	11.88	12.66			± 1.90
		%	58.2	56.9	56.5			± 10.0	55.9	55.0	56.0			± 10.0	70.0	70.3	71.5			± 10.0
		#	0.69	0.69	0.78			± 0.33	2.42	2.42	2.57			± 0.70	2.64	2.62	2.83			± 1.50
LYMPHO		%	28.7	30.0	31.0			± 12.0	31.9	32.2	32.5			± 8.0	15.2	15.5	16.0			± 8.0
		#	0.07	0.08	0.08			± 0.07	0.33	0.41	0.36			± 0.33	0.80	0.76	0.62			± 0.62
MONO		%	3.1	3.5	3.0			± 3.0	4.4	5.4	4.5			± 4.4	4.6	4.5	3.5			± 3.5
		#	0.15	0.14	0.15			± 0.14	0.30	0.29	0.28			± 0.28	0.96	0.88	0.80			± 0.80
EOS		%	6.4	6.1	6.0			± 6.0	4.0	3.9	3.5			± 3.5	5.5	5.2	4.5			± 4.5
		#	0.09	0.08	0.09			± 0.08	0.29	0.26	0.28			± 0.26	0.82	0.76	0.80			± 0.76
BASO		%	3.6	3.5	3.5			± 3.5	3.8	3.5	3.5			± 3.5	4.7	4.5	4.5			± 4.5

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.43