


LOT PX 429
Rev 2


CONTROL

 (Exp.) 2021-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA					PENTRA					PENTRA					
			XL80	XLR				XL80	XLR				XL80	XLR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7			± 0.4	7.5	7.5			± 1.0	17.7	17.7			± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.11	2.11			± 0.16	4.59	4.59			± 0.20	5.14	5.14			± 0.25	
HB	HGB	g/dl	6.0	6.0			± 0.4	13.9	13.9			± 0.5	16.8	16.8			± 0.6	
		g/l	60	60			± 4	139	139			± 5	168	168			± 6	
		mmol/l	3.73	3.73			± 0.25	8.63	8.63			± 0.31	10.43	10.43			± 0.37	
HT	HCT	%	17.5	17.5			± 1.5	39.0	39.0			± 2.0	45.7	45.7			± 2.5	
		l/l	0.175	0.175			± 0.015	0.390	0.390			± 0.020	0.457	0.457			± 0.025	
VGM	MCV	µm ³ ; fl	83	83			± 5	85	85			± 5	89	89			± 5	
TGMH	MCH	pg	28.4	28.4			± 2.0	30.3	30.3			± 2.0	32.7	32.7			± 2.5	
		fmol	1.77	1.77			± 0.12	1.88	1.88			± 0.12	2.03	2.03			± 0.16	
CCMH	MCHC	g/dl	34.3	34.3			± 3.0	35.6	35.6			± 3.0	36.7	36.7			± 3.0	
		g/l	343	343			± 30	356	356			± 30	367	367			± 30	
		mmol/l	21.28	21.28			± 1.86	22.12	22.12			± 1.86	22.81	22.81			± 1.86	
IDR	RDW	%	14.0	14.0			± 5.0	13.5	13.5			± 5.0	12.7	12.7			± 5.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67			± 20	238	238			± 30	485	485			± 50	
VPM	MPV	µm ³ ; fl	10.0	10.0			± 2.0	10.3	10.3			± 2.0	10.0	10.0			± 2.0	
NEUT		#	1.36	1.36			± 0.35	4.12	4.12			± 0.90	12.09	12.09			± 1.90	
		%	50.5	50.5			± 10.0	54.9	54.9			± 10.0	68.3	68.3			± 10.0	
LYMPHO		#	0.80	0.80			± 0.33	2.33	2.33			± 0.70	2.64	2.64			± 1.50	
		%	29.7	29.7			± 12.0	31.1	31.1			± 8.0	14.9	14.9			± 8.0	
MONO		#	0.15	0.15			± 0.15	0.52	0.52			± 0.52	1.29	1.29			± 1.29	
		%	5.6	5.6			± 5.6	6.9	6.9			± 6.9	7.3	7.3			± 7.3	
EOS		#	0.30	0.30			± 0.30	0.29	0.29			± 0.29	0.92	0.92			± 0.92	
		%	11.1	11.1			± 11.1	3.8	3.8			± 3.8	5.2	5.2			± 5.2	
BASO		#	0.08	0.08			± 0.08	0.25	0.25			± 0.25	0.76	0.76			± 0.76	
		%	3.1	3.1			± 3.1	3.3	3.3			± 3.3	4.3	4.3			± 4.3	

LOT PX 429
Rev 2

CONTROL

 (Exp.) 2021-07-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCRP		60 60C+ ES60		MS60	MSCRP		60 60C+ ES60	MS60		MSCRP		60 60C+ ES60	MS60	MSCRP	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.6	2.6		± 0.4	7.5	7.4	7.3		± 1.0	18.3	17.8	17.6		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.17	2.13	2.14		± 0.16	4.58	4.55	4.58		± 0.20	5.15	5.08	5.16		± 0.25			
HB HGB	g/dl	6.0	5.9	6.0		± 0.4	13.9	13.7	13.8		± 0.5	16.8	16.5	16.7		± 0.6			
	g/l	60	59	60		± 4	139	137	138		± 5	168	165	167		± 6			
HT HCT	mmol/l	3.73	3.66	3.73		± 0.25	8.63	8.51	8.57		± 0.31	10.43	10.25	10.37		± 0.37			
	%	17.1	16.6	16.9		± 1.5	38.0	37.3	38.0		± 2.0	45.8	44.7	44.9		± 2.5			
VGM MCV	l/l	0.171	0.166	0.169		± 0.015	0.380	0.373	0.380		± 0.020	0.458	0.447	0.449		± 0.025			
	µm ³ /fl	79	78	79		± 5	83	82	83		± 5	89	88	87		± 5			
TGMH MCH	pg	27.6	27.7	28.0		± 2.0	30.3	30.1	30.1		± 2.0	32.6	32.5	32.4		± 2.5			
	fmol	1.72	1.72	1.74		± 0.12	1.88	1.87	1.87		± 0.12	2.03	2.02	2.01		± 0.16			
CCMH MCHC	g/dl	35.0	35.5	35.5		± 3.0	36.6	36.7	36.3		± 3.0	36.7	36.9	37.2		± 3.0			
	g/l	350	355	355		± 30	366	367	363		± 30	367	369	372		± 30			
IDR RDW	mmol/l	21.73	22.05	22.04		± 1.86	22.71	22.80	22.54		± 1.86	22.76	22.92	23.10		± 1.86			
	%	13.5	12.5	13.5		± 4.0	12.5	12.5	12.5		± 4.0	12.0	11.5	11.5		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	69	70		± 20	247	239	240		± 30	494	489	482		± 50			
VPM MPV	µm ³ /fl	9.4	9.6	9.0		± 2.0	9.5	9.8	9.1		± 2.0	9.1	9.3	8.7		± 2.0			
NEUT	#	1.35	1.29	1.43		± 0.35	4.04	4.00	4.12		± 0.90	12.70	12.28	12.58		± 1.90			
	%	50.1	49.8	55.0		± 10.0	53.8	54.0	56.5		± 10.0	69.4	69.0	71.5		± 10.0			
LYMPHO	#	0.82	0.82	0.74		± 0.33	2.41	2.52	2.34		± 0.70	2.75	2.83	2.55		± 1.50			
	%	30.4	31.4	28.5		± 12.0	32.1	34.0	32.0		± 8.0	15.0	15.9	14.5		± 8.0			
MONO	#	0.18	0.14	0.13		± 0.13	0.56	0.38	0.37		± 0.37	1.15	1.07	0.88		± 0.88			
	%	6.6	5.2	5.0		± 5.0	7.5	5.1	5.0		± 5.0	6.3	6.0	5.0		± 5.0			
EOS	#	0.27	0.27	0.21		± 0.21	0.26	0.27	0.22		± 0.22	0.93	0.89	0.79		± 0.79			
	%	9.9	10.5	8.0		± 8.0	3.4	3.7	3.0		± 3.0	5.1	5.0	4.5		± 4.5			
BASO	#	0.08	0.08	0.09		± 0.08	0.24	0.24	0.26		± 0.24	0.77	0.73	0.79		± 0.73			
	%	3.0	3.1	3.5		± 3.0	3.2	3.2	3.5		± 3.2	4.2	4.1	4.5		± 4.1			

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.44