


LOT PX 427
Rev 2


CONTROL

 (Exp.) 2021-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio															
		CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA					
		80 XL80	XLR			80 XL80	XLR			80 XL80	XLR			80 XL80	XLR		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7			± 0.4	7.8	7.8			± 1.0	17.8	17.8			± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.29	2.29			± 0.16	4.50	4.50			± 0.20	5.11	5.11			± 0.25	
HB HGB	g/dl	6.6	6.6			± 0.4	13.7	13.7			± 0.5	16.8	16.8			± 0.6	
	g/l	66	66			± 4	137	137			± 5	168	168			± 6	
	mmol/l	4.10	4.10			± 0.25	8.51	8.51			± 0.31	10.43	10.43			± 0.37	
HT HCT	%	19.2	19.2			± 1.5	37.8	37.8			± 2.0	46.5	46.5			± 2.5	
	l/l	0.192	0.192			± 0.015	0.378	0.378			± 0.020	0.465	0.465			± 0.025	
VGM MCV	µm ³ fl	84	84			± 5	84	84			± 5	91	91			± 5	
TGMH MCH	pg	28.8	28.8			± 2.0	30.4	30.4			± 2.0	32.9	32.9			± 2.5	
	fmol	1.79	1.79			± 0.12	1.89	1.89			± 0.12	2.04	2.04			± 0.16	
CCMH MCHC	g/dl	34.3	34.3			± 3.0	36.2	36.2			± 3.0	36.1	36.1			± 3.0	
	g/l	343	343			± 30	362	362			± 30	361	361			± 30	
	mmol/l	21.31	21.31			± 1.86	22.51	22.51			± 1.86	22.44	22.44			± 1.86	
IDR RDW	%	13.7	13.3			± 4.0	14.0	13.1			± 4.0	13.0	12.5			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	68	68			± 20	252	252			± 30	490	490			± 50	
VPM MPV	µm ³ fl	9.3	9.3			± 2.0	9.1	9.1			± 2.0	9.5	9.5			± 2.0	
NEUT	#	1.41	1.41			± 0.35	4.02	4.02			± 0.90	11.71	11.71			± 1.90	
	%	52.3	52.3			± 10.0	51.5	51.5			± 10.0	65.8	65.8			± 10.0	
LYMPHO	#	0.77	0.77			± 0.33	2.36	2.36			± 0.70	2.74	2.74			± 1.50	
	%	28.7	28.7			± 12.0	30.2	30.2			± 8.0	15.4	15.4			± 8.0	
MONO	#	0.23	0.23			± 0.23	0.76	0.76			± 0.76	1.66	1.66			± 1.66	
	%	8.4	8.4			± 8.4	9.7	9.7			± 9.7	9.3	9.3			± 9.3	
EOS	#	0.19	0.19			± 0.19	0.37	0.37			± 0.37	0.87	0.87			± 0.87	
	%	7.2	7.2			± 7.2	4.8	4.8			± 4.8	4.9	4.9			± 4.9	
BASO	#	0.09	0.09			± 0.09	0.30	0.30			± 0.30	0.82	0.82			± 0.82	
	%	3.4	3.4			± 3.4	3.8	3.8			± 3.8	4.6	4.6			± 4.6	

LOT PX 427
Rev 2

CONTROL

 (Exp.) 2021-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			60 60C+ ES60	MS60	MSCRP	60 60C+ ES60	MS60	MSCRP	60 60C+ ES60	MS60	MSCRP	60 60C+ ES60	MS60	MSCRP						
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.7			± 0.4	7.7	7.8	7.8			± 1.0	18.5	18.2	18.4			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.34	2.31	2.33			± 0.16	4.51	4.50	4.51			± 0.20	5.08	5.04	5.14			± 0.25
		g/dl	6.6	6.5	6.6			± 0.4	13.6	13.4	13.5			± 0.5	16.7	16.5	16.6			± 0.6
HB	HGB	g/l	66	65	66			± 4	136	134	135			± 5	167	165	166			± 6
		mmol/l	4.10	4.04	4.10			± 0.25	8.45	8.32	8.38			± 0.31	10.37	10.25	10.31			± 0.37
HT	HCT	%	18.7	18.5	18.5			± 1.5	37.4	36.9	37.0			± 2.0	45.7	44.9	45.2			± 2.5
		l/l	0.187	0.185	0.185			± 0.015	0.374	0.369	0.370			± 0.020	0.457	0.449	0.452			± 0.025
VGM	MCV	µm ³ ; fl	80	80	79.5			± 5	83	82	82.0			± 5	90	89	88.0			± 5
TGMH	MCH	pg	28.2	28.1	28.3			± 2.0	30.2	29.8	29.9			± 2.0	32.9	32.7	32.3			± 2.5
		fmol	1.75	1.75	1.76			± 0.12	1.87	1.85	1.86			± 0.12	2.04	2.03	2.01			± 0.16
		g/dl	35.3	35.2	35.6			± 3.0	36.3	36.3	36.5			± 3.0	36.5	36.8	36.7			± 3.0
CCMH	MCHC	g/l	353	352	356			± 30	363	363	365			± 30	365	368	367			± 30
		mmol/l	21.89	21.84	22.13			± 1.86	22.56	22.55	22.67			± 1.86	22.68	22.84	22.79			± 1.86
IDR	RDW	%	13.5	12.5	13.0			± 4.0	12.5	12.5	12.8			± 4.0	12.0	12.0	12.2			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	70	71			± 20	250	251	252			± 30	491	490	489			± 50
VPM	MPV	µm ³ ; fl	8.5	8.6	8.4			± 2.0	8.5	8.3	8.1			± 2.0	9.0	8.8	8.5			± 2.0
		#	1.46	1.42	1.51			± 0.35	4.27	4.20	4.33			± 0.90	12.95	12.38	13.06			± 1.90
		%	56.2	54.6	56.0			± 10.0	55.5	53.9	55.5			± 10.0	70.0	68.0	71.0			± 10.0
		#	0.77	0.79	0.78			± 0.33	2.51	2.50	2.46			± 0.70	2.96	2.84	2.76			± 1.50
		%	29.8	30.5	29.0			± 12.0	32.6	32.1	31.5			± 8.0	16.0	15.6	15.0			± 8.0
		#	0.11	0.14	0.14			± 0.11	0.32	0.51	0.39			± 0.32	0.81	1.22	0.92			± 0.81
		%	4.2	5.3	5.0			± 4.2	4.2	6.6	5.0			± 4.2	4.4	6.7	5.0			± 4.4
		#	0.16	0.16	0.16			± 0.16	0.29	0.29	0.31			± 0.29	0.96	0.95	0.92			± 0.92
		%	6.3	6.1	6.0			± 6.0	3.8	3.7	4.0			± 3.7	5.2	5.2	5.0			± 5.0
		#	0.09	0.09	0.11			± 0.09	0.30	0.29	0.31			± 0.29	0.81	0.82	0.74			± 0.74
		%	3.5	3.5	4.0			± 3.5	3.9	3.7	4.0			± 3.7	4.4	4.5	4.0			± 4.0

1300040596
FRONT / RECTO
Ref: TEMP-0821 Rev.44