


LOT PX 414
Rev 2


CONTROL

 (Exp.) 2019-01-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		80 XL80	XLR					80 XL80	XLR					80 XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4			± 0.4	7.6	7.6			± 1.0	17.5	17.5			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.37	2.37			± 0.16	4.65	4.65			± 0.20	5.15	5.15			± 0.25			
HB HGB	g/dl	7.0	7.0			± 0.4	13.9	13.9			± 0.5	16.3	16.3			± 0.6			
	g/l	70	70			± 4	139	139			± 5	163	163			± 6			
	mmol/l	4.35	4.35			± 0.25	8.63	8.63			± 0.31	10.12	10.12			± 0.37			
HT HCT	%	20.1	20.1			± 1.5	38.1	38.1			± 2.0	44.8	44.8			± 2.5			
	l/l	0.201	0.201			± 0.015	0.381	0.381			± 0.020	0.448	0.448			± 0.025			
VGM MCV	µm ³ fl	85	85			± 5	82	82			± 5	87	87			± 5			
TGMH MCH	pg	29.5	29.5			± 2.0	29.9	29.9			± 2.0	31.7	31.7			± 2.5			
	fmol	1.83	1.83			± 0.12	1.86	1.86			± 0.12	1.97	1.97			± 0.16			
CCMH MCHC	g/dl	34.7	34.7			± 3.0	36.5	36.5			± 3.0	36.4	36.4			± 3.0			
	g/l	347	347			± 30	365	365			± 30	364	364			± 30			
	mmol/l	21.58	21.58			± 1.86	22.64	22.64			± 1.86	22.59	22.59			± 1.86			
IDR RDW	%	13.5	12.5			± 4.0	13.0	12.0			± 4.0	13.0	12.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	68	68			± 20	247	247			± 30	497	497			± 50			
VPM MPV	µm ³ fl	10.5	10.5			± 2.0	10.6	10.6			± 2.0	10.1	10.1			± 2.0			
NEUT	#	1.28	1.28			± 0.35	4.15	4.15			± 0.90	12.04	12.04			± 1.90			
	%	53.5	53.5			± 10.0	54.6	54.6			± 10.0	68.8	68.8			± 10.0			
LYMPHO	#	0.74	0.74			± 0.33	2.52	2.52			± 0.70	2.78	2.78			± 1.50			
	%	30.7	30.7			± 12.0	33.1	33.1			± 8.0	15.9	15.9			± 8.0			
MONO	#	0.11	0.11			± 0.11	0.31	0.31			± 0.31	0.68	0.68			± 0.68			
	%	4.4	4.4			± 4.4	4.1	4.1			± 4.1	3.9	3.9			± 3.9			
EOS	#	0.19	0.19			± 0.19	0.34	0.34			± 0.34	1.19	1.19			± 1.19			
	%	8.0	8.0			± 8.0	4.5	4.5			± 4.5	6.8	6.8			± 6.8			
BASO	#	0.08	0.08			± 0.08	0.28	0.28			± 0.28	0.81	0.81			± 0.81			
	%	3.4	3.4			± 3.4	3.7	3.7			± 3.7	4.6	4.6			± 4.6			

LOT PX 414
Rev 2

CONTROL

 (Exp.) 2019-01-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.5		± 0.4	7.5	7.5	7.8			± 1.0	17.7	17.4	18.3		± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.42	2.40	2.42		± 0.16	4.65	4.66	4.70			± 0.20	5.13	5.11	5.17		± 0.25		
	g/dl	7.0	7.0	7.0		± 0.4	13.9	13.8	13.8			± 0.5	16.2	16.1	16.2		± 0.6		
HB HGB	g/l	70	70	70		± 4	139	138	138			± 5	162	161	162		± 6		
	mmol/l	4.35	4.35	4.35		± 0.25	8.63	8.57	8.57			± 0.31	10.06	10.00	10.06		± 0.37		
HT HCT	%	19.8	19.7	19.6		± 1.5	38.1	37.7	37.8			± 2.0	44.6	43.9	44.2		± 2.5		
	l/l	0.198	0.197	0.196		± 0.015	0.381	0.377	0.378			± 0.020	0.446	0.439	0.442		± 0.025		
VGM MCV	µm ³ ·fl	82	82	81.0		± 5	82	81	80.5			± 5	87	86	85.5		± 5		
TGMH MCH	pg	28.9	29.2	28.9		± 2.0	29.9	29.6	29.4			± 2.0	31.6	31.5	31.3		± 2.5		
	fmol	1.80	1.81	1.80		± 0.12	1.86	1.84	1.82			± 0.12	1.96	1.96	1.95		± 0.16		
	g/dl	35.3	35.6	35.7		± 3.0	36.5	36.6	36.5			± 3.0	36.3	36.6	36.6		± 3.0		
CCMH MCHC	g/l	353	356	357		± 30	365	366	365			± 30	363	366	366		± 30		
	mmol/l	21.91	22.09	22.18		± 1.86	22.64	22.70	22.65			± 1.86	22.54	22.75	22.76		± 1.86		
IDR RDW	%	13.5	13.5	12.6		± 4.0	12.8	13.0	12.2			± 4.0	12.5	12.4	12.4		± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	74	70		± 20	250	255	248			± 30	500	505	495		± 50		
VPM MPV	µm ³ ·fl	10.4	11.5	9.6		± 2.0	10.3	10.4	9.5			± 2.0	9.8	9.8	9.0		± 2.0		
NEUT	#	1.33	1.34	1.40		± 0.35	4.24	4.22	4.46			± 0.90	12.43	12.23	13.14		± 1.90		
	%	55.5	55.7	56.0		± 10.0	56.5	56.2	57.2			± 10.0	70.2	70.3	71.8		± 10.0		
LYMPHO	#	0.72	0.73	0.75		± 0.33	2.42	2.45	2.50			± 0.70	2.74	2.66	2.75		± 1.50		
	%	30.0	30.5	30.0		± 12.0	32.2	32.6	32.1			± 8.0	15.5	15.3	15.0		± 8.0		
MONO	#	0.07	0.07	0.10		± 0.07	0.23	0.23	0.23			± 0.23	0.55	0.52	0.48		± 0.48		
	%	3.1	2.8	3.9		± 2.8	3.1	3.0	3.0			± 3.0	3.1	3.0	2.6		± 2.6		
EOS	#	0.19	0.18	0.17		± 0.17	0.33	0.35	0.32			± 0.32	1.17	1.20	1.15		± 1.15		
	%	7.9	7.6	6.6		± 6.6	4.4	4.6	4.1			± 4.1	6.6	6.9	6.3		± 6.3		
BASO	#	0.08	0.08	0.09		± 0.08	0.29	0.27	0.28			± 0.27	0.81	0.78	0.79		± 0.78		
	%	3.5	3.4	3.5		± 3.4	3.8	3.6	3.6			± 3.6	4.6	4.5	4.3		± 4.3		

1300040596 FRONT / RECTO Ref: TEMP-0821 Rev.42