


**LOT** PX 435  
Rev 1


**CONTROL**

 (Exp.) **2022-07-05**  
(YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XL9					XL80	XL9					XL80	XL9				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.1	3.1			± 0.4	8.6	8.6			± 1.0	17.9	17.9			± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.39	2.39			± 0.16	4.53	4.53			± 0.20	4.97	4.97			± 0.25			
HB HGB	g/dl	6.4	6.4			± 0.4	13.6	13.6			± 0.5	16.0	16.0			± 0.6			
	g/l	64	64			± 4	136	136			± 5	160	160			± 6			
HT HCT	mmol/l	3.97	3.97			± 0.25	8.45	8.45			± 0.31	9.94	9.94			± 0.37			
	%	19.6	19.6			± 1.5	39.9	39.9			± 2.0	46.7	46.7			± 2.5			
VGM MCV	l/l	0.196	0.196			± 0.015	0.399	0.399			± 0.020	0.467	0.467			± 0.025			
	µm <sup>3</sup> /fl	82	82			± 5	88	88			± 5	94	94			± 5			
TGMH MCH	pg	26.8	26.8			± 2.0	30.0	30.0			± 2.0	32.2	32.2			± 2.5			
	fmol	1.66	1.66			± 0.12	1.86	1.86			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	32.7	32.7			± 3.0	34.1	34.1			± 3.0	34.2	34.2			± 3.0			
	g/l	327	327			± 30	341	341			± 30	342	342			± 30			
IDR RDW	mmol/l	20.28	20.28			± 1.86	21.19	21.19			± 1.86	21.27	21.27			± 1.86			
	%	15.0	15.0			± 5.0	13.0	13.0			± 5.0	12.0	12.0			± 5.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	74	74			± 20	221	221			± 30	457	457			± 50			
VPM MPV	µm <sup>3</sup> /fl	10.1	10.1			± 2.0	10.9	10.9			± 2.0	10.4	10.4			± 2.0			
NEUT	#	1.56	1.56			± 0.35	4.45	4.45			± 0.90	12.55	12.55			± 1.90			
	%	50.4	50.4			± 10.0	51.8	51.8			± 10.0	70.1	70.1			± 10.0			
LYMPHO	#	1.10	1.10			± 0.33	3.27	3.27			± 0.70	3.40	3.40			± 1.50			
	%	35.4	35.4			± 12.0	38.0	38.0			± 8.0	19.0	19.0			± 8.0			
MONO	#	0.20	0.20			± 0.20	0.32	0.32			± 0.32	0.38	0.38			± 0.38			
	%	6.5	6.5			± 6.5	3.7	3.7			± 3.7	2.1	2.1			± 2.1			
EOS	#	0.16	0.16			± 0.16	0.31	0.31			± 0.31	0.86	0.86			± 0.86			
	%	5.0	5.0			± 5.0	3.6	3.6			± 3.6	4.8	4.8			± 4.8			
BASO	#	0.08	0.08			± 0.08	0.25	0.25			± 0.25	0.72	0.72			± 0.72			
	%	2.7	2.7			± 2.7	2.9	2.9			± 2.9	4.0	4.0			± 4.0			

**LOT** PX 435  
Rev 1

**CONTROL**

 (Exp.) **2022-07-05**  
(YYYY-MM-DD)

		ABX Lysebio													
CONTROL				L	CONTROL				N	CONTROL				H	
PENTRA				PENTRA				PENTRA							

PARAMETRES PARAMETERS	UNITES UNITS	60 60C+ ES60	MS60	MSCRIP	TOLERANCES TOLERANCE	60 60C+ ES60	MS60	MSCRIP	TOLERANCES TOLERANCE	60 60C+ ES60	MS60	MSCRIP	TOLERANCES TOLERANCE
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>3.1</b>	<b>3.1</b>	<b>3.1</b>	± 0.4	<b>8.8</b>	<b>8.7</b>	<b>8.4</b>	± 1.0	<b>18.5</b>	<b>18.0</b>	<b>18.0</b>	± 2.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.47</b>	<b>2.45</b>	<b>2.41</b>	± 0.16	<b>4.55</b>	<b>4.55</b>	<b>4.50</b>	± 0.20	<b>5.01</b>	<b>4.96</b>	<b>4.97</b>	± 0.25
HB HGB	g/dl	<b>6.4</b>	<b>6.3</b>	<b>6.4</b>	± 0.4	<b>13.6</b>	<b>13.5</b>	<b>13.6</b>	± 0.5	<b>16.0</b>	<b>15.7</b>	<b>15.9</b>	± 0.6
	g/l	<b>64</b>	<b>63</b>	<b>64</b>	± 4	<b>136</b>	<b>135</b>	<b>136</b>	± 5	<b>160</b>	<b>157</b>	<b>159</b>	± 6
HT HCT	mmol/l	<b>3.97</b>	<b>3.91</b>	<b>3.97</b>	± 0.25	<b>8.45</b>	<b>8.38</b>	<b>8.45</b>	± 0.31	<b>9.94</b>	<b>9.75</b>	<b>9.87</b>	± 0.37
	%	<b>19.5</b>	<b>19.1</b>	<b>18.8</b>	± 1.5	<b>39.6</b>	<b>39.1</b>	<b>38.7</b>	± 2.0	<b>46.6</b>	<b>46.1</b>	<b>46.2</b>	± 2.5
VGM MCV	l/l	<b>0.195</b>	<b>0.191</b>	<b>0.188</b>	± 0.015	<b>0.396</b>	<b>0.391</b>	<b>0.387</b>	± 0.020	<b>0.466</b>	<b>0.461</b>	<b>0.462</b>	± 0.025
	µm <sup>3</sup> ; fl	<b>79</b>	<b>78</b>	<b>78</b>	± 5	<b>87</b>	<b>86</b>	<b>86</b>	± 5	<b>93</b>	<b>93</b>	<b>93</b>	± 5
TGMH MCH	pg	<b>25.9</b>	<b>25.7</b>	<b>26.6</b>	± 2.0	<b>29.9</b>	<b>29.7</b>	<b>30.2</b>	± 2.0	<b>31.9</b>	<b>31.7</b>	<b>32.0</b>	± 2.5
	fmol	<b>1.61</b>	<b>1.60</b>	<b>1.65</b>	± 0.12	<b>1.86</b>	<b>1.84</b>	<b>1.88</b>	± 0.12	<b>1.98</b>	<b>1.97</b>	<b>1.99</b>	± 0.16
CCMH MCHC	g/dl	<b>32.8</b>	<b>33.0</b>	<b>34.0</b>	± 3.0	<b>34.4</b>	<b>34.5</b>	<b>35.1</b>	± 3.0	<b>34.3</b>	<b>34.0</b>	<b>34.4</b>	± 3.0
	g/l	<b>328</b>	<b>330</b>	<b>340</b>	± 30	<b>344</b>	<b>345</b>	<b>351</b>	± 30	<b>343</b>	<b>340</b>	<b>344</b>	± 30
IDR RDW	mmol/l	<b>20.37</b>	<b>20.47</b>	<b>21.14</b>	± 1.86	<b>21.34</b>	<b>21.42</b>	<b>21.82</b>	± 1.86	<b>21.33</b>	<b>21.14</b>	<b>21.36</b>	± 1.86
	%	<b>15.5</b>	<b>15.5</b>	<b>14.0</b>	± 4.0	<b>12.5</b>	<b>12.5</b>	<b>11.5</b>	± 4.0	<b>12.0</b>	<b>12.0</b>	<b>11.0</b>	± 4.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>78</b>	<b>76</b>	<b>75</b>	± 20	<b>223</b>	<b>225</b>	<b>219</b>	± 30	<b>455</b>	<b>465</b>	<b>445</b>	± 50
VPM MPV	µm <sup>3</sup> ; fl	<b>9.8</b>	<b>10.0</b>	<b>9.0</b>	± 2.0	<b>10.2</b>	<b>10.5</b>	<b>9.5</b>	± 2.0	<b>9.7</b>	<b>10.0</b>	<b>9.0</b>	± 2.0
NEUT	#	<b>1.53</b>	<b>1.55</b>	<b>1.52</b>	± 0.35	<b>4.44</b>	<b>4.49</b>	<b>4.33</b>	± 0.90	<b>12.84</b>	<b>12.62</b>	<b>12.60</b>	± 1.90
	%	<b>49.2</b>	<b>50.1</b>	<b>49.0</b>	± 10.0	<b>50.4</b>	<b>51.6</b>	<b>51.5</b>	± 10.0	<b>69.4</b>	<b>70.1</b>	<b>70.0</b>	± 10.0
LYMPHO	#	<b>1.13</b>	<b>1.09</b>	<b>1.13</b>	± 0.33	<b>3.48</b>	<b>3.31</b>	<b>3.28</b>	± 0.70	<b>3.64</b>	<b>3.37</b>	<b>3.60</b>	± 1.50
	%	<b>36.3</b>	<b>35.1</b>	<b>36.5</b>	± 12.0	<b>39.5</b>	<b>38.0</b>	<b>39.0</b>	± 8.0	<b>19.7</b>	<b>18.7</b>	<b>20.0</b>	± 8.0
MONO	#	<b>0.21</b>	<b>0.21</b>	<b>0.20</b>	± 0.20	<b>0.34</b>	<b>0.34</b>	<b>0.29</b>	± 0.29	<b>0.41</b>	<b>0.38</b>	<b>0.27</b>	± 0.27
	%	<b>6.7</b>	<b>6.8</b>	<b>6.5</b>	± 6.5	<b>3.9</b>	<b>3.9</b>	<b>3.5</b>	± 3.5	<b>2.2</b>	<b>2.1</b>	<b>1.5</b>	± 1.5
EOS	#	<b>0.16</b>	<b>0.16</b>	<b>0.16</b>	± 0.16	<b>0.29</b>	<b>0.31</b>	<b>0.25</b>	± 0.25	<b>0.87</b>	<b>0.90</b>	<b>0.81</b>	± 0.81
	%	<b>5.1</b>	<b>5.3</b>	<b>5.0</b>	± 5.0	<b>3.3</b>	<b>3.6</b>	<b>3.0</b>	± 3.0	<b>4.7</b>	<b>5.0</b>	<b>4.5</b>	± 4.5
BASO	#	<b>0.08</b>	<b>0.08</b>	<b>0.09</b>	± 0.08	<b>0.26</b>	<b>0.25</b>	<b>0.25</b>	± 0.25	<b>0.74</b>	<b>0.74</b>	<b>0.72</b>	± 0.72
	%	<b>2.7</b>	<b>2.7</b>	<b>3.0</b>	± 2.7	<b>2.9</b>	<b>2.9</b>	<b>3.0</b>	± 2.9	<b>4.0</b>	<b>4.1</b>	<b>4.0</b>	± 4.0

Ref: TEMP-0821 Rev.45 FRONT / RECTO 1300090306