


LOT PX 415
Rev 2

CONTROL

 (Exp.) **2019-03-05**
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE				
			CONTROL			L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE		
			PENTRA				PENTRA						PENTRA								
			80 XL80	XLR		80 XL80	XLR				80 XL80	XLR				80 XL80		XLR			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4				± 0.4	7.4	7.4				± 1.0	17.5	17.5				± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.37				± 0.16	4.58	4.58				± 0.20	5.13	5.13				± 0.25	
HB	HGB	g/dl	6.9	6.9				± 0.4	13.7	13.7				± 0.5	16.0	16.0				± 0.6	
		g/l	69	69					± 4	137	137				± 5	160	160				± 6
		mmol/l	4.28	4.28					± 0.25	8.51	8.51				± 0.31	9.94	9.94				± 0.37
HT	HCT	%	19.9	19.9				± 1.5	37.6	37.6				± 2.0	44.6	44.6				± 2.5	
		l/l	0.199	0.199					± 0.015	0.376	0.376				± 0.020	0.446	0.446				± 0.025
VGM	MCV	µm ³ ; fl	84	84				± 5	82	82				± 5	87	87				± 5	
TGMH	MCH	pg	29.1	29.1				± 2.0	29.9	29.9				± 2.0	31.2	31.2				± 2.5	
		fmol	1.81	1.81					± 0.12	1.86	1.86				± 0.12	1.94	1.94				± 0.16
CCMH	MCHC	g/dl	34.7	34.7				± 3.0	36.5	36.5				± 3.0	35.8	35.8				± 3.0	
		g/l	347	347					± 30	365	365				± 30	358	358				± 30
		mmol/l	21.52	21.52					± 1.86	22.65	22.65				± 1.86	22.26	22.26				± 1.86
IDR	RDW	%	13.0	12.0				± 4.0	13.4	12.3				± 4.0	13.4	12.3				± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	65	65				± 20	250	250				± 30	505	505				± 50	
VPM	MPV	µm ³ ; fl	9.2	9.2				± 2.0	9.4	9.4				± 2.0	9.7	9.7				± 2.0	
NEUT		#	1.37	1.37				± 0.35	4.33	4.33				± 0.90	12.62	12.62				± 1.90	
		%	56.9	56.9					± 10.0	58.5	58.5				± 10.0	72.1	72.1				± 10.0
LYMPHO		#	0.67	0.67				± 0.33	2.24	2.24				± 0.70	2.54	2.54				± 1.50	
		%	28.0	28.0					± 12.0	30.3	30.3				± 8.0	14.5	14.5				± 8.0
MONO		#	0.09	0.09				± 0.09	0.27	0.27				± 0.27	0.70	0.70				± 0.70	
		%	3.6	3.6					± 3.6	3.7	3.7				± 3.7	4.0	4.0				± 4.0
EOS		#	0.19	0.19				± 0.19	0.29	0.29				± 0.29	0.86	0.86				± 0.86	
		%	8.1	8.1					± 8.1	3.9	3.9				± 3.9	4.9	4.9				± 4.9
BASO		#	0.08	0.08				± 0.08	0.27	0.27				± 0.27	0.79	0.79				± 0.79	
		%	3.4	3.4					± 3.4	3.6	3.6				± 3.6	4.5	4.5				± 4.5

LOT PX 415
Rev 2

CONTROL

 (Exp.) **2019-03-05**
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE	
			CONTROL			L	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA					PENTRA							
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP						
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.4	2.4		± 0.4	7.5	7.3	7.4			± 1.0	17.8	17.2	17.8			± 2.2
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.44	2.40	2.39		± 0.16	4.62	4.60	4.61			± 0.20	5.13	5.10	5.16			± 0.25
HB	HGB	g/dl	7.0	6.9	6.9		± 0.4	13.7	13.6	13.6			± 0.5	15.9	15.7	15.9			± 0.6
		g/l	70	69	69		± 4	137	136	136			± 5	159	157	159			± 6
HT	HCT	mmol/l	4.35	4.28	4.28		± 0.25	8.51	8.45	8.45			± 0.31	9.87	9.75	9.87			± 0.37
		%	19.8	19.4	19.4		± 1.5	37.4	37.3	37.1			± 2.0	44.6	43.9	43.6			± 2.5
		l/l	0.198	0.194	0.194		± 0.015	0.374	0.373	0.371			± 0.020	0.446	0.439	0.436			± 0.025
VGM	MCV	µm ³ ; fl	81	81	81.0		± 5	81	81	80.5			± 5	87	86	84.5			± 5
TGMH	MCH	pg	28.7	28.8	28.9		± 2.0	29.7	29.6	29.5			± 2.0	31.0	30.8	30.8			± 2.5
		fmol	1.78	1.79	1.79		± 0.12	1.84	1.84	1.83			± 0.12	1.92	1.91	1.91			± 0.16
CCMH	MCHC	g/dl	35.4	35.5	35.6		± 3.0	36.6	36.5	36.6			± 3.0	35.6	35.8	36.5			± 3.0
		g/l	354	355	356		± 30	366	365	366			± 30	356	358	365			± 30
		mmol/l	21.99	22.04	22.13		± 1.86	22.73	22.67	22.76			± 1.86	22.12	22.23	22.65			± 1.86
IDR	RDW	%	13.0	12.8	12.0		± 4.0	12.8	12.8	12.0			± 4.0	12.5	12.5	12.0			± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	65	65		± 20	250	253	245			± 30	500	510	491			± 50
VPM	MPV	µm ³ ; fl	9.1	9.1	8.8		± 2.0	9.3	9.3	8.9			± 2.0	9.5	9.6	9.0			± 2.0
NEUT		#	1.43	1.39	1.40		± 0.35	4.42	4.25	4.43			± 0.90	12.85	12.40	13.30			± 1.90
		%	57.3	57.8	58.2		± 10.0	58.9	58.2	59.9			± 10.0	72.2	72.1	74.7			± 10.0
LYMPHO		#	0.73	0.68	0.67		± 0.33	2.29	2.23	2.23			± 0.70	2.67	2.48	2.40			± 1.50
		%	29.0	28.4	28.0		± 12.0	30.5	30.6	30.1			± 8.0	15.0	14.4	13.5			± 8.0
MONO		#	0.06	0.07	0.07		± 0.06	0.24	0.27	0.22			± 0.22	0.64	0.71	0.53			± 0.53
		%	2.5	2.9	2.9		± 2.5	3.2	3.7	3.0			± 3.0	3.6	4.1	3.0			± 3.0
EOS		#	0.19	0.18	0.18		± 0.18	0.28	0.29	0.26			± 0.26	0.84	0.84	0.80			± 0.80
		%	7.7	7.5	7.4		± 7.4	3.7	4.0	3.5			± 3.5	4.7	4.9	4.5			± 4.5
BASO		#	0.09	0.08	0.08		± 0.08	0.28	0.26	0.26			± 0.26	0.80	0.77	0.77			± 0.77
		%	3.5	3.4	3.5		± 3.4	3.7	3.5	3.5			± 3.5	4.5	4.5	4.3			± 4.3

Ref: TEMP-0821 Rev.42 FRONT / RECTO 1300040596