


LOT PX 405
Rev 1


CONTROL

 (Exp.) 2017-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE	
			CONTROL					L	CONTROL					N	CONTROL						H
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR	60 60C+ ES60	80 XL80	MS60	XLR	MSCR	60 60C+ ES60	80 XL80	MS60	XLR	MSCR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	2.5	2.6	± 0.4	7.5	7.4	7.5	7.4	7.6	± 1.0	18.2	17.9	18.1	17.9	18.2	± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.37	2.31	2.37	2.31	2.36	± 0.16	4.62	4.63	4.64	4.63	4.65	± 0.20	5.14	5.19	5.19	5.19	5.20	± 0.25	
HB	HGB	g/dl	6.9	6.9	6.9	6.9	6.9	± 0.4	13.6	13.6	13.6	13.6	13.5	± 0.5	16.7	16.7	16.8	16.7	16.6	± 0.6	
		g/l	69	69	69	69	69	± 4	136	136	136	136	135	± 5	167	167	168	167	166	± 6	
		mmol/l	4.28	4.28	4.28	4.28	4.28	± 0.25	8.45	8.45	8.45	8.45	8.38	± 0.31	10.37	10.37	10.43	10.37	10.31	± 0.37	
HT	HCT	%	19.2	19.4	19.2	19.4	19.4	± 1.5	37.0	37.5	37.1	37.0	37.7	± 2.0	45.7	46.2	45.7	45.7	45.8	± 2.5	
		l/l	0.192	0.194	0.192	0.194	0.194	± 0.015	0.370	0.375	0.371	0.370	0.377	± 0.020	0.457	0.462	0.457	0.457	0.458	± 0.025	
VGM	MCV	µm ³ ; fl	81	84	81	84	82.0	± 5	80	81	80	80	81.0	± 5	89	89	88	88	88.0	± 5	
TGMH	MCH	pg	29.1	29.9	29.1	29.9	29.2	± 2.0	29.4	29.4	29.3	29.4	29.0	± 2.0	32.5	32.2	32.4	32.2	31.9	± 2.5	
		fmol	1.81	1.85	1.81	1.85	1.82	± 0.12	1.83	1.82	1.82	1.82	1.80	± 0.12	2.02	2.00	2.01	2.00	1.98	± 0.16	
CCMH	MCHC	g/dl	35.9	35.6	35.9	35.6	35.7	± 3.0	36.8	36.3	36.6	36.7	35.8	± 3.0	36.5	36.2	36.8	36.6	36.3	± 3.0	
		g/l	359	356	359	356	357	± 30	368	363	366	367	358	± 30	365	362	368	366	363	± 30	
		mmol/l	22.32	22.08	22.32	22.08	22.14	± 1.86	22.85	22.52	22.75	22.80	22.26	± 1.86	22.67	22.45	22.84	22.71	22.53	± 1.86	
IDR	RDW	%	13.0	12.8	12.7	12.8	12.1	± 4.0	12.6	13.1	12.5	13.1	12.5	± 4.0	12.8	13.1	12.4	13.1	12.6	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	72	75	72	75	± 20	250	248	255	248	250	± 30	497	497	515	497	490	± 50	
VPM	MPV	µm ³ ; fl	9.6	9.8	9.6	9.8	9.4	± 2.0	9.5	9.8	9.6	9.8	9.1	± 2.0	8.6	8.9	8.8	8.9	8.4	± 2.0	
NEUT	#	%	1.40	1.39	1.39	1.39	1.45	± 0.35	4.05	4.01	4.09	4.01	4.03	± 0.90	12.65	12.53	12.85	12.53	13.01	± 1.90	
		%	56.0	55.5	55.7	55.5	55.6	± 10.0	54.0	54.2	54.5	54.2	53.0	± 10.0	69.5	70.0	71.0	70.0	71.5	± 10.0	
LYMPHO	#	%	0.75	0.78	0.79	0.78	0.85	± 0.33	2.51	2.54	2.63	2.54	2.81	± 0.70	2.82	2.94	2.90	2.94	3.06	± 1.50	
		%	30.1	31.3	31.5	31.3	32.7	± 12.0	33.5	34.3	35.0	34.3	37.0	± 8.0	15.5	16.4	16.0	16.4	16.8	± 8.0	
MONO	#	%	0.11	0.10	0.08	0.10	0.09	± 0.08	0.45	0.36	0.30	0.36	0.30	± 0.30	1.00	0.79	0.63	0.79	0.64	± 0.63	
		%	4.2	3.9	3.2	3.9	3.4	± 3.2	6.0	4.9	4.0	4.9	4.0	± 4.0	5.5	4.4	3.5	4.4	3.5	± 3.5	
EOS	#	%	0.16	0.15	0.16	0.15	0.13	± 0.13	0.23	0.24	0.23	0.24	0.23	± 0.23	0.91	0.86	0.91	0.86	0.73	± 0.73	
		%	6.5	6.1	6.3	6.1	5.0	± 5.0	3.1	3.2	3.0	3.2	3.0	± 3.0	5.0	4.8	5.0	4.8	4.0	± 4.0	
BASO	#	%	0.08	0.08	0.08	0.08	0.09	± 0.08	0.26	0.25	0.26	0.25	0.23	± 0.23	0.82	0.79	0.81	0.79	0.76	± 0.76	
		%	3.2	3.2	3.3	3.2	3.3	± 3.2	3.4	3.4	3.5	3.4	3.0	± 3.0	4.5	4.4	4.5	4.4	4.2	± 4.2	

LOT PX 405
Rev 1

CONTROL

 (Exp.) 2017-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA				PENTRA				PENTRA									
			120	DX120	DX NEXUS		120	DX120	DX NEXUS		120	DX120	DX NEXUS							
120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS												
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6			± 0.4	7.7	7.7	7.7			± 1.0	18.8	18.8	18.8			± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.36	2.36	2.36			± 0.16	4.66	4.66	4.66			± 0.20	5.25	5.25	5.25			± 0.25
		g/dl	7.0	7.0	7.0			± 0.4	13.5	13.5	13.5			± 0.5	16.6	16.6	16.6			± 0.6
HB	HGB	g/l	70	70	70			± 4	135	135	135			± 5	166	166	166			± 6
		mmol/l	4.35	4.35	4.35			± 0.25	8.38	8.38	8.38			± 0.31	10.31	10.31	10.31			± 0.37
HT	HCT	%	19.8	19.8	19.8			± 1.5	38.2	38.2	38.2			± 2.0	46.7	46.7	46.7			± 2.5
		l/l	0.198	0.198	0.198			± 0.015	0.382	0.382	0.382			± 0.020	0.467	0.467	0.467			± 0.025
VGM	MCV	µm ³ ; fl	84	84	84			± 5	82	82	82			± 5	89	89	89			± 5
TGMH	MCH	pg	29.7	29.7	29.7			± 2.0	29.0	29.0	29.0			± 2.0	31.6	31.6	31.6			± 2.5
		fmol	1.84	1.84	1.84			± 0.12	1.80	1.80	1.80			± 0.12	1.96	1.96	1.96			± 0.16
		g/dl	35.3	35.3	35.3			± 3.0	35.3	35.3	35.3			± 3.0	35.5	35.5	35.5			± 3.0
CCMH	MCHC	g/l	353	353	353			± 30	353	353	353			± 30	355	355	355			± 30
		mmol/l	21.93	21.93	21.93			± 1.86	21.94	21.94	21.94			± 1.86	22.06	22.06	22.06			± 1.86
IDR	RDW	%	15.0	15.0	15.0			± 4.0	16.0	16.0	16.0			± 4.0	15.0	15.0	15.0			± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	77	77	77			± 20	252	252	252			± 30	495	495	495			± 50
VPM	MPV	µm ³ ; fl	9.8	9.8	9.8			± 2.0	9.6	9.6	9.6			± 2.0	8.7	8.7	8.7			± 2.0
		#	1.52	1.57	1.57			± 0.35	4.22	4.34	4.34			± 0.90	13.40	13.60	13.60			± 1.90
		%	58.4	60.5	60.5			± 10.0	54.8	56.4	56.4			± 10.0	71.4	72.5	72.5			± 10.0
		#	0.75	0.68	0.68			± 0.33	2.53	2.36	2.36			± 0.70	2.80	2.59	2.59			± 1.50
		%	28.8	26.2	26.2			± 12.0	32.8	30.7	30.7			± 8.0	14.9	13.8	13.8			± 8.0
		#	0.12	0.12	0.12			± 0.12	0.49	0.50	0.50			± 0.49	1.17	1.15	1.15			± 1.15
		%	4.5	4.6	4.6			± 4.5	6.3	6.5	6.5			± 6.3	6.2	6.1	6.1			± 6.1
		#	0.15	0.16	0.16			± 0.15	0.24	0.26	0.26			± 0.24	0.94	0.96	0.96			± 0.94
		%	5.8	6.2	6.2			± 5.8	3.1	3.4	3.4			± 3.1	5.0	5.1	5.1			± 5.0
		#	0.07	0.07	0.07			± 0.07	0.23	0.23	0.23			± 0.23	0.47	0.47	0.47			± 0.47
		%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5

Ref: TEMP-0821 Rev.40 FRONT / RECTO 1300032435-A