


LOT PX 434
Rev 2


CONTROL

 (Exp.) 2022-05-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5			± 0.4	7.4	7.4			± 1.0	16.9	16.9			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.27	2.27			± 0.16	4.55	4.55			± 0.20	5.13	5.13			± 0.25			
HB HGB	g/dl	6.4	6.4			± 0.4	14.0	14.0			± 0.5	16.7	16.7			± 0.6			
	g/l	64	64			± 4	140	140			± 5	167	167			± 6			
	mmol/l	3.97	3.97			± 0.25	8.69	8.69			± 0.31	10.37	10.37			± 0.37			
HT HCT	%	18.8	18.8			± 1.5	38.7	38.7			± 2.0	45.7	45.7			± 2.5			
	l/l	0.188	0.188			± 0.015	0.387	0.387			± 0.020	0.457	0.457			± 0.025			
VGM MCV	µm ³ fl	83	83			± 5	85	85			± 5	89	89			± 5			
TGMH MCH	pg	28.2	28.2			± 2.0	30.8	30.8			± 2.0	32.6	32.6			± 2.5			
	fmol	1.75	1.75			± 0.12	1.91	1.91			± 0.12	2.02	2.02			± 0.16			
CCMH MCHC	g/dl	34.0	34.0			± 3.0	36.2	36.2			± 3.0	36.6	36.6			± 3.0			
	g/l	340	340			± 30	362	362			± 30	366	366			± 30			
	mmol/l	21.09	21.09			± 1.86	22.48	22.48			± 1.86	22.71	22.71			± 1.86			
IDR RDW	%	12.5	12.5			± 5.0	13.0	13.0			± 5.0	12.5	12.5			± 5.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67			± 20	253	253			± 30	485	485			± 50			
VPM MPV	µm ³ fl	9.9	9.9			± 2.0	9.7	9.7			± 2.0	9.8	9.8			± 2.0			
NEUT	#	1.39	1.39			± 0.35	4.37	4.37			± 0.90	12.15	12.15			± 1.90			
	%	55.6	55.6			± 10.0	59.0	59.0			± 10.0	71.9	71.9			± 10.0			
LYMPHO	#	0.75	0.75			± 0.33	2.32	2.32			± 0.70	2.70	2.70			± 1.50			
	%	29.9	29.9			± 12.0	31.4	31.4			± 8.0	16.0	16.0			± 8.0			
MONO	#	0.10	0.10			± 0.10	0.16	0.16			± 0.16	0.41	0.41			± 0.41			
	%	3.9	3.9			± 3.9	2.2	2.2			± 2.2	2.4	2.4			± 2.4			
EOS	#	0.18	0.18			± 0.18	0.30	0.30			± 0.30	0.91	0.91			± 0.91			
	%	7.3	7.3			± 7.3	4.0	4.0			± 4.0	5.4	5.4			± 5.4			
BASO	#	0.08	0.08			± 0.08	0.25	0.25			± 0.25	0.73	0.73			± 0.73			
	%	3.3	3.3			± 3.3	3.4	3.4			± 3.4	4.3	4.3			± 4.3			

LOT PX 434
Rev 2

CONTROL

 (Exp.) 2022-05-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.5	7.4	7.3		± 1.0	17.3	17.1	17.1		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.32	2.30	2.31		± 0.16	4.61	4.57	4.65		± 0.20	5.16	5.12	5.09		± 0.25			
	g/dl	6.5	6.4	6.4		± 0.4	14.0	13.7	13.9		± 0.5	16.7	16.3	16.4		± 0.6			
HB HGB	g/l	65	64	64		± 4	140	137	139		± 5	167	163	164		± 6			
	mmol/l	4.04	3.97	3.97		± 0.25	8.69	8.51	8.63		± 0.31	10.37	10.12	10.18		± 0.37			
HT HCT	%	18.6	18.2	18.0		± 1.5	38.7	37.9	38.4		± 2.0	45.9	44.8	45.0		± 2.5			
	l/l	0.186	0.182	0.180		± 0.015	0.387	0.379	0.384		± 0.020	0.459	0.448	0.450		± 0.025			
VGM MCV	µm ³ ·fl	80	79	78.0		± 5	84	83	82.5		± 5	89	88	87.0		± 5			
TGMH MCH	pg	28.0	27.8	27.7		± 2.0	30.4	30.0	29.9		± 2.0	32.4	31.8	32.2		± 2.5			
	fmol	1.74	1.73	1.72		± 0.12	1.89	1.86	1.86		± 0.12	2.01	1.98	2.00		± 0.16			
	g/dl	35.0	35.2	35.5		± 3.0	36.2	36.1	36.2		± 3.0	36.4	36.2	37.0		± 3.0			
CCMH MCHC	g/l	350	352	355		± 30	362	361	362		± 30	364	362	370		± 30			
	mmol/l	21.75	21.87	22.06		± 1.86	22.45	22.43	22.50		± 1.86	22.58	22.47	23.00		± 1.86			
IDR RDW	%	13.0	12.0	12.5		± 4.0	13.0	12.5	12.0		± 4.0	12.0	11.5	12.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	69	68		± 20	255	255	258		± 30	482	492	488		± 50			
VPM MPV	µm ³ ·fl	9.5	9.6	8.9		± 2.0	9.1	9.2	8.7		± 2.0	9.0	9.0	8.7		± 2.0			
NEUT	#	1.31	1.31	1.30		± 0.35	4.28	4.33	4.16		± 0.90	12.23	12.16	12.31		± 1.90			
	%	54.5	54.6	54.0		± 10.0	57.0	58.5	57.0		± 10.0	70.7	71.1	72.0		± 10.0			
LYMPHO	#	0.75	0.74	0.76		± 0.33	2.49	2.30	2.45		± 0.70	2.92	2.75	2.82		± 1.50			
	%	31.3	30.7	31.5		± 12.0	33.2	31.1	33.5		± 8.0	16.9	16.1	16.5		± 8.0			
MONO	#	0.10	0.10	0.13		± 0.10	0.17	0.19	0.22		± 0.17	0.43	0.50	0.43		± 0.43			
	%	4.0	4.0	5.5		± 4.0	2.2	2.5	3.0		± 2.2	2.5	2.9	2.5		± 2.5			
EOS	#	0.17	0.18	0.14		± 0.14	0.31	0.33	0.26		± 0.26	0.97	0.96	0.86		± 0.86			
	%	6.9	7.4	6.0		± 6.0	4.1	4.4	3.5		± 3.5	5.6	5.6	5.0		± 5.0			
BASO	#	0.08	0.08	0.07		± 0.07	0.26	0.26	0.22		± 0.22	0.74	0.74	0.68		± 0.68			
	%	3.3	3.3	3.0		± 3.0	3.5	3.5	3.0		± 3.0	4.3	4.3	4.0		± 4.0			

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.45