

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 092

CONTROL

(Exp.) 2012-11-05
 (YYYY - MM - DD)

Rev 1

		ABX Lysebio														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA				
		60 60C+ ES60	80 XL80	MS60			60 60C+ ES60	80 XL80	MS60			60 60C+ ES60	80 XL80	MS60		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3	2.3		± 0.4	7.5	7.5	7.4		± 1.0	17.6	17.3	17.3		± 2.2
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.48	2.42	2.47		± 0.12	4.58	4.57	4.60		± 0.15	5.17	5.18	5.22		± 0.20
HB HGB	g/dl	6.9	7.0	6.9		± 0.4	13.4	13.4	13.3		± 0.5	16.5	16.4	16.5		± 0.6
	g/l	69	70	69		± 4	134	134	133		± 5	165	164	165		± 6
	mmol/l	4.28	4.35	4.28		± 0.25	8.32	8.32	8.26		± 0.31	10.25	10.18	10.25		± 0.37
HT HCT	%	19.8	19.8	19.3		± 1.5	37.1	37.5	36.3		± 2.0	45.5	45.6	44.9		± 2.5
	l/l	0.198	0.198	0.193		± 0.015	0.371	0.375	0.363		± 0.020	0.455	0.456	0.449		± 0.025
VGM MCV	µm ³ ; fl	80	82	78		± 5.0	81	82	79		± 5.0	88	88	86		± 5.0
TGMH MCH	pg	27.8	28.9	27.9		± 2.0	29.3	29.3	28.9		± 2.0	31.9	31.7	31.6		± 2.5
	f mol	1.73	1.80	1.73		± 0.12	1.82	1.82	1.80		± 0.12	1.98	1.97	1.96		± 0.16
CCMH MCHC	g/dl	34.8	35.3	35.8		± 3.0	36.1	35.8	36.6		± 3.0	36.3	36.0	36.8		± 3.0
	g/l	348	353	358		± 30	361	358	366		± 30	363	360	368		± 30
	mmol/l	21.60	21.91	22.24		± 1.86	22.43	22.21	22.73		± 1.86	22.52	22.34	22.82		± 1.86
IDR RDW	%	12.9	13.5	12.8		± 4.0	12.5	13.4	12.5		± 4.0	12.5	13.4	12.0		± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75	70		± 20	258	258	260		± 30	495	500	505		± 50
VPM MPV	µm ³ ; fl	8.6	8.6	8.6		± 2.0	9.1	9.1	9.1		± 2.0	8.2	8.2	8.1		± 2.0
NEUT	#	1.29	1.31	1.22		± 0.30	4.16	4.13	4.01		± 0.80	12.18	11.76	11.94		± 1.90
	%	56.0	57.0	53.0		± 10.0	55.5	55.0	54.2		± 10.0	69.2	68.0	69.0		± 10.0
LYMPHO	#	0.71	0.68	0.76		± 0.20	2.48	2.48	2.55		± 0.70	3.10	3.11	3.17		± 1.50
	%	31.0	29.5	33.0		± 8.0	33.0	33.0	34.5		± 8.0	17.6	18.0	18.3		± 8.0
MONO	#	0.09	0.09	0.12		± 0.09	0.30	0.30	0.33		± 0.30	0.70	0.70	0.70		± 0.70
	%	4.0	4.0	5.0		± 4.0	4.0	4.0	4.5		± 4.0	4.0	4.0	4.0		± 4.0
EOS	#	0.14	0.14	0.13		± 0.12	0.30	0.34	0.26		± 0.24	0.88	1.04	0.78		± 0.73
	%	6.0	6.0	5.8		± 5.0	4.0	4.5	3.5		± 3.0	5.0	6.0	4.5		± 4.0
BASO	#	0.07	0.08	0.07		± 0.06	0.26	0.26	0.24		± 0.20	0.74	0.78	0.73		± 0.37
	%	3.0	3.5	3.2		± 2.5	3.5	3.5	3.3		± 2.5	4.2	4.5	4.2		± 2.0

Ref: TEMP-0821 Rev.18 BACK / VERSO 9930001-A

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 092
 Rev 1

CONTROL

(Exp.) 2012-11-05
 (YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.9	7.9	7.9		± 1.0	18.3	18.3	18.3		± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.48	2.48	2.48		± 0.12	4.65	4.65	4.65		± 0.15	5.29	5.29	5.29		± 0.20			
HB HGB	g/dl	7.0	7.0	7.0		± 0.4	13.3	13.3	13.3		± 0.5	16.3	16.3	16.3		± 0.6			
	g/l	70	70	70		± 4	133	133	133		± 5	163	163	163		± 6			
HT HCT	mmol/l	4.35	4.35	4.35		± 0.25	8.26	8.26	8.26		± 0.31	10.12	10.12	10.12		± 0.37			
	%	20.3	20.3	20.3		± 1.5	38.1	38.1	38.1		± 2.0	46.6	46.6	46.6		± 2.5			
VGM MCV	l/l	0.203	0.203	0.203		± 0.015	0.381	0.381	0.381		± 0.020	0.466	0.466	0.466		± 0.025			
	µm ³ ·fl	82	82	82		± 5.0	82	82	82		± 5.0	88	88	88		± 5.0			
TGMH MCH	pg	28.2	28.2	28.2		± 2.0	28.6	28.6	28.6		± 2.0	30.8	30.8	30.8		± 2.5			
	f·mol	1.75	1.75	1.75		± 0.12	1.78	1.78	1.78		± 0.12	1.91	1.91	1.91		± 0.16			
CCMH MCHC	g/dl	34.4	34.4	34.4		± 3.0	34.9	34.9	34.9		± 3.0	35.0	35.0	35.0		± 3.0			
	g/l	344	344	344		± 30	349	349	349		± 30	350	350	350		± 30			
	mmol/l	21.38	21.38	21.38		± 1.86	21.66	21.66	21.66		± 1.86	21.74	21.74	21.74		± 1.86			
IDR RDW	%	15.0	15.0	15.0		± 4.0	16.0	16.0	16.0		± 4.0	15.0	15.0	15.0		± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	77	77	77		± 20	265	265	265		± 30	495	495	495		± 50			
VPM MPV	µm ³ ·fl	8.6	8.6	8.6		± 2.0	9.0	9.0	9.0		± 2.0	8.2	8.2	8.2		± 2.0			
NEUT	#	1.36	1.36	1.36		± 0.30	4.46	4.51	4.51		± 0.80	13.10	13.00	13.00		± 1.90			
	%	56.5	56.5	56.5		± 10.0	56.5	57.1	57.1		± 10.0	71.5	71.0	71.0		± 10.0			
LYMPHO	#	0.74	0.70	0.70		± 0.20	2.65	2.46	2.46		± 0.70	3.29	3.02	3.02		± 1.50			
	%	31.0	29.0	29.0		± 8.0	33.5	31.1	31.1		± 8.0	18.0	16.5	16.5		± 8.0			
MONO	#	0.12	0.16	0.16		± 0.09	0.36	0.43	0.43		± 0.30	0.82	1.01	1.01		± 0.70			
	%	5.0	6.5	6.5		± 4.0	4.5	5.5	5.5		± 4.0	4.5	5.5	5.5		± 4.0			
EOS	#	0.12	0.13	0.13		± 0.12	0.24	0.24	0.24		± 0.24	0.73	0.73	0.73		± 0.73			
	%	5.0	5.5	5.5		± 5.0	3.0	3.0	3.0		± 3.0	4.0	4.0	4.0		± 4.0			
BASO	#	0.06	0.06	0.06		± 0.06	0.20	0.26	0.26		± 0.20	0.37	0.55	0.55		± 0.37			
	%	2.5	2.5	2.5		± 2.5	2.5	3.3	3.3		± 2.5	2.0	3.0	3.0		± 2.0			

9930066-A
 BACK / VERSO
 Ref: TEMP-0821 Rev.18