

ABX Diffrol



LOT

PX 032

CONTROL

(Exp.) **2012-05-05**
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																				
			CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA							PENTRA							PENTRA						
			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.4	2.4	2.4	2.3	± 0.4	7.4	7.4	7.8	7.8	7.4	± 1.0	17.7	17.6	18.4	18.4	17.4	± 2.2			
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.43	2.36	2.44	2.44	2.40	± 0.12	4.66	4.62	4.68	4.68	4.64	± 0.15	5.15	5.16	5.26	5.26	5.13	± 0.20			
HB	HGB	g/dl	6.8	6.8	6.9	6.9	6.7	± 0.4	13.7	13.7	13.7	13.7	13.7	± 0.5	16.5	16.5	16.5	16.5	16.5	± 0.6			
		g/l	68	68	69	69	67	± 4	137	137	137	137	137	± 5	165	165	165	165	165	± 6			
		mmol/l	4.22	4.22	4.28	4.28	4.16	± 0.25	8.51	8.51	8.51	8.51	8.51	± 0.31	10.25	10.25	10.25	10.25	10.25	± 0.37			
HT	HCT	%	19.4	19.4	20.0	20.0	18.7	± 1.5	38.2	37.9	38.8	38.8	36.7	± 2.0	45.8	45.9	46.8	46.8	44.1	± 2.5			
		l/l	0.194	0.194	0.200	0.200	0.187	± 0.015	0.382	0.379	0.388	0.388	0.367	± 0.020	0.458	0.459	0.468	0.468	0.441	± 0.025			
VGM	MCV	µm ³ fl	80	82	82	82	78	± 4.0	82	82	83	83	79	± 4.0	89	89	89	89	86	± 4.0			
TGMH	MCH	pg	28.0	28.8	28.3	28.3	27.9	± 2.0	29.4	29.7	29.3	29.3	29.5	± 2.0	32.0	32.0	31.4	31.4	32.2	± 2.5			
		fmol	1.74	1.79	1.76	1.76	1.73	± 0.12	1.83	1.84	1.82	1.82	1.83	± 0.12	1.99	1.99	1.95	1.95	2.00	± 0.16			
CCMH	MCHC	g/dl	35.0	35.1	34.5	34.5	35.8	± 3.0	35.9	36.2	35.3	35.3	37.4	± 3.0	36.0	35.9	35.2	35.2	37.4	± 3.0			
		g/l	350	351	345	345	358	± 30	359	362	353	353	374	± 30	360	359	352	352	374	± 30			
		mmol/l	21.72	21.82	21.42	21.42	22.23	± 1.86	22.26	22.46	21.90	21.90	23.21	± 1.86	22.36	22.31	21.89	21.89	23.23	± 1.86			
IDR	RDW	%	12.8	13.5	14.5	14.5	12.5	± 4.0	13.0	13.2	14.8	14.8	12.7	± 4.0	12.0	13.2	14.5	14.5	12.2	± 4.0			
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	72	72	66	± 20	240	245	245	245	240	± 30	490	500	495	495	500	± 50			
VPM	MPV	µm ³ fl	8.6	8.7	8.9	8.9	8.7	± 2.0	8.1	8.3	8.5	8.5	8.2	± 2.0	8.1	8.2	8.4	8.4	8.1	± 2.0			
NEUT	#	%	1.28	1.36	1.37	1.42	1.25	± 0.30	4.07	4.03	4.37	4.51	4.00	± 0.80	12.60	12.58	13.70	13.80	12.21	± 1.90			
		%	55.8	56.5	57.0	59.0	54.5	± 10.0	55.0	54.5	56.0	57.8	54.0	± 10.0	71.2	71.5	74.0	74.8	70.2	± 10.0			
LYMPHO	#	%	0.75	0.76	0.76	0.70	0.75	± 0.25	2.52	2.55	2.69	2.50	2.55	± 0.70	2.76	2.64	2.82	2.58	2.75	± 1.50			
		%	32.4	31.5	31.5	29.0	32.5	± 8.0	34.0	34.5	34.5	32.0	34.5	± 8.0	15.6	15.0	15.3	14.0	15.8	± 8.0			
MONO	#	%	0.10	0.10	0.10	0.12	0.10	± 0.10	0.31	0.31	0.31	0.33	0.33	± 0.31	0.80	0.75	0.83	0.83	0.87	± 0.75			
		%	4.0	4.0	4.0	5.0	4.5	± 4.0	4.0	4.0	4.0	4.2	4.5	± 4.0	4.5	4.0	4.5	4.5	5.0	± 4.0			
EOS	#	%	0.11	0.12	0.12	0.11	0.12	± 0.11	0.26	0.26	0.23	0.23	0.26	± 0.23	0.74	0.88	0.74	0.77	0.78	± 0.74			
		%	4.8	5.0	5.0	4.5	5.0	± 4.5	3.5	3.5	3.0	3.0	3.5	± 3.0	4.2	5.0	4.0	4.2	4.5	± 4.0			
BASO	#	%	0.07	0.07	0.06	0.06	0.08	± 0.06	0.26	0.26	0.20	0.23	0.26	± 0.20	0.80	0.79	0.40	0.46	0.78	± 0.40			
		%	3.0	3.0	2.5	2.5	3.5	± 2.5	3.5	3.5	2.5	3.0	3.5	± 2.5	4.5	4.5	2.2	2.5	4.5	± 2.2			

Ref: FORM-0821 Rev. 15 BACK / VERSO 9930052-B