

ABX Difftrol



LOT

PX 111

CONTROL



(Exp.) 2012-01-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	
			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.4	2.4	2.4	2.4	± 0.4		7.3	7.3	7.6	7.6	7.5		± 1.0	17.5	17.4	18.2	18.2		
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.45	2.40	2.44	2.44	2.44	± 0.12	4.58	4.55	4.59	4.59	4.55	± 0.15	5.02	5.02	5.09	5.09	5.00	± 0.20		
HB HGB	g/dl	6.7	6.7	6.7	6.7	6.6	± 0.4	13.5	13.5	13.4	13.4	13.4	± 0.5	15.8	15.8	15.6	15.6	15.8	± 0.6		
	g/l	67	67	67	67	66	± 4	135	135	134	134	134	± 5	158	158	156	156	158	± 6		
	mmol/l	4.16	4.16	4.16	4.16	4.10	± 0.25	8.38	8.38	8.32	8.32	8.32	± 0.31	9.81	9.81	9.69	9.69	9.81	± 0.37		
HT HCT	%	19.6	19.9	20.0	20.0	19.0	± 1.5	37.6	37.3	38.6	38.6	36.4	± 2.0	44.7	44.7	45.3	45.3	43.5	± 2.5		
	l/l	0.196	0.199	0.200	0.200	0.190	± 0.015	0.376	0.373	0.386	0.386	0.364	± 0.020	0.447	0.447	0.453	0.453	0.435	± 0.025		
VGM MCV	µm ³ ; fl	80	83	82	82	78	± 5.0	82	82	84	84	80	± 5.0	89	89	89	89	87	± 5.0		
TGMH MCH	pg	27.3	27.9	27.5	27.5	27.0	± 2.0	29.5	29.7	29.2	29.2	29.5	± 2.0	31.5	31.5	30.6	30.6	31.6	± 2.5		
	fmol	1.70	1.73	1.71	1.71	1.68	± 0.12	1.83	1.84	1.81	1.81	1.83	± 0.12	1.95	1.95	1.90	1.90	1.96	± 0.16		
CCMH MCHC	g/dl	34.2	33.6	33.5	33.5	34.7	± 3.0	35.9	36.2	34.8	34.8	36.8	± 3.0	35.4	35.4	34.4	34.4	36.3	± 3.0		
	g/l	342	336	335	335	347	± 30	359	362	348	348	368	± 30	354	354	344	344	363	± 30		
	mmol/l	21.23	20.89	20.80	20.80	21.54	± 1.86	22.32	22.47	21.58	21.58	22.86	± 1.86	21.96	21.96	21.38	21.38	22.56	± 1.86		
IDR RDW	%	12.9	12.7	14.4	14.4	12.8	± 4.0	12.5	13.0	14.9	14.9	12.6	± 4.0	12.2	12.7	13.4	13.4	12.2	± 4.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	78	74	80	80	74	± 20	260	262	260	260	264	± 30	505	505	495	495	506	± 50		
VPM MPV	µm ³ ; fl	8.8	9.0	8.3	8.3	8.8	± 2.0	8.3	8.5	8.3	8.3	8.3	± 2.0	8.0	8.1	8.0	8.0	7.8	± 2.0		
NEUT	#	1.30	1.36	1.38	1.37	1.33	± 0.30	3.80	3.69	4.03	3.95	3.81	± 0.80	11.94	11.66	12.90	12.80	11.96	± 1.90		
	%	56.5	56.5	57.5	57.0	55.6	± 10.0	52.0	50.5	53.0	52.0	50.8	± 10.0	68.2	67.0	70.5	70.0	67.2	± 10.0		
LYMPHO	#	0.69	0.73	0.73	0.72	0.75	± 0.20	2.63	2.74	2.77	2.77	2.81	± 0.70	3.06	3.22	3.46	3.46	3.38	± 1.50		
	%	30.0	30.5	30.5	30.0	31.2	± 8.0	36.0	37.5	36.5	36.5	37.5	± 8.0	17.5	18.5	19.0	19.0	19.0	± 8.0		
MONO	#	0.10	0.10	0.10	0.10	0.10	± 0.10	0.37	0.37	0.34	0.38	0.34	± 0.34	0.91	0.91	0.91	0.91	0.91	± 0.91		
	%	4.0	4.0	4.0	4.0	4.0	± 4.0	5.0	5.0	4.5	5.0	4.5	± 4.5	5.0	5.0	5.0	5.0	5.0	± 5.0		
EOS	#	0.15	0.15	0.13	0.16	0.14	± 0.13	0.29	0.29	0.23	0.27	0.30	± 0.23	0.88	0.96	0.64	0.73	0.80	± 0.64		
	%	6.5	6.4	5.5	6.5	6.0	± 5.5	4.0	4.0	3.0	3.5	4.0	± 3.0	5.0	5.5	3.5	4.0	4.5	± 3.5		
BASO	#	0.07	0.06	0.06	0.06	0.08	± 0.06	0.23	0.23	0.23	0.23	0.24	± 0.23	0.75	0.70	0.36	0.36	0.77	± 0.36		
	%	3.0	2.6	2.5	2.5	3.2	± 2.5	3.0	3.0	3.0	3.0	3.2	± 3.0	4.3	4.0	2.0	2.0	4.3	± 2.0		

9930052-B
BACK / VERSO
Ref: FORM-0821 Rev.14