

**HORIBA ABX SAS**  
 Parc Euromédecine  
 34184 Montpellier cedex 4 - France  
 Tél.: 33 (0) 4 67 14 15 16  
 Fax.: 33 (0) 4 67 14 15 17

**LOT** PX 051

**CONTROL**

(Exp.) 2011-07-05  
 (YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL					L	CONTROL					N	CONTROL					H
			PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE
			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60	
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.6	2.6		± 0.4	7.5	7.4	7.7	7.7		± 1.0	17.7	17.5	18.2	18.2		± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.50	2.42	2.49	2.49		± 0.12	4.66	4.64	4.73	4.73		± 0.15	5.05	5.04	5.17	5.17		± 0.20
HB	HGB	g/dl	6.2	6.1	6.3	6.3		± 0.4	13.9	13.8	13.9	13.9		± 0.5	16.5	16.5	16.6	16.6		± 0.6
		g/l	62	61	63	63		± 4	139	138	139	139		± 5	165	165	166	166		± 6
		mmol/l	3.85	3.79	3.91	3.91		± 0.25	8.63	8.57	8.63	8.63		± 0.31	10.25	10.25	10.31	10.31		± 0.37
HT	HCT	%	18.8	18.9	19.7	19.7		± 1.5	40.5	40.4	41.6	41.6		± 2.0	47.5	47.4	48.6	48.6		± 2.5
		l/l	0.188	0.189	0.197	0.197		± 0.015	0.405	0.404	0.416	0.416		± 0.020	0.475	0.474	0.486	0.486		± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	75	78	79	79		± 4.0	87	87	88	88		± 4.0	94	94	94	94		± 4.0
TGMH	MCH	pg	24.8	25.2	25.3	25.3		± 2.0	29.8	29.7	29.4	29.4		± 2.0	32.7	32.7	32.1	32.1		± 2.5
		fmol	1.54	1.57	1.57	1.57		± 0.12	1.85	1.85	1.82	1.82		± 0.12	2.03	2.03	1.99	1.99		± 0.16
CCMH	MCHC	g/dl	33.1	32.3	32.0	32.0		± 3.0	34.3	34.2	33.4	33.4		± 3.0	34.8	34.8	34.2	34.2		± 3.0
		g/l	331	323	320	320		± 30	343	342	334	334		± 30	348	348	342	342		± 30
		mmol/l	20.53	20.07	19.89	19.89		± 1.86	21.29	21.23	20.74	20.74		± 1.86	21.59	21.63	21.21	21.21		± 1.86
IDR	RDW	%	13.7	13.3	15.3	15.3		± 4.0	11.7	11.8	13.3	13.3		± 4.0	11.2	11.4	13.3	13.3		± 4.0
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	70	73	73		± 20	260	255	260	260		± 30	505	505	505	505		± 50
VPM	MPV	µm <sup>3</sup> ; fl	9.2	9.2	9.3	9.3		± 2.0	8.3	8.5	8.4	8.4		± 2.0	8.3	8.5	8.5	8.5		± 2.0
NEUT	#		1.44	1.44	1.48	1.48		± 0.30	4.67	4.48	4.70	4.70		± 0.80	13.28	13.04	14.20	14.00		± 1.90
		%	57.5	57.5	57.0	57.0		± 10.0	62.3	60.5	61.0	61.0		± 10.0	75.0	74.5	78.1	77.0		± 10.0
LYMPHO	#		0.79	0.79	0.82	0.82		± 0.30	2.19	2.26	2.35	2.25		± 0.70	2.48	2.63	2.66	2.53		± 1.50
		%	31.5	31.5	31.5	31.5		± 8.0	29.2	30.5	30.5	29.2		± 8.0	14.0	15.0	14.6	13.9		± 8.0
MONO	#		0.04	0.04	0.04	0.04		± 0.04	0.12	0.12	0.12	0.12		± 0.12	0.27	0.22	0.22	0.22		± 0.22
		%	1.5	1.5	1.5	1.5		± 1.5	1.5	1.5	1.5	1.5		± 1.5	1.5	1.2	1.2	1.2		± 1.2
EOS	#		0.16	0.16	0.18	0.18		± 0.16	0.27	0.30	0.33	0.42		± 0.27	0.92	0.88	0.80	1.02		± 0.80
		%	6.5	6.5	7.0	7.0		± 6.0	3.6	4.0	4.3	5.5		± 3.6	5.2	5.0	4.1	5.6		± 4.0
BASO	#		0.08	0.08	0.08	0.08		± 0.08	0.26	0.26	0.21	0.22		± 0.20	0.76	0.76	0.40	0.42		± 0.40
		%	3.0	3.0	3.0	3.0		± 3.0	3.4	3.5	2.7	2.8		± 2.5	4.3	4.3	2.0	2.3		± 2.0

Ref: FORM-0621 Rev.11 BACK / VERSO 9930052-A