


LOT PX 056
Rev 4


CONTROL

 (Exp.) 2016-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																		
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		60 60C+ ES60	80 XL80	MS60	XLR	MSCR		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.5	2.4	2.5	2.4	± 0.4	7.6	7.5	7.4	7.5	7.4	± 1.0	17.9	17.5	17.3	17.5	17.6	± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.41	2.35	2.38	2.35	2.36	± 0.16	4.60	4.60	4.58	4.60	4.57	± 0.20	5.08	5.10	5.09	5.10	5.08	± 0.25	
HB	HGB	g/dl	6.8	6.8	6.8	6.8	6.7	± 0.4	13.5	13.6	13.5	13.6	13.5	± 0.5	15.7	15.8	15.8	15.8	15.7	± 0.6	
		g/l	68	68	68	68	67	± 4	135	136	135	136	135	± 5	157	158	158	158	157	± 6	
		mmol/l	4.22	4.22	4.22	4.22	4.16	± 0.25	8.38	8.45	8.38	8.45	8.38	± 0.31	9.75	9.81	9.81	9.81	9.75	± 0.37	
HT	HCT	%	19.3	19.5	19.0	19.3	18.6	± 1.5	37.7	37.7	37.1	37.3	37.5	± 2.0	43.2	43.9	42.8	43.4	43.7	± 2.5	
		l/l	0.193	0.195	0.190	0.193	0.186	± 0.015	0.377	0.377	0.371	0.373	0.375	± 0.020	0.432	0.439	0.428	0.434	0.437	± 0.025	
VGM	MCV	µm ³ ; fl	80	83	80	82	79.0	± 5	82	82	81	81	82.0	± 5	85	86	84	85	86.0	± 5	
TGMH	MCH	pg	28.2	28.9	28.6	28.9	28.4	± 2.0	29.3	29.6	29.5	29.6	29.5	± 2.0	30.9	31.0	31.0	31.0	30.9	± 2.5	
		fmol	1.75	1.80	1.77	1.80	1.76	± 0.12	1.82	1.84	1.83	1.84	1.83	± 0.12	1.92	1.92	1.93	1.92	1.92	± 0.16	
CCMH	MCHC	g/dl	35.3	34.9	35.7	35.3	35.9	± 3.0	35.8	36.1	36.4	36.5	36.0	± 3.0	36.4	36.0	37.0	36.4	35.9	± 3.0	
		g/l	353	349	357	353	359	± 30	358	361	364	365	360	± 30	364	360	370	364	359	± 30	
		mmol/l	21.90	21.65	22.18	21.91	22.32	± 1.86	22.23	22.39	22.60	22.67	22.37	± 1.86	22.58	22.37	22.95	22.63	22.32	± 1.86	
IDR	RDW	%	14.1	14.1	12.8	14.1	12.5	± 4.0	13.5	14.1	12.7	14.1	12.8	± 4.0	13.4	13.3	11.9	13.3	12.0	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	73	70	70	70	68	± 20	250	250	255	250	245	± 30	500	495	515	495	495	± 50	
VPM	MPV	µm ³ ; fl	9.5	9.7	9.7	9.7	9.0	± 2.0	9.4	9.8	9.5	9.8	8.9	± 2.0	9.0	9.3	9.1	9.3	8.4	± 2.0	
NEUT	#		1.55	1.64	1.60	1.64	1.60	± 0.35	4.18	4.07	4.14	4.07	4.20	± 0.90	12.71	12.21	12.56	12.21	12.92	± 1.90	
		%	64.5	65.4	66.5	65.4	66.7	± 10.0	55.0	54.2	55.9	54.2	56.7	± 10.0	71.0	69.8	72.6	69.8	73.4	± 10.0	
LYMPHO	#		0.53	0.52	0.51	0.52	0.49	± 0.33	2.36	2.33	2.36	2.33	2.28	± 0.70	2.36	2.21	2.23	2.21	2.11	± 1.50	
		%	22.0	20.9	21.2	20.9	20.5	± 12.0	31.0	31.0	31.9	31.0	30.8	± 8.0	13.2	12.6	12.9	12.6	12.0	± 8.0	
MONO	#		0.11	0.11	0.08	0.11	0.11	± 0.08	0.44	0.46	0.31	0.46	0.37	± 0.31	0.90	0.93	0.64	0.93	0.70	± 0.64	
		%	4.5	4.6	3.3	4.6	4.5	± 3.3	5.8	6.3	4.2	6.3	5.0	± 4.2	5.0	5.3	3.7	5.3	4.0	± 3.7	
EOS	#		0.13	0.14	0.13	0.14	0.11	± 0.11	0.34	0.38	0.33	0.38	0.30	± 0.30	1.16	1.40	1.12	1.40	1.13	± 1.12	
		%	5.5	5.4	5.3	5.4	4.5	± 4.5	4.5	5.0	4.5	5.0	4.0	± 4.0	6.5	8.0	6.5	8.0	6.4	± 6.4	
BASO	#		0.08	0.09	0.09	0.09	0.09	± 0.08	0.28	0.26	0.26	0.26	0.26	± 0.26	0.77	0.75	0.74	0.75	0.74	± 0.74	
		%	3.5	3.7	3.7	3.7	3.8	± 3.5	3.7	3.5	3.5	3.5	3.5	± 3.5	4.3	4.3	4.3	4.3	4.2	± 4.2	

LOT PX 056
Rev 4

CONTROL

 (Exp.) 2016-07-05
(YYYY - MM - DD)

		ABX Lysebio																							
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE					
		PENTRA						PENTRA						PENTRA											
		120	DX120	DX NEXUS				120	DX120	DX NEXUS				120	DX120	DX NEXUS					120	DX120	DX NEXUS		
		120 RETIC	DF120	DF NEXUS				120 RETIC	DF120	DF NEXUS				120 RETIC	DF120	DF NEXUS					120 RETIC	DF120	DF NEXUS		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5			± 0.4	7.8	7.8	7.8			± 1.0	18.4	18.4	18.4			± 2.2						
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.42	2.42	2.42			± 0.16	4.62	4.62	4.62			± 0.20	5.17	5.17	5.17			± 0.25						
	g/dl	6.9	6.9	6.9			± 0.4	13.6	13.6	13.6			± 0.5	15.7	15.7	15.7			± 0.6						
HB HGB	g/l	69	69	69			± 4	136	136	136			± 5	157	157	157			± 6						
	mmol/l	4.28	4.28	4.28			± 0.25	8.45	8.45	8.45			± 0.31	9.75	9.75	9.75			± 0.37						
HT HCT	%	20.1	20.1	20.1			± 1.5	38.3	38.3	38.3			± 2.0	44.5	44.5	44.5			± 2.5						
	l/l	0.201	0.201	0.201			± 0.015	0.383	0.383	0.383			± 0.020	0.445	0.445	0.445			± 0.025						
VGM MCV	µm ³ ·fl	83	83	83			± 5	83	83	83			± 5	86	86	86			± 5						
TGMH MCH	pg	28.5	28.5	28.5			± 2.0	29.4	29.4	29.4			± 2.0	30.4	30.4	30.4			± 2.5						
	fmol	1.77	1.77	1.77			± 0.12	1.83	1.83	1.83			± 0.12	1.89	1.89	1.89			± 0.16						
	g/dl	34.4	34.4	34.4			± 3.0	35.5	35.5	35.5			± 3.0	35.3	35.3	35.3			± 3.0						
CCMH MCHC	g/l	344	344	344			± 30	355	355	355			± 30	353	353	353			± 30						
	mmol/l	21.33	21.33	21.33			± 1.86	22.02	22.02	22.02			± 1.86	21.93	21.93	21.93			± 1.86						
IDR RDW	%	16.4	16.4	16.4			± 4.0	16.7	16.7	16.7			± 4.0	15.6	15.6	15.6			± 4.0						
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	74	74			± 20	260	260	260			± 30	505	505	505			± 50						
VPM MPV	µm ³ ·fl	9.7	9.7	9.7			± 2.0	9.6	9.6	9.6			± 2.0	9.1	9.1	9.1			± 2.0						
NEUT	#	1.66	1.70	1.70			± 0.35	4.37	4.48	4.48			± 0.90	13.60	13.80	13.80			± 1.90						
	%	66.5	67.9	67.9			± 10.0	56.0	57.4	57.4			± 10.0	73.9	74.8	74.8			± 10.0						
LYMPHO	#	0.51	0.47	0.47			± 0.33	2.33	2.18	2.18			± 0.70	2.24	2.08	2.08			± 1.50						
	%	20.5	18.6	18.6			± 12.0	29.9	27.9	27.9			± 8.0	12.2	11.3	11.3			± 8.0						
MONO	#	0.13	0.14	0.14			± 0.13	0.52	0.54	0.54			± 0.52	0.99	0.99	0.99			± 0.99						
	%	5.3	5.4	5.4			± 5.3	6.7	6.9	6.9			± 6.7	5.4	5.4	5.4			± 5.4						
EOS	#	0.13	0.14	0.14			± 0.13	0.34	0.37	0.37			± 0.34	1.10	1.10	1.10			± 1.10						
	%	5.2	5.6	5.6			± 5.2	4.4	4.8	4.8			± 4.4	6.0	6.0	6.0			± 6.0						
BASO	#	0.06	0.06	0.06			± 0.06	0.23	0.23	0.23			± 0.23	0.46	0.46	0.46			± 0.46						
	%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5						

Ref: TEMP-0821 Rev.37 FRONT / RECTO 1300019740-A