


**LOT** PX 095  
Rev 1


**CONTROL**

 (Exp.) 2015-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE	
			CONTROL					L	CONTROL					N	CONTROL						H
			PENTRA						PENTRA						PENTRA						
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP	60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP	60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP				
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.5	2.5	2.5	2.6	± 0.4	7.4	7.4	7.6	7.4	7.7	± 1.0	18.3	18.1	18.3	18.1	18.6	± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.38	2.40	2.38	2.41	± 0.16	4.54	4.54	4.52	4.54	4.54	± 0.20	5.28	5.30	5.30	5.30	5.26	± 0.25	
HB	HGB	g/dl	6.9	7.0	6.9	7.0	7.0	± 0.4	13.4	13.4	13.4	13.4	13.5	± 0.5	17.1	17.1	17.1	17.1	17.1	± 0.6	
		g/l	69	70	69	70	70	± 4	134	134	134	134	135	± 5	171	171	171	171	171	± 6	
		mmol/l	4.28	4.35	4.28	4.35	4.35	± 0.25	8.32	8.32	8.32	8.32	8.38	± 0.31	10.62	10.62	10.62	10.62	10.62	± 0.37	
HT	HCT	%	19.8	20.0	19.2	20.0	19.3	± 1.5	37.2	37.2	36.6	37.2	36.8	± 2.0	47.0	47.2	46.1	46.6	46.0	± 2.5	
		l/l	0.198	0.200	0.192	0.200	0.193	± 0.015	0.372	0.372	0.366	0.372	0.368	± 0.020	0.470	0.472	0.461	0.466	0.460	± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	81	84	80	84	80.0	± 5	82	82	81	82	81.0	± 5	89	89	87	88	87.5	± 5	
TGMH	MCH	pg	28.3	29.4	28.8	29.4	29.0	± 2.0	29.5	29.5	29.6	29.5	29.7	± 2.0	32.4	32.3	32.3	32.3	32.5	± 2.5	
		fmol	1.76	1.83	1.79	1.83	1.80	± 0.12	1.83	1.83	1.84	1.83	1.85	± 0.12	2.01	2.00	2.00	2.00	2.02	± 0.16	
CCMH	MCHC	g/dl	34.9	35.0	35.9	35.0	36.3	± 3.0	36.0	36.0	36.6	36.0	36.7	± 3.0	36.4	36.3	37.1	36.7	37.2	± 3.0	
		g/l	349	350	359	350	363	± 30	360	360	366	360	367	± 30	364	363	371	367	372	± 30	
		mmol/l	21.68	21.74	22.32	21.74	22.55	± 1.86	22.35	22.35	22.73	22.35	22.80	± 1.86	22.60	22.51	23.03	22.77	23.07	± 1.86	
IDR	RDW	%	13.0	13.0	12.8	13.0	10.8	± 4.0	12.8	13.2	12.5	13.2	11.0	± 4.0	12.8	13.2	11.8	13.2	11.2	± 4.0	
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	70	73	70	74	± 20	246	245	247	245	240	± 30	485	485	505	485	475	± 50	
VPM	MPV	µm <sup>3</sup> ; fl	9.5	9.7	9.5	9.7	9.0	± 2.0	9.5	9.7	9.7	9.7	9.1	± 2.0	9.2	9.5	9.4	9.5	8.8	± 2.0	
NEUT	#		1.32	1.38	1.38	1.38	1.45	± 0.35	4.14	4.19	4.33	4.19	4.45	± 0.90	12.99	12.92	13.23	12.92	13.45	± 1.90	
		%	55.0	55.0	55.3	55.0	55.6	± 10.0	56.0	56.6	57.0	56.6	57.8	± 10.0	71.0	71.4	72.3	71.4	72.3	± 10.0	
LYMPHO	#		0.79	0.83	0.83	0.83	0.83	± 0.33	2.48	2.44	2.51	2.44	2.45	± 0.70	2.93	2.90	2.80	2.90	2.79	± 1.50	
		%	33.0	33.0	33.2	33.0	31.8	± 12.0	33.5	33.0	33.0	33.0	31.8	± 8.0	16.0	16.0	15.3	16.0	15.0	± 8.0	
MONO	#		0.08	0.08	0.08	0.08	0.08	± 0.08	0.22	0.22	0.23	0.22	0.23	± 0.22	0.64	0.54	0.55	0.54	0.56	± 0.54	
		%	3.5	3.3	3.0	3.3	3.0	± 3.0	3.0	3.0	3.0	3.0	3.0	± 3.0	3.5	3.0	3.0	3.0	3.0	± 3.0	
EOS	#		0.13	0.14	0.14	0.14	0.16	± 0.13	0.30	0.30	0.30	0.30	0.31	± 0.30	0.92	0.94	0.92	0.94	1.02	± 0.92	
		%	5.3	5.5	5.5	5.5	6.2	± 5.3	4.0	4.0	4.0	4.0	4.0	± 4.0	5.0	5.2	5.0	5.2	5.5	± 5.0	
BASO	#		0.08	0.08	0.08	0.08	0.09	± 0.08	0.26	0.25	0.23	0.25	0.26	± 0.23	0.82	0.80	0.81	0.80	0.78	± 0.78	
		%	3.2	3.2	3.0	3.2	3.4	± 3.0	3.5	3.4	3.0	3.4	3.4	± 3.0	4.5	4.4	4.4	4.4	4.2	± 4.2	

**LOT** PX 095  
Rev 1

**CONTROL**

 (Exp.) 2015-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA				PENTRA							
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	TOLERANCES TOLERANCE				
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.6	2.6	2.6	± 0.4	7.8	7.8	7.8	± 1.0	19.2	19.2	19.2	± 2.2				
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.44	2.44	2.44	± 0.16	4.58	4.58	4.58	± 0.20	5.39	5.39	5.39	± 0.25				
		g/dl	7.0	7.0	7.0	± 0.4	13.4	13.4	13.4	± 0.5	16.9	16.9	16.9	± 0.6				
HB	HGB	g/l	70	70	70	± 4	134	134	134	± 5	169	169	169	± 6				
		mmol/l	4.35	4.35	4.35	± 0.25	8.32	8.32	8.32	± 0.31	10.49	10.49	10.49	± 0.37				
HT	HCT	%	20.3	20.3	20.3	± 1.5	38.0	38.0	38.0	± 2.0	48.0	48.0	48.0	± 2.5				
		l/l	0.203	0.203	0.203	± 0.015	0.380	0.380	0.380	± 0.020	0.480	0.480	0.480	± 0.025				
VGM	MCV	µm <sup>3</sup> ; fl	83	83	83	± 5	83	83	83	± 5	89	89	89	± 5				
TGMH	MCH	pg	28.7	28.7	28.7	± 2.0	29.3	29.3	29.3	± 2.0	31.4	31.4	31.4	± 2.5				
		fmol	1.78	1.78	1.78	± 0.12	1.82	1.82	1.82	± 0.12	1.95	1.95	1.95	± 0.16				
CCMH	MCHC	g/dl	34.6	34.6	34.6	± 3.0	35.3	35.3	35.3	± 3.0	35.2	35.2	35.2	± 3.0				
		g/l	346	346	346	± 30	353	353	353	± 30	352	352	352	± 30				
		mmol/l	21.46	21.46	21.46	± 1.86	21.89	21.89	21.89	± 1.86	21.88	21.88	21.88	± 1.86				
IDR	RDW	%	15.0	15.0	15.0	± 4.0	15.6	15.6	15.6	± 4.0	14.8	14.8	14.8	± 4.0				
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	77	77	77	± 20	257	257	257	± 30	500	500	500	± 50				
VPM	MPV	µm <sup>3</sup> ; fl	9.7	9.7	9.7	± 2.0	9.5	9.5	9.5	± 2.0	9.3	9.3	9.3	± 2.0				
NEUT	#		1.45	1.52	1.52	± 0.35	4.52	4.67	4.67	± 0.90	14.20	14.50	14.50	± 1.90				
	%		55.6	58.3	58.3	± 10.0	58.0	59.9	59.9	± 10.0	73.8	75.3	75.3	± 10.0				
LYMPHO	#		0.82	0.74	0.74	± 0.33	2.46	2.29	2.29	± 0.70	2.92	2.63	2.63	± 1.50				
	%		31.6	28.4	28.4	± 12.0	31.6	29.3	29.3	± 8.0	15.2	13.7	13.7	± 8.0				
MONO	#		0.15	0.16	0.16	± 0.15	0.30	0.30	0.30	± 0.30	0.77	0.75	0.75	± 0.75				
	%		5.8	6.0	6.0	± 5.8	3.8	3.9	3.9	± 3.8	4.0	3.9	3.9	± 3.9				
EOS	#		0.12	0.12	0.12	± 0.12	0.28	0.30	0.30	± 0.28	0.86	0.88	0.88	± 0.86				
	%		4.5	4.8	4.8	± 4.5	3.6	3.9	3.9	± 3.6	4.5	4.6	4.6	± 4.5				
BASO	#		0.07	0.07	0.07	± 0.07	0.23	0.23	0.23	± 0.23	0.48	0.48	0.48	± 0.48				
	%		2.5	2.5	2.5	± 2.5	3.0	3.0	3.0	± 3.0	2.5	2.5	2.5	± 2.5				

9930096-A  
FRONT / RECTO  
Ref: TEMP-0821 Rev.36