


**LOT** PX 055  
Rev 2


**CONTROL**

 (Exp.) 2015-07-05  
(YYYY - MM - DD)

		<b>ABX Lysebio</b>																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA							PENTRA							PENTRA						
		60 60C+ ES60	80 XL80	MS60	XLR	MSCR			60 60C+ ES60	80 XL80	MS60	XLR	MSCR			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.3	2.4	2.4	2.4	2.4	± 0.4	7.4	7.5	7.5	7.5	7.4	± 1.0	17.8	17.6	18.0	17.6	17.7	± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.50	2.44	2.46	2.44	2.43	± 0.16	4.68	4.70	4.70	4.70	4.64	± 0.20	5.20	5.25	5.24	5.25	5.13	± 0.25			
HB HGB	g/dl	7.1	7.1	7.1	7.1	7.0	± 0.4	14.1	14.2	14.1	14.2	14.0	± 0.5	16.6	16.7	16.6	16.7	16.5	± 0.6			
	g/l	71	71	71	71	70	± 4	141	142	141	142	140	± 5	166	167	166	167	165	± 6			
	mmol/l	4.41	4.41	4.41	4.41	4.35	± 0.25	8.76	8.82	8.76	8.82	8.69	± 0.31	10.31	10.37	10.31	10.37	10.25	± 0.37			
HT HCT	%	20.0	20.3	19.7	20.3	19.3	± 1.5	38.4	38.5	37.6	38.5	37.6	± 2.0	45.8	46.2	45.1	46.2	44.6	± 2.5			
	l/l	0.200	0.203	0.197	0.203	0.193	± 0.015	0.384	0.385	0.376	0.385	0.376	± 0.020	0.458	0.462	0.451	0.462	0.446	± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	80	83	80	83	79.5	± 5	82	82	80	82	81.0	± 5	88	88	86	88	87.0	± 5			
TGMH MCH	pg	28.4	29.1	28.9	29.1	28.8	± 2.0	30.1	30.2	30.0	30.2	30.2	± 2.0	31.9	31.8	31.7	31.8	32.2	± 2.5			
	fmol	1.76	1.81	1.79	1.81	1.79	± 0.12	1.87	1.88	1.86	1.88	1.87	± 0.12	1.98	1.98	1.97	1.98	2.00	± 0.16			
CCMH MCHC	g/dl	35.5	35.1	36.1	35.1	36.2	± 3.0	36.7	36.8	37.5	36.8	37.2	± 3.0	36.3	36.1	36.8	36.1	37.0	± 3.0			
	g/l	355	351	361	351	362	± 30	367	368	375	368	372	± 30	363	361	368	361	370	± 30			
	mmol/l	22.05	21.77	22.40	21.77	22.50	± 1.86	22.82	22.88	23.29	22.88	23.13	± 1.86	22.53	22.45	22.88	22.45	22.96	± 1.86			
IDR RDW	%	12.5	13.7	12.5	13.7	10.7	± 4.0	13.0	14.3	12.5	14.3	10.9	± 4.0	12.7	13.4	11.7	13.4	10.5	± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	68	68	68	70	± 20	262	258	262	258	255	± 30	485	482	495	482	480	± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.3	9.3	9.3	9.3	8.9	± 2.0	9.0	9.2	9.0	9.2	8.5	± 2.0	8.8	9.0	8.9	9.0	8.4	± 2.0			
NEUT	#	1.31	1.42	1.44	1.42	1.52	± 0.35	4.17	4.29	4.35	4.29	4.37	± 0.90	12.37	12.27	12.78	12.27	12.83	± 1.90			
	%	57.0	59.2	60.0	59.2	63.5	± 10.0	56.3	57.2	58.0	57.2	59.0	± 10.0	69.5	69.7	71.0	69.7	72.5	± 10.0			
LYMPHO	#	0.68	0.66	0.66	0.66	0.60	± 0.33	2.46	2.44	2.44	2.44	2.29	± 0.70	2.90	2.90	2.88	2.90	2.60	± 1.50			
	%	29.5	27.5	27.5	27.5	25.0	± 12.0	33.2	32.5	32.5	32.5	31.0	± 8.0	16.3	16.5	16.0	16.5	14.7	± 8.0			
MONO	#	0.09	0.07	0.07	0.07	0.07	± 0.07	0.26	0.23	0.23	0.23	0.22	± 0.22	0.75	0.62	0.63	0.62	0.53	± 0.53			
	%	4.0	3.0	3.0	3.0	3.0	± 3.0	3.5	3.0	3.0	3.0	3.0	± 3.0	4.2	3.5	3.5	3.5	3.0	± 3.0			
EOS	#	0.14	0.16	0.16	0.16	0.12	± 0.12	0.26	0.29	0.26	0.29	0.26	± 0.26	0.98	1.02	0.99	1.02	0.97	± 0.97			
	%	6.0	6.8	6.5	6.8	5.0	± 5.0	3.5	3.8	3.5	3.8	3.5	± 3.5	5.5	5.8	5.5	5.8	5.5	± 5.5			
BASO	#	0.08	0.08	0.07	0.08	0.08	± 0.07	0.26	0.26	0.23	0.26	0.26	± 0.23	0.80	0.79	0.72	0.79	0.76	± 0.72			
	%	3.5	3.5	3.0	3.5	3.5	± 3.0	3.5	3.5	3.0	3.5	3.5	± 3.0	4.5	4.5	4.0	4.5	4.3	± 4.0			

**LOT** PX 055  
Rev 2

**CONTROL**

 (Exp.) 2015-07-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.3	2.3	2.3		± 0.4	7.6	7.6	7.6		± 1.0	18.2	18.2	18.2		± 2.2	
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.48	2.48	2.48		± 0.16	4.70	4.70	4.70		± 0.20	5.27	5.27	5.27		± 0.25	
		g/dl	7.2	7.2	7.2		± 0.4	14.1	14.1	14.1		± 0.5	16.5	16.5	16.5		± 0.6	
HB	HGB	g/l	72	72	72		± 4	141	141	141		± 5	165	165	165		± 6	
		mmol/l	4.47	4.47	4.47		± 0.25	8.76	8.76	8.76		± 0.31	10.25	10.25	10.25		± 0.37	
HT	HCT	%	20.6	20.6	20.6		± 1.5	39.0	39.0	39.0		± 2.0	46.9	46.9	46.9		± 2.5	
		l/l	0.206	0.206	0.206		± 0.015	0.390	0.390	0.390		± 0.020	0.469	0.469	0.469		± 0.025	
VGM	MCV	µm <sup>3</sup> ; fl	83	83	83		± 5	83	83	83		± 5	89	89	89		± 5	
TGMH	MCH	pg	29.0	29.0	29.0		± 2.0	30.0	30.0	30.0		± 2.0	31.3	31.3	31.3		± 2.5	
		fmol	1.80	1.80	1.80		± 0.12	1.86	1.86	1.86		± 0.12	1.94	1.94	1.94		± 0.16	
		g/dl	35.0	35.0	35.0		± 3.0	36.1	36.1	36.1		± 3.0	35.2	35.2	35.2		± 3.0	
CCMH	MCHC	g/l	350	350	350		± 30	361	361	361		± 30	352	352	352		± 30	
		mmol/l	21.72	21.72	21.72		± 1.86	22.45	22.45	22.45		± 1.86	21.85	21.85	21.85		± 1.86	
IDR	RDW	%	14.5	14.5	14.5		± 4.0	16.0	16.0	16.0		± 4.0	14.3	14.3	14.3		± 4.0	
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	72		± 20	265	265	265		± 30	495	495	495		± 50	
VPM	MPV	µm <sup>3</sup> ; fl	9.3	9.3	9.3		± 2.0	8.9	8.9	8.9		± 2.0	8.7	8.7	8.7		± 2.0	
		#	1.37	1.42	1.42		± 0.35	4.31	4.46	4.46		± 0.90	12.70	13.00	13.00		± 1.90	
NEUT		%	59.6	61.8	61.8		± 10.0	56.7	58.7	58.7		± 10.0	69.5	71.2	71.2		± 10.0	
		#	0.62	0.56	0.56		± 0.33	2.48	2.30	2.30		± 0.70	3.19	2.88	2.88		± 1.50	
LYMPHO		%	27.1	24.4	24.4		± 12.0	32.6	30.2	30.2		± 8.0	17.5	15.8	15.8		± 8.0	
		#	0.11	0.12	0.12		± 0.11	0.40	0.42	0.42		± 0.40	1.20	1.20	1.20		± 1.20	
MONO		%	4.9	5.0	5.0		± 4.9	5.3	5.5	5.5		± 5.3	6.6	6.6	6.6		± 6.6	
		#	0.14	0.14	0.14		± 0.14	0.18	0.20	0.20		± 0.18	0.71	0.71	0.71		± 0.71	
EOS		%	5.9	6.3	6.3		± 5.9	2.4	2.6	2.6		± 2.4	3.9	3.9	3.9		± 3.9	
		#	0.06	0.06	0.06		± 0.06	0.23	0.23	0.23		± 0.23	0.46	0.46	0.46		± 0.46	
BASO		%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5	

9930096-A  
FRONT / RECTO  
Ref: TEMP-0821 Rev.36