

LOT **PX 093**
 Rev 1

CONTROL

(Exp.) **2013-11-05**
 (YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE		
			PENTRA					PENTRA					PENTRA							
			60 60C+ ES60	80 XL80	MS60	XL80	60 60C+ ES60	80 XL80	MS60	XL80	60 60C+ ES60	80 XL80	MS60	XL80	60 60C+ ES60	80 XL80	MS60		XL80	
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.2	2.3	2.3	2.3		± 0.4	7.2	7.3	7.3	7.3		± 1.0	17.2	17.1	17.3		17.1	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.38	2.40	2.38		± 0.12	4.62	4.60	4.60	4.60		± 0.15	5.12	5.12	5.10	5.12		± 0.20
HB	HGB	g/dl	6.8	6.7	6.7	6.7		± 0.4	13.6	13.6	13.5	13.6		± 0.5	16.3	16.3	16.3	16.3		± 0.6
		g/l	68	67	67	67		± 4	136	136	135	136		± 5	163	163	163	163		± 6
		mmol/l	4.22	4.16	4.16	4.16		± 0.25	8.45	8.45	8.38	8.45		± 0.31	10.12	10.12	10.12	10.12		± 0.37
HT	HCT	%	19.2	19.5	18.7	19.5		± 1.5	37.4	37.7	36.8	37.7		± 2.0	45.1	45.6	43.9	45.6		± 2.5
		l/l	0.192	0.195	0.187	0.195		± 0.015	0.374	0.377	0.368	0.377		± 0.020	0.451	0.456	0.439	0.456		± 0.025
VGM	MCV	µm ³ ; fl	79	82	78	82		± 5	81	82	80	82		± 5	88	89	86	89		± 5
TGMH	MCH	pg	28.0	28.2	27.9	28.2		± 2.0	29.4	29.6	29.3	29.6		± 2.0	31.8	31.8	32.0	31.8		± 2.5
		fmol	1.74	1.75	1.73	1.75		± 0.12	1.83	1.84	1.82	1.84		± 0.12	1.98	1.98	1.98	1.98		± 0.16
CCMH	MCHC	g/dl	35.4	34.3	35.8	34.3		± 3.0	36.3	36.1	36.7	36.1		± 3.0	36.2	35.8	37.2	35.8		± 3.0
		g/l	354	343	358	343		± 30	363	361	367	361		± 30	362	358	372	358		± 30
		mmol/l	22.00	21.32	22.23	21.32		± 1.86	22.57	22.39	22.78	22.39		± 1.86	22.47	22.21	23.08	22.21		± 1.86
IDR	RDW	%	13.0	13.8	12.9	13.8		± 4.0	13.0	13.7	12.5	13.7		± 4.0	12.5	13.7	12.2	13.7		± 4.0
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	65	67	65		± 20	240	245	252	245		± 30	500	505	515	505		± 50
VPM	MPV	µm ³ ; fl	9.1	9.2	8.6	9.2		± 2.0	9.0	9.2	8.8	9.2		± 2.0	9.0	9.2	8.9	9.2		± 2.0
NEUT	#		1.18	1.29	1.24	1.29		± 0.25	4.08	4.16	4.09	4.16		± 0.75	12.13	12.06	12.25	12.06		± 1.80
		%	53.5	56.0	54.0	56.0		± 10.0	56.6	57.0	56.0	57.0		± 10.0	70.5	70.5	70.8	70.5		± 10.0
LYMPHO	#		0.70	0.70	0.74	0.70		± 0.20	2.28	2.26	2.34	2.26		± 0.60	2.84	2.74	2.85	2.74		± 1.40
		%	32.0	30.3	32.2	30.3		± 8.0	31.6	31.0	32.0	31.0		± 8.0	16.5	16.0	16.5	16.0		± 8.0
MONO	#		0.11	0.09	0.09	0.09		± 0.09	0.29	0.29	0.29	0.29		± 0.29	0.69	0.69	0.69	0.69		± 0.69
		%	5.0	4.0	4.0	4.0		± 4.0	4.0	4.0	4.0	4.0		± 4.0	4.0	4.0	4.0	4.0		± 4.0
EOS	#		0.15	0.15	0.15	0.15		± 0.15	0.31	0.33	0.33	0.33		± 0.30	0.77	0.86	0.78	0.86		± 0.62
		%	6.5	6.5	6.5	6.5		± 6.5	4.3	4.5	4.5	4.5		± 4.0	4.5	5.0	4.5	5.0		± 3.5
BASO	#		0.07	0.07	0.08	0.07		± 0.07	0.25	0.26	0.26	0.26		± 0.19	0.77	0.77	0.73	0.77		± 0.35
		%	3.0	3.2	3.3	3.2		± 3.0	3.5	3.5	3.5	3.5		± 2.5	4.5	4.5	4.2	4.5		± 2.0

LOT PX 093

CONTROL

(Exp.) 2013-11-05
 (YYYY-MM-DD)

Rev 1

		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA						
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				TOLERANCES TOLERANCE	120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3	2.3			± 0.4	7.5	7.5	7.5			± 1.0	17.7	17.7	17.7			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.42	2.42	2.42			± 0.12	4.63	4.63	4.63			± 0.15	5.20	5.20	5.20			± 0.20	
HB HGB	g/dl	6.8	6.8	6.8			± 0.4	13.4	13.4	13.4			± 0.5	16.1	16.1	16.1			± 0.6	
	g/l	68	68	68			± 4	134	134	134			± 5	161	161	161			± 6	
	mmol/l	4.22	4.22	4.22			± 0.25	8.32	8.32	8.32			± 0.31	10.00	10.00	10.00			± 0.37	
HT HCT	%	19.6	19.6	19.6			± 1.5	38.4	38.4	38.4			± 2.0	46.3	46.3	46.3			± 2.5	
	l/l	0.196	0.196	0.196			± 0.015	0.384	0.384	0.384			± 0.020	0.463	0.463	0.463			± 0.025	
VGM MCV	µm ³ ; fl	81	81	81			± 5	83	83	83			± 5	89	89	89			± 5	
TGMH MCH	pg	28.1	28.1	28.1			± 2.0	28.9	28.9	28.9			± 2.0	31.0	31.0	31.0			± 2.5	
	fmol	1.74	1.74	1.74			± 0.12	1.80	1.80	1.80			± 0.12	1.92	1.92	1.92			± 0.16	
CCMH MCHC	g/dl	34.7	34.7	34.7			± 3.0	34.9	34.9	34.9			± 3.0	34.8	34.8	34.8			± 3.0	
	g/l	347	347	347			± 30	349	349	349			± 30	348	348	348			± 30	
	mmol/l	21.54	21.54	21.54			± 1.86	21.65	21.65	21.65			± 1.86	21.60	21.60	21.60			± 1.86	
IDR RDW	%	16.0	16.0	16.0			± 4.0	16.5	16.5	16.5			± 4.0	15.3	15.3	15.3			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	70			± 20	248	248	248			± 30	510	510	510			± 50	
VPM MPV	µm ³ ; fl	9.0	9.0	9.0			± 2.0	9.1	9.1	9.1			± 2.0	9.0	9.0	9.0			± 2.0	
NEUT	#	1.21	1.34	1.34			± 0.25	4.16	4.49	4.49			± 0.75	12.50	13.20	13.20			± 1.80	
	%	52.7	58.3	58.3			± 10.0	55.5	59.9	59.9			± 10.0	70.7	74.1	74.1			± 10.0	
LYMPHO	#	0.72	0.64	0.64			± 0.20	2.29	2.06	2.06			± 0.60	2.92	2.53	2.53			± 1.40	
	%	31.3	27.7	27.7			± 8.0	30.5	27.4	27.4			± 8.0	16.5	14.3	14.3			± 8.0	
MONO	#	0.15	0.10	0.10			± 0.09	0.53	0.45	0.45			± 0.29	1.15	0.99	0.99			± 0.69	
	%	6.5	4.2	4.2			± 4.0	7.0	6.0	6.0			± 4.0	6.5	5.6	5.6			± 4.0	
EOS	#	0.15	0.16	0.16			± 0.15	0.30	0.32	0.32			± 0.30	0.62	0.71	0.71			± 0.62	
	%	6.5	6.8	6.8			± 6.5	4.0	4.2	4.2			± 4.0	3.5	4.0	4.0			± 3.5	
BASO	#	0.07	0.07	0.07			± 0.07	0.23	0.19	0.19			± 0.19	0.50	0.35	0.35			± 0.35	
	%	3.0	3.0	3.0			± 3.0	3.0	2.5	2.5			± 2.5	2.8	2.0	2.0			± 2.0	

Ref: TEMP-0821 Rev.33 BACK / VERSO 9930080-B