

LOT PX 114
 Rev 1

CONTROL

(Exp.) 2015-01-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE	
			CONTROL					L	CONTROL					N	CONTROL						H
			PENTRA					MSCRCP	PENTRA					MSCRCP	PENTRA						MSCRCP
			60 60C+ ES60	80 XL80	MS60	XLR			60 60C+ ES60	80 XL80	MS60	XLR			60 60C+ ES60	80 XL80	MS60	XLR			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6	2.6	2.6	± 0.4	7.8	7.8	7.8	7.8	7.7	± 1.0	18.1	17.8	18.1	17.8	17.9	± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.32	2.37	2.32	2.30	± 0.16	4.50	4.50	4.50	4.50	4.41	± 0.20	5.08	5.09	5.10	5.09	5.01	± 0.25	
HB	HGB	g/dl	6.7	6.7	6.7	6.7	6.6	± 0.4	13.3	13.4	13.3	13.4	13.2	± 0.5	16.3	16.5	16.3	16.5	16.2	± 0.6	
		g/l	67	67	67	67	66	± 4	133	134	133	134	132	± 5	163	165	163	165	162	± 6	
		mmol/l	4.16	4.16	4.16	4.16	4.10	± 0.25	8.26	8.32	8.26	8.32	8.20	± 0.31	10.12	10.25	10.12	10.25	10.06	± 0.37	
HT	HCT	%	19.0	19.3	18.7	19.3	18.4	± 1.5	36.9	37.4	36.0	37.4	36.5	± 2.0	45.2	45.3	44.4	45.3	44.3	± 2.5	
		l/l	0.190	0.193	0.187	0.193	0.184	± 0.015	0.369	0.374	0.360	0.374	0.365	± 0.020	0.452	0.453	0.444	0.453	0.443	± 0.025	
VGM	MCV	µm ³ ·fl	80	83	79	83	80.2	± 5	82	83	80	83	82.7	± 5	89	89	87	89	88.4	± 5	
TGMH	MCH	pg	28.3	28.9	28.3	28.9	28.7	± 2.0	29.6	29.8	29.6	29.8	29.9	± 2.0	32.1	32.4	32.0	32.4	32.3	± 2.5	
		fmol	1.76	1.79	1.76	1.79	1.78	± 0.12	1.84	1.85	1.84	1.85	1.86	± 0.12	1.99	2.01	1.98	2.01	2.01	± 0.16	
CCMH	MCHC	g/dl	35.3	34.8	35.8	34.8	35.8	± 3.0	36.0	35.9	36.9	35.9	36.2	± 3.0	36.1	36.4	36.7	36.4	36.6	± 3.0	
		g/l	353	348	358	348	358	± 30	360	359	369	359	362	± 30	361	364	367	364	366	± 30	
		mmol/l	21.94	21.61	22.22	21.61	22.22	± 1.86	22.38	22.28	22.94	22.28	22.48	± 1.86	22.39	22.62	22.81	22.62	22.72	± 1.86	
IDR	RDW	%	12.5	13.3	12.5	13.3	10.7	± 4.0	12.5	13.6	12.0	13.6	10.7	± 4.0	12.0	13.4	11.8	13.4	11.1	± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	78	75	74	75	73	± 20	249	250	250	250	250	± 30	490	490	495	490	490	± 50	
VPM	MPV	µm ³ ·fl	9.1	9.4	9.2	9.4	8.6	± 2.0	9.6	10.0	9.8	10.0	9.0	± 2.0	9.6	9.9	9.9	9.9	8.9	± 2.0	
NEUT	#	%	1.49	1.48	1.47	1.48	1.51	± 0.35	4.43	4.46	4.41	4.46	4.40	± 0.90	12.67	12.50	12.60	12.50	12.60	± 1.90	
		%	57.3	57.1	56.5	57.1	57.5	± 10.0	56.8	57.2	56.5	57.2	57.1	± 10.0	70.0	70.2	69.6	70.2	70.4	± 10.0	
LYMPHO	#	%	0.79	0.79	0.79	0.79	0.75	± 0.35	2.48	2.48	2.46	2.48	2.45	± 0.70	3.08	3.03	3.08	3.03	2.97	± 1.50	
		%	30.2	30.5	30.5	30.5	30.0	± 12.0	31.8	31.8	31.5	31.8	31.8	± 8.0	17.0	17.0	17.0	17.0	16.6	± 8.0	
MONO	#	%	0.10	0.09	0.10	0.09	0.10	± 0.09	0.30	0.27	0.31	0.27	0.29	± 0.27	0.72	0.68	0.81	0.68	0.72	± 0.68	
		%	4.0	3.5	4.0	3.5	4.0	± 3.5	3.8	3.5	4.0	3.5	3.8	± 3.5	4.0	3.8	4.5	3.8	4.0	± 3.8	
EOS	#	%	0.13	0.14	0.14	0.14	0.13	± 0.13	0.30	0.31	0.35	0.31	0.29	± 0.29	0.81	0.80	0.81	0.80	0.81	± 0.80	
		%	5.0	5.5	5.5	5.5	5.0	± 5.0	3.8	4.0	4.5	4.0	3.8	± 3.8	4.5	4.5	4.5	4.5	4.5	± 4.5	
BASO	#	%	0.09	0.09	0.09	0.09	0.09	± 0.09	0.30	0.27	0.27	0.27	0.27	± 0.27	0.81	0.80	0.80	0.80	0.81	± 0.80	
		%	3.5	3.4	3.5	3.4	3.5	± 3.4	3.8	3.5	3.5	3.5	3.5	± 3.5	4.5	4.5	4.4	4.5	4.5	± 4.4	

LOT PX 114
 Rev 1

CONTROL

(Exp.) 2015-01-05
 (YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6		± 0.4	8.0	8.0	8.0		± 1.0	18.7	18.7	18.7		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38	2.38	2.38		± 0.16	4.50	4.50	4.50		± 0.20	5.15	5.15	5.15		± 0.25			
HB HGB	g/dl	6.8	6.8	6.8		± 0.4	13.3	13.3	13.3		± 0.5	16.3	16.3	16.3		± 0.6			
	g/l	68	68	68		± 4	133	133	133		± 5	163	163	163		± 6			
	mmol/l	4.22	4.22	4.22		± 0.25	8.26	8.26	8.26		± 0.31	10.12	10.12	10.12		± 0.37			
HT HCT	%	19.8	19.8	19.8		± 1.5	37.4	37.4	37.4		± 2.0	45.8	45.8	45.8		± 2.5			
	l/l	0.198	0.198	0.198		± 0.015	0.374	0.374	0.374		± 0.020	0.458	0.458	0.458		± 0.025			
VGM MCV	µm ³ ·fl	83	83	83		± 5	83	83	83		± 5	89	89	89		± 5			
TGMH MCH	pg	28.6	28.6	28.6		± 2.0	29.6	29.6	29.6		± 2.0	31.7	31.7	31.7		± 2.5			
	fmol	1.77	1.77	1.77		± 0.12	1.84	1.84	1.84		± 0.12	1.97	1.97	1.97		± 0.16			
CCMH MCHC	g/dl	34.4	34.4	34.4		± 3.0	35.6	35.6	35.6		± 3.0	35.6	35.6	35.6		± 3.0			
	g/l	344	344	344		± 30	356	356	356		± 30	356	356	356		± 30			
	mmol/l	21.38	21.38	21.38		± 1.86	22.11	22.11	22.11		± 1.86	22.08	22.08	22.08		± 1.86			
IDR RDW	%	14.9	14.9	14.9		± 4.0	16.4	16.4	16.4		± 4.0	15.8	15.8	15.8		± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	80	80	80		± 20	260	260	260		± 30	505	505	505		± 50			
VPM MPV	µm ³ ·fl	9.2	9.2	9.2		± 2.0	9.8	9.8	9.8		± 2.0	9.5	9.5	9.5		± 2.0			
NEUT	#	1.49	1.54	1.54		± 0.35	4.44	4.58	4.58		± 0.90	12.90	13.20	13.20		± 1.90			
	%	57.2	59.3	59.3		± 10.0	55.5	57.2	57.2		± 10.0	69.1	70.3	70.3		± 10.0			
LYMPHO	#	0.78	0.71	0.71		± 0.35	2.54	2.38	2.38		± 0.70	3.18	2.94	2.94		± 1.50			
	%	29.9	27.4	27.4		± 12.0	31.8	29.7	29.7		± 8.0	17.0	15.7	15.7		± 8.0			
MONO	#	0.14	0.14	0.14		± 0.14	0.50	0.52	0.52		± 0.50	1.33	1.33	1.33		± 1.33			
	%	5.4	5.5	5.5		± 5.4	6.3	6.5	6.5		± 6.3	7.1	7.1	7.1		± 7.1			
EOS	#	0.13	0.14	0.14		± 0.13	0.27	0.29	0.29		± 0.27	0.80	0.82	0.82		± 0.80			
	%	5.0	5.3	5.3		± 5.0	3.4	3.6	3.6		± 3.4	4.3	4.4	4.4		± 4.3			
BASO	#	0.07	0.07	0.07		± 0.07	0.24	0.24	0.24		± 0.24	0.47	0.47	0.47		± 0.47			
	%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5			

9930091-A
 Ref: TEMP-0821 Rev.35