

LOT PX 113

CONTROL

(Exp.) **2014-01-05**
 (YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE	
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			60 60C+ ES60	80 XL80	MS60	XL80		60 60C+ ES60	80 XL80	MS60	XL80		60 60C+ ES60	80 XL80	MS60	XL80			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.5	2.4	2.5	± 0.4		7.4	7.3	7.4	7.3		± 1.0	17.7	17.5	17.9			17.5
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.33	2.37	2.33	± 0.12	4.58	4.56	4.55	4.56	± 0.15	5.15	5.15	5.14	5.15	± 0.20			
HB HGB	g/dl	6.7	6.7	6.7	6.7	± 0.4	13.4	13.5	13.3	13.5	± 0.5	16.4	16.4	16.3	16.4	± 0.6			
	g/l	67	67	67	67	± 4	134	135	133	135	± 5	164	164	163	164	± 6			
	mmol/l	4.16	4.16	4.16	4.16	± 0.25	8.32	8.38	8.26	8.38	± 0.31	10.18	10.18	10.12	10.18	± 0.37			
HT HCT	%	18.9	19.1	18.5	19.1	± 1.5	37.1	37.4	35.9	37.4	± 2.0	44.8	45.3	43.7	45.3	± 2.5			
	l/l	0.189	0.191	0.185	0.191	± 0.015	0.371	0.374	0.359	0.374	± 0.020	0.448	0.453	0.437	0.453	± 0.025			
VGM MCV	µm ³ ; fl	79	82	78	82	± 5	81	82	79	82	± 5	87	88	85	88	± 5			
TGMH MCH	pg	28.0	28.8	28.3	28.8	± 2.0	29.3	29.6	29.2	29.6	± 2.0	31.8	31.8	31.7	31.8	± 2.5			
	fmol	1.74	1.79	1.76	1.79	± 0.12	1.82	1.84	1.82	1.84	± 0.12	1.98	1.98	1.97	1.98	± 0.16			
CCMH MCHC	g/dl	35.5	35.1	36.2	35.1	± 3.0	36.1	36.1	37.0	36.1	± 3.0	36.6	36.2	37.3	36.2	± 3.0			
	g/l	355	351	362	351	± 30	361	361	370	361	± 30	366	362	373	362	± 30			
	mmol/l	22.04	21.78	22.51	21.78	± 1.86	22.43	22.42	22.98	22.42	± 1.86	22.73	22.47	23.17	22.47	± 1.86			
IDR RDW	%	13.0	14.0	12.5	14.0	± 4.0	12.5	13.8	12.5	13.8	± 4.0	11.6	13.3	11.6	13.3	± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	70	72	70	± 20	258	260	265	260	± 30	505	515	520	515	± 50			
VPM MPV	µm ³ ; fl	9.0	9.2	9.2	9.2	± 2.0	8.7	9.0	8.9	9.0	± 2.0	8.8	8.9	8.9	8.9	± 2.0			
NEUT	#	1.36	1.39	1.30	1.39	± 0.30	3.96	3.83	4.00	3.83	± 0.80	12.74	12.43	12.80	12.43	± 1.90			
	%	56.5	55.5	54.0	55.5	± 10.0	53.5	52.5	54.0	52.5	± 10.0	72.0	71.0	71.5	71.0	± 10.0			
LYMPHO	#	0.73	0.79	0.79	0.79	± 0.25	2.59	2.63	2.63	2.63	± 0.60	2.74	2.80	2.77	2.80	± 1.50			
	%	30.5	31.5	33.0	31.5	± 8.0	35.0	36.0	35.5	36.0	± 8.0	15.5	16.0	15.5	16.0	± 8.0			
MONO	#	0.08	0.08	0.08	0.08	± 0.08	0.33	0.30	0.30	0.30	± 0.30	0.72	0.72	0.72	0.72	± 0.72			
	%	3.0	3.0	3.0	3.0	± 3.0	4.5	4.0	4.0	4.0	± 4.0	4.0	4.0	4.0	4.0	± 4.0			
EOS	#	0.16	0.16	0.16	0.16	± 0.11	0.26	0.29	0.24	0.29	± 0.19	0.71	0.79	0.81	0.79	± 0.45			
	%	6.5	6.5	6.7	6.5	± 4.5	3.5	4.0	3.2	4.0	± 2.5	4.0	4.5	4.5	4.5	± 2.5			
BASO	#	0.08	0.09	0.08	0.09	± 0.07	0.26	0.26	0.24	0.26	± 0.23	0.80	0.79	0.81	0.79	± 0.45			
	%	3.5	3.5	3.3	3.5	± 3.0	3.5	3.5	3.3	3.5	± 3.0	4.5	4.5	4.5	4.5	± 2.5			

LOT PX 113
Rev 2

CONTROL

 (Exp.) 2014-01-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE			
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.6	7.6	7.6		± 1.0	18.1	18.1	18.1		± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41	2.41		± 0.12	4.60	4.60	4.60		± 0.15	5.22	5.22	5.22		± 0.20
HB	HGB	g/dl	6.8	6.8	6.8		± 0.4	13.3	13.3	13.3		± 0.5	16.2	16.2	16.2		± 0.6
		g/l	68	68	68		± 4	133	133	133		± 5	162	162	162		± 6
		mmol/l	4.22	4.22	4.22		± 0.25	8.26	8.26	8.26		± 0.31	10.06	10.06	10.06		± 0.37
HT	HCT	%	19.5	19.5	19.5		± 1.5	37.7	37.7	37.7		± 2.0	45.9	45.9	45.9		± 2.5
		l/l	0.195	0.195	0.195		± 0.015	0.377	0.377	0.377		± 0.020	0.459	0.459	0.459		± 0.025
VGM	MCV	µm ³ ; fl	81	81	81		± 5	82	82	82		± 5	88	88	88		± 5
TGMH	MCH	pg	28.2	28.2	28.2		± 2.0	28.9	28.9	28.9		± 2.0	31.0	31.0	31.0		± 2.5
		fmol	1.75	1.75	1.75		± 0.12	1.80	1.80	1.80		± 0.12	1.93	1.93	1.93		± 0.16
CCMH	MCHC	g/dl	34.8	34.8	34.8		± 3.0	35.3	35.3	35.3		± 3.0	35.3	35.3	35.3		± 3.0
		g/l	348	348	348		± 30	353	353	353		± 30	353	353	353		± 30
		mmol/l	21.63	21.63	21.63		± 1.86	21.90	21.90	21.90		± 1.86	21.90	21.90	21.90		± 1.86
IDR	RDW	%	17.0	17.0	17.0		± 4.0	17.7	17.7	17.7		± 4.0	15.7	15.7	15.7		± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75	75		± 20	263	263	263		± 30	515	515	515		± 50
VPM	MPV	µm ³ ; fl	9.0	9.0	9.0		± 2.0	9.0	9.0	9.0		± 2.0	8.8	8.8	8.8		± 2.0
NEUT	#		1.33	1.41	1.41		± 0.30	3.95	4.20	4.20		± 0.80	13.20	13.19	13.19		± 1.90
		%	55.5	58.7	58.7		± 10.0	52.0	55.2	55.2		± 10.0	73.0	72.9	72.9		± 10.0
LYMPHO	#		0.74	0.65	0.65		± 0.25	2.66	2.41	2.41		± 0.60	2.72	2.57	2.57		± 1.50
		%	31.0	27.2	27.2		± 8.0	35.0	31.7	31.7		± 8.0	15.0	14.2	14.2		± 8.0
MONO	#		0.14	0.11	0.11		± 0.08	0.57	0.55	0.55		± 0.30	1.27	1.19	1.19		± 0.72
		%	6.0	4.7	4.7		± 3.0	7.5	7.3	7.3		± 4.0	7.0	6.6	6.6		± 4.0
EOS	#		0.11	0.15	0.15		± 0.11	0.19	0.21	0.21		± 0.19	0.45	0.69	0.69		± 0.45
		%	4.5	6.4	6.4		± 4.5	2.5	2.8	2.8		± 2.5	2.5	3.8	3.8		± 2.5
BASO	#		0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.45	0.45	0.45		± 0.45
		%	3.0	3.0	3.0		± 3.0	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5

Ref: TEMP-0821 Rev.33 BACK / VERSO 9930080-B