

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT

PX 013

Unprotect.AssaySheet_PX

CONTROL

Generate .csv_PX

(Exp.) **2013-03-05**
 (YYYY - MM - DD)

Rev 1

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80					60 60C+ ES60	80 XL80				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4			± 0.4	7.2	7.2			± 1.0	17.6	17.4			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.25	2.20			± 0.12	4.70	4.70			± 0.15	5.14	5.18			± 0.20			
HB HGB	g/dl	6.3	6.2			± 0.4	13.7	13.8			± 0.5	16.1	16.2			± 0.6			
	g/l	63	62			± 4	137	138			± 5	161	162			± 6			
	mmol/l	3.91	3.85			± 0.25	8.51	8.57			± 0.31	10.00	10.06			± 0.37			
HT HCT	%	17.6	17.8			± 1.5	38.1	38.1			± 2.0	44.7	45.1			± 2.5			
	l/l	0.176	0.178			± 0.015	0.381	0.381			± 0.020	0.447	0.451			± 0.025			
VGM MCV	µm ³ /fl	78	81			± 5	81	81			± 5	87	87			± 5			
TGMH MCH	pg	28.0	28.2			± 2.0	29.1	29.4			± 2.0	31.3	31.3			± 2.5			
	fmol	1.74	1.75			± 0.12	1.81	1.82			± 0.12	1.95	1.94			± 0.16			
CCMH MCHC	g/dl	35.9	34.8			± 3.0	36.0	36.2			± 3.0	36.0	35.9			± 3.0			
	g/l	359	348			± 30	360	362			± 30	360	359			± 30			
	mmol/l	22.29	21.61			± 1.86	22.35	22.51			± 1.86	22.36	22.32			± 1.86			
IDR RDW	%	12.5	13.2			± 4.0	12.4	13.0			± 4.0	12.3	13.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75			± 20	255	255			± 30	505	520			± 50			
VPM MPV	µm ³ /fl	9.3	9.6			± 2.0	9.1	9.4			± 2.0	8.7	9.0			± 2.0			
NEUT	#	1.32	1.32			± 0.30	3.78	3.71			± 0.80	12.06	11.57			± 1.90			
	%	55.0	55.0			± 10.0	52.5	51.5			± 10.0	68.5	66.5			± 10.0			
LYMPHO	#	0.74	0.74			± 0.30	2.45	2.48			± 0.70	3.08	3.13			± 1.50			
	%	31.0	31.0			± 8.0	34.0	34.5			± 8.0	17.5	18.0			± 8.0			
MONO	#	0.13	0.13			± 0.13	0.38	0.38			± 0.38	0.70	0.78			± 0.70			
	%	5.0	5.0			± 5.0	5.0	5.0			± 5.0	4.0	4.5			± 4.0			
EOS	#	0.13	0.14			± 0.13	0.36	0.43			± 0.30	0.97	1.13			± 0.74			
	%	5.5	6.0			± 5.5	5.0	6.0			± 4.0	5.5	6.5			± 4.0			
BASO	#	0.08	0.07			± 0.07	0.25	0.23			± 0.23	0.79	0.78			± 0.46			
	%	3.5	3.0			± 2.5	3.5	3.0			± 3.0	4.5	4.5			± 2.5			

Ref: TEMP-0821 Rev.18 FRONT / RECTO 9930061-A

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 013
 Rev 1

CONTROL

(Exp.) 2013-03-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse												TOLERANCES TOLERANCE						
				CONTROL				L				CONTROL						N				
				PENTRA								PENTRA										
				120	DX 120			120	DX 120			120	DX 120					120	DX 120			
120 RETIC	DF120			120 RETIC	DF120			120 RETIC	DF120			120 RETIC	DF120									
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5					± 0.4	7.5	7.5					± 1.0	18.5	18.5					± 2.2
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.26	2.26					± 0.12	4.74	4.74					± 0.15	5.25	5.25					± 0.20
HB HGB	g/dl	6.4	6.4					± 0.4	13.8	13.8					± 0.5	16.2	16.2					± 0.6
	g/l	64	64					± 4	138	138					± 5	162	162					± 6
	mmol/l	3.97	3.97					± 0.25	8.57	8.57					± 0.31	10.06	10.06					± 0.37
HT HCT	%	18.5	18.5					± 1.5	38.9	38.9					± 2.0	46.2	46.2					± 2.5
	l/l	0.185	0.185					± 0.015	0.389	0.389					± 0.020	0.462	0.462					± 0.025
VGM MCV	µm ³ ·fl	82	82					± 5	82	82					± 5	88	88					± 5
TGMH MCH	pg	28.3	28.3					± 2.0	29.1	29.1					± 2.0	30.9	30.9					± 2.5
	fmol	1.76	1.76					± 0.12	1.81	1.81					± 0.12	1.92	1.92					± 0.16
CCMH MCHC	g/dl	34.5	34.5					± 3.0	35.5	35.5					± 3.0	35.1	35.1					± 3.0
	g/l	345	345					± 30	355	355					± 30	351	351					± 30
	mmol/l	21.45	21.45					± 1.86	22.05	22.05					± 1.86	21.78	21.78					± 1.86
IDR RDW	%	15.1	15.1					± 4.0	15.9	15.9					± 4.0	15.0	15.0					± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75					± 20	262	262					± 30	515	515					± 50
VPM MPV	µm ³ ·fl	9.9	9.9					± 2.0	9.5	9.5					± 2.0	9.1	9.1					± 2.0
NEUT	#	1.39	1.40					± 0.30	4.05	4.28					± 0.80	13.00	13.00					± 1.90
	%	55.5	56.0					± 10.0	54.0	57.0					± 10.0	70.0	70.0					± 10.0
LYMPHO	#	0.78	0.73					± 0.30	2.55	2.25					± 0.70	3.15	2.96					± 1.50
	%	31.0	29.0					± 8.0	34.0	30.0					± 8.0	17.0	16.0					± 8.0
MONO	#	0.13	0.15					± 0.13	0.38	0.41					± 0.38	1.20	1.30					± 0.70
	%	5.0	6.0					± 5.0	5.0	5.5					± 5.0	6.5	7.0					± 4.0
EOS	#	0.15	0.15					± 0.13	0.30	0.30					± 0.30	0.74	0.74					± 0.74
	%	6.0	6.0					± 5.5	4.0	4.0					± 4.0	4.0	4.0					± 4.0
BASO	#	0.07	0.08					± 0.07	0.23	0.26					± 0.23	0.46	0.56					± 0.46
	%	2.5	3.0					± 2.5	3.0	3.5					± 3.0	2.5	3.0					± 2.5

Ref: TEMP-0821 Rev.18 FRONT / RECTO 9930066-A