

LOT PX 090
 Rev 1

CONTROL

(Exp.) 2010-11-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS		ABX Lyse																		
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE	
		PENTRA				MS60		PENTRA				MS60		PENTRA				MS60		
		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			60 60C+ ES60
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.6	2.5	2.5		± 0.4	7.6	7.6	7.8	7.8		± 1.0	17.9	17.7	18.5	18.5		± 2.2
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.40	2.44	2.44		± 0.12	4.57	4.57	4.63	4.63		± 0.15	5.07	5.10	5.20	5.20		± 0.20
HB	HGB	g/dl	6.2	6.2	6.2	6.2		± 0.4	13.5	13.5	13.5	13.5		± 0.5	16.4	16.4	16.4	16.4		± 0.6
		g/l	62	62	62	62		± 4	135	135	135	135		± 5	164	164	164	164		± 6
		mmol/l	3.85	3.85	3.85	3.85		± 0.25	8.38	8.38	8.38	8.38		± 0.31	10.18	10.18	10.18	10.18		± 0.37
HT	HCT	%	18.7	19.2	19.5	19.5		± 1.5	39.8	39.8	40.7	40.7		± 2.0	47.2	47.4	48.4	48.4		± 2.5
		l/l	0.187	0.192	0.195	0.195		± 0.015	0.398	0.398	0.407	0.407		± 0.020	0.472	0.474	0.484	0.484		± 0.025
VGM	MCV	µm ³ ; fl	77	80	80	80		± 4	87	87	88	88		± 4	93	93	93	93		± 4
TGMH	MCH	pg	25.5	25.8	25.4	25.4		± 2.0	29.5	29.5	29.2	29.2		± 2.0	32.3	32.2	31.5	31.5		± 2.5
		fmol	1.58	1.60	1.58	1.58		± 0.12	1.83	1.83	1.81	1.81		± 0.12	2.01	2.00	1.96	1.96		± 0.16
CCMH	MCHC	g/dl	33.1	32.3	31.8	31.8		± 3.0	34.0	34.0	33.1	33.1		± 3.0	34.8	34.6	33.9	33.9		± 3.0
		g/l	331	323	318	318		± 30	340	340	331	331		± 30	348	346	339	339		± 30
		mmol/l	20.58	20.05	19.72	19.72		± 1.86	21.09	21.09	20.58	20.58		± 1.86	21.60	21.47	21.06	21.06		± 1.86
IDR	RDW	%	14.1	13.8	15.2	15.2		± 4.0	11.8	12.6	14.2	14.2		± 4.0	11.4	12.2	13.6	13.6		± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	67	72	72		± 20	252	250	260	260		± 30	495	495	500	500		± 50
VPM	MPV	µm ³ ; fl	9.4	9.4	9.3	9.3		± 2.0	8.5	8.6	8.7	8.7		± 2.0	8.5	8.6	8.6	8.6		± 2.0
NEUT	#		1.43	1.51	1.39	1.39		± 0.30	4.71	4.67	4.99	4.99		± 0.80	13.51	13.40	14.40	14.40		± 1.90
		%	57.0	58.0	55.7	55.7		± 10.0	62.0	61.5	64.0	64.0		± 10.0	75.5	75.7	77.5	77.5		± 10.0
LYMPHO	#		0.78	0.78	0.80	0.80		± 0.30	2.24	2.28	2.18	2.18		± 0.70	2.42	2.44	2.41	2.41		± 1.50
		%	31.0	30.0	32.0	32.0		± 8.0	29.5	30.0	28.0	28.0		± 8.0	13.5	13.8	13.0	13.0		± 8.0
MONO	#		0.05	0.05	0.05	0.05		± 0.05	0.12	0.12	0.12	0.12		± 0.12	0.28	0.28	0.28	0.28		± 0.28
		%	1.8	1.8	1.8	1.8		± 1.8	1.5	1.5	1.5	1.5		± 1.5	1.5	1.5	1.5	1.5		± 1.5
EOS	#		0.18	0.18	0.18	0.18		± 0.18	0.27	0.27	0.27	0.27		± 0.27	0.93	0.93	0.93	0.93		± 0.93
		%	7.0	7.0	7.0	7.0		± 7.0	3.5	3.5	3.5	3.5		± 3.5	5.0	5.0	5.0	5.0		± 5.0
BASO	#		0.08	0.08	0.09	0.09		± 0.08	0.27	0.27	0.23	0.23		± 0.23	0.81	0.71	0.56	0.56		± 0.56
		%	3.2	3.2	3.5	3.5		± 3.0	3.5	3.5	3.0	3.0		± 3.0	4.5	4.0	3.0	3.0		± 3.0

Ref: FORM-0821 Rev.11 FRONT / RECTO 9930052-A