

ABX Difftrol



LOT PX 052
Rev 1

CONTROL

(Exp.) 2012-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse												TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	
			60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80			60 60C+ ES60	80 XL80			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3		± 0.4	7.3	7.3		± 1.0	17.8	17.6		± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.38	2.33		± 0.12	4.48	4.48		± 0.15	5.06	5.06		± 0.20	
HB	HGB	g/dl	6.8	6.8		± 0.4	13.4	13.4		± 0.5	16.2	16.2		± 0.6	
		g/l	68	68		± 4	134	134		± 5	162	162		± 6	
		mmol/l	4.22	4.22		± 0.25	8.32	8.32		± 0.31	10.06	10.06		± 0.37	
HT	HCT	%	19.0	19.6		± 1.5	36.7	37.2		± 2.0	44.5	44.5		± 2.5	
		l/l	0.190	0.196		± 0.015	0.367	0.372		± 0.020	0.445	0.445		± 0.025	
VGM	MCV	µm ³ fl	80	84		± 4	82	83		± 4	88	88		± 4	
TGMH	MCH	pg	28.6	29.2		± 2.0	29.9	29.9		± 2.0	32.0	32.0		± 2.5	
		fmol	1.77	1.81		± 0.12	1.86	1.86		± 0.12	1.99	1.99		± 0.16	
CCMH	MCHC	g/dl	35.7	34.7		± 3.0	36.5	36.0		± 3.0	36.4	36.4		± 3.0	
		g/l	357	347		± 30	365	360		± 30	364	364		± 30	
		mmol/l	22.18	21.58		± 1.86	22.65	22.38		± 1.86	22.59	22.59		± 1.86	
IDR	RDW	%	13.5	14.0		± 4.0	12.2	13.5		± 4.0	12.0	13.7		± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70		± 20	252	255		± 30	500	505		± 50	
VPM	MPV	µm ³ fl	8.7	8.9		± 2.0	8.2	8.3		± 2.0	8.2	8.3		± 2.0	
NEUT		#	1.27	1.35		± 0.30	3.96	3.96		± 0.80	12.32	12.06		± 1.80	
		%	55.0	58.5		± 10.0	54.3	54.2		± 10.0	69.2	68.5		± 10.0	
LYMPHO		#	0.72	0.67		± 0.20	2.52	2.50		± 0.60	3.03	2.90		± 1.50	
		%	31.5	29.0		± 8.0	34.5	34.2		± 8.0	17.0	16.5		± 8.0	
MONO		#	0.10	0.08		± 0.08	0.31	0.22		± 0.22	0.85	0.62		± 0.62	
		%	4.5	3.5		± 3.5	4.2	3.0		± 3.0	4.8	3.5		± 3.5	
EOS		#	0.13	0.13		± 0.12	0.26	0.35		± 0.26	0.80	1.23		± 0.80	
		%	5.8	5.8		± 5.0	3.5	4.8		± 3.5	4.5	7.0		± 4.5	
BASO		#	0.07	0.07		± 0.07	0.26	0.28		± 0.26	0.80	0.80		± 0.80	
		%	3.2	3.2		± 3.2	3.5	3.8		± 3.5	4.5	4.5		± 4.0	

9930061-A FRONT / RECTO Ref. TEMP-0821 Rev.17

ABX Diffrol



LOT PX 052
Rev 1

CONTROL

(Exp.) 2012-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse												TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	
			120 120 RETIC	DX120 DF120			120 120 RETIC	DX120 DF120			120 120 RETIC	DX120 DF120			
GB WBC	$10^9/mm^3 ; 10^9/l$	2.4	2.4		± 0.4	7.6	7.6		± 1.0	18.7	18.7		± 2.2		
GR RBC	$10^9/mm^3 ; 10^{12}/l$	2.38	2.38		± 0.12	4.52	4.52		± 0.15	5.18	5.18		± 0.20		
HB HGB	g/dl	6.8	6.8		± 0.4	13.4	13.4		± 0.5	16.1	16.1		± 0.6		
	g/l	68	68		± 4	134	134		± 5	161	161		± 6		
	mmol/l	4.22	4.22		± 0.25	8.32	8.32		± 0.31	10.00	10.00		± 0.37		
HT HCT	%	20.0	20.0		± 1.5	38.0	38.0		± 2.0	46.1	46.1		± 2.5		
	l/l	0.200	0.200		± 0.015	0.380	0.380		± 0.020	0.461	0.461		± 0.025		
VGM MCV	$\mu m^3 ; fl$	84	84		± 4	84	84		± 4	89	89		± 4		
TGMH MCH	pg	28.6	28.6		± 2.0	29.6	29.6		± 2.0	31.1	31.1		± 2.5		
	fmol	1.77	1.77		± 0.12	1.84	1.84		± 0.12	1.93	1.93		± 0.16		
CCMH MCHC	g/dl	34.0	34.0		± 3.0	35.3	35.3		± 3.0	34.9	34.9		± 3.0		
	g/l	340	340		± 30	353	353		± 30	349	349		± 30		
	mmol/l	21.12	21.12		± 1.86	21.92	21.92		± 1.86	21.69	21.69		± 1.86		
IDR RDW	%	15.6	15.6		± 4.0	16.0	16.0		± 4.0	15.5	15.5		± 4.0		
PLAQ. PLTS	$10^9/mm^3 ; 10^9/l$	70	70		± 20	253	253		± 30	500	500		± 50		
VPM MPV	$\mu m^3 ; fl$	9.0	9.0		± 2.0	8.2	8.2		± 2.0	8.2	8.2		± 2.0		
NEUT	#	1.36	1.43		± 0.30	4.18	4.29		± 0.80	13.50	13.40		± 1.90		
	%	56.5	59.5		± 10.0	55.0	56.5		± 10.0	72.0	71.8		± 10.0		
LYMPHO	#	0.74	0.65		± 0.20	2.66	2.32		± 0.70	3.18	2.84		± 1.50		
	%	31.0	27.0		± 8.0	35.0	30.5		± 8.0	17.0	15.2		± 8.0		
MONO	#	0.12	0.14		± 0.12	0.34	0.46		± 0.34	0.94	1.22		± 0.94		
	%	5.0	6.0		± 5.0	4.5	6.0		± 4.5	5.0	6.5		± 5.0		
EOS	#	0.12	0.12		± 0.12	0.23	0.27		± 0.23	0.71	0.75		± 0.71		
	%	5.0	5.0		± 5.0	3.0	3.5		± 3.0	3.8	4.0		± 3.8		
BASO	#	0.06	0.06		± 0.06	0.19	0.27		± 0.19	0.41	0.47		± 0.41		
	%	2.5	2.5		± 2.5	2.5	3.5		± 2.5	2.2	2.5		± 2.2		

Ref: TEMP-0821 Rev.17 FRONT / RECTO 9930066-A