

ABX Diffrol



LOT PX 032
Rev 1

CONTROL

(Exp.) 2012-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse																	
			CONTROL					L	CONTROL					N	CONTROL					H
			PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE
			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120	MS60	
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.4	2.4	2.4		± 0.4	7.4	7.4	7.8	7.8		± 1.0	17.7	17.6	18.4	18.4		± 2.2
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.43	2.36	2.44	2.44		± 0.12	4.66	4.62	4.68	4.68		± 0.15	5.15	5.16	5.26	5.26		± 0.20
HB	HGB	g/dl	6.8	6.8	6.9	6.9		± 0.4	13.7	13.6	13.7	13.7		± 0.5	16.4	16.4	16.5	16.5		± 0.6
		g/l	68	68	69	69		± 4	137	136	137	137		± 5	164	164	165	165		± 6
		mmol/l	4.22	4.22	4.28	4.28		± 0.25	8.51	8.45	8.51	8.51		± 0.31	10.18	10.18	10.25	10.25		± 0.37
HT	HCT	%	19.4	19.4	20.0	20.0		± 1.5	38.2	37.9	38.8	38.8		± 2.0	45.8	45.9	46.8	46.8		± 2.5
		l/l	0.194	0.194	0.200	0.200		± 0.015	0.382	0.379	0.388	0.388		± 0.020	0.458	0.459	0.468	0.468		± 0.025
VGM	MCV	µm ³ fl	80	82	82	82		± 4	82	82	83	83		± 4	89	89	89	89		± 4
TGMH	MCH	pg	28.0	28.8	28.3	28.3		± 2.0	29.4	29.4	29.3	29.3		± 2.0	31.8	31.8	31.4	31.4		± 2.5
		fmol	1.74	1.79	1.76	1.76		± 0.12	1.83	1.83	1.82	1.82		± 0.12	1.98	1.97	1.95	1.95		± 0.16
CCMH	MCHC	g/dl	35.0	35.1	34.5	34.5		± 3.0	35.9	35.9	35.3	35.3		± 3.0	35.8	35.7	35.2	35.2		± 3.0
		g/l	350	351	345	345		± 30	359	359	353	353		± 30	358	357	352	352		± 30
		mmol/l	21.72	21.82	21.42	21.42		± 1.86	22.26	22.29	21.90	21.90		± 1.86	22.22	22.18	21.89	21.89		± 1.86
IDR	RDW	%	12.8	13.5	14.5	14.5		± 4.0	13.0	13.2	14.8	14.8		± 4.0	12.0	13.2	14.5	14.5		± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	72	72		± 20	240	245	245	245		± 30	490	500	495	495		± 50
VPM	MPV	µm ³ fl	8.6	8.7	8.9	8.9		± 2.0	8.1	8.3	8.5	8.5		± 2.0	8.1	8.2	8.4	8.4		± 2.0
NEUT	#		1.28	1.36	1.37	1.42		± 0.30	4.07	4.03	4.37	4.51		± 0.80	12.60	12.58	13.70	13.80		± 1.90
		%	55.8	56.5	57.0	59.0		± 10.0	55.0	54.5	56.0	57.8		± 10.0	71.2	71.5	74.0	74.8		± 10.0
LYMPHO	#		0.75	0.76	0.76	0.70		± 0.25	2.52	2.55	2.69	2.50		± 0.70	2.76	2.64	2.82	2.58		± 1.50
		%	32.4	31.5	31.5	29.0		± 8.0	34.0	34.5	34.5	32.0		± 8.0	15.6	15.0	15.3	14.0		± 8.0
MONO	#		0.10	0.10	0.10	0.12		± 0.10	0.31	0.31	0.31	0.33		± 0.31	0.80	0.75	0.83	0.83		± 0.75
		%	4.0	4.0	4.0	5.0		± 4.0	4.0	4.0	4.0	4.2		± 4.0	4.5	4.0	4.5	4.5		± 4.0
EOS	#		0.11	0.12	0.12	0.11		± 0.11	0.26	0.26	0.23	0.23		± 0.23	0.74	0.88	0.74	0.77		± 0.74
		%	4.8	5.0	5.0	4.5		± 4.5	3.5	3.5	3.0	3.0		± 3.0	4.2	5.0	4.0	4.2		± 4.0
BASO	#		0.07	0.07	0.06	0.06		± 0.06	0.26	0.26	0.20	0.23		± 0.20	0.80	0.79	0.40	0.46		± 0.40
		%	3.0	3.0	2.5	2.5		± 2.5	3.5	3.5	2.5	3.0		± 2.5	4.5	4.5	2.2	2.5		± 2.2

Ref: FORM-0821 Rev.15 FRONT / RECTO 9930052-B