

# ABX Diffrol



**LOT** PX 111  
Rev 1

**CONTROL**

(Exp.) 2012-01-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lyse																
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE	
			PENTRA				MS60	PENTRA				MS60	PENTRA				MS60		
			60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.3	2.4	2.4	2.4	± 0.4	7.3	7.3	7.6	7.6	± 1.0	17.5	17.4	18.2	18.2	± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.45	2.40	2.44	2.44	± 0.12	4.58	4.55	4.59	4.59	± 0.15	5.02	5.02	5.09	5.09	± 0.20			
HB HGB	g/dl	6.7	6.7	6.8	6.8	± 0.4	13.5	13.5	13.4	13.4	± 0.5	15.7	15.8	15.7	15.7	± 0.6			
	g/l	67	67	68	68	± 4	135	135	134	134	± 5	157	158	157	157	± 6			
	mmol/l	4.16	4.16	4.22	4.22	± 0.25	8.38	8.38	8.32	8.32	± 0.31	9.75	9.81	9.75	9.75	± 0.37			
HT HCT	%	19.6	19.9	20.0	20.0	± 1.5	37.6	37.3	38.6	38.6	± 2.0	44.7	44.7	45.3	45.3	± 2.5			
	l/l	0.196	0.199	0.200	0.200	± 0.015	0.376	0.373	0.386	0.386	± 0.020	0.447	0.447	0.453	0.453	± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	80	83	82	82	± 5	82	82	84	84	± 5	89	89	89	89	± 5			
TGMH MCH	pg	27.3	27.9	27.9	27.9	± 2.0	29.5	29.7	29.2	29.2	± 2.0	31.3	31.5	30.8	30.8	± 2.5			
	fmol	1.70	1.73	1.73	1.73	± 0.12	1.83	1.84	1.81	1.81	± 0.12	1.94	1.95	1.92	1.92	± 0.16			
CCMH MCHC	g/dl	34.2	33.6	34.0	34.0	± 3.0	35.9	36.2	34.8	34.8	± 3.0	35.1	35.4	34.7	34.7	± 3.0			
	g/l	342	336	340	340	± 30	359	362	348	348	± 30	351	354	347	347	± 30			
	mmol/l	21.23	20.89	21.11	21.11	± 1.86	22.32	22.47	21.58	21.58	± 1.86	21.82	21.96	21.52	21.52	± 1.86			
IDR RDW	%	12.9	12.7	14.4	14.4	± 4.0	12.5	13.0	14.9	14.9	± 4.0	12.2	12.7	13.4	13.4	± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	74	80	80	± 20	260	262	260	260	± 30	505	505	495	495	± 50			
VPM MPV	µm <sup>3</sup> ; fl	8.8	9.0	8.3	8.3	± 2.0	8.3	8.5	8.3	8.3	± 2.0	8.0	8.1	8.0	8.0	± 2.0			
NEUT	#	1.30	1.36	1.38	1.37	± 0.30	3.80	3.69	4.03	3.95	± 0.80	11.94	11.66	12.90	12.80	± 1.90			
	%	56.5	56.5	57.5	57.0	± 10.0	52.0	50.5	53.0	52.0	± 10.0	68.2	67.0	70.5	70.0	± 10.0			
LYMPHO	#	0.69	0.73	0.73	0.72	± 0.20	2.63	2.74	2.77	2.77	± 0.70	3.06	3.22	3.46	3.46	± 1.50			
	%	30.0	30.5	30.5	30.0	± 8.0	36.0	37.5	36.5	36.5	± 8.0	17.5	18.5	19.0	19.0	± 8.0			
MONO	#	0.10	0.10	0.10	0.10	± 0.10	0.37	0.37	0.34	0.38	± 0.34	0.91	0.91	0.91	0.91	± 0.91			
	%	4.0	4.0	4.0	4.0	± 4.0	5.0	5.0	4.5	5.0	± 4.5	5.0	5.0	5.0	5.0	± 5.0			
EOS	#	0.15	0.15	0.13	0.16	± 0.13	0.29	0.29	0.23	0.27	± 0.23	0.88	0.96	0.64	0.73	± 0.64			
	%	6.5	6.4	5.5	6.5	± 5.5	4.0	4.0	3.0	3.5	± 3.0	5.0	5.5	3.5	4.0	± 3.5			
BASO	#	0.07	0.06	0.06	0.06	± 0.06	0.23	0.23	0.23	0.23	± 0.23	0.75	0.70	0.36	0.36	± 0.36			
	%	3.0	2.6	2.5	2.5	± 2.5	3.0	3.0	3.0	3.0	± 3.0	4.3	4.0	2.0	2.0	± 2.0			

9930052-B FRONT / RECTO Ref: FORM-0821 Rev.14