

LOT PX 071

CONTROL

(Exp.) 2011-09-05
 (YYYY-MM-DD)

Rev 1

		ABX Lyse																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA					MS60		PENTRA					MS60		PENTRA					MS60	
		60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120				60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120				60 60C+ ES60	80 XL80	120 120 RETIC	DX120 DF120			
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.6	2.6		± 0.4	7.4	7.4	7.8	7.8		± 1.0	17.8	17.6	18.4	18.4		± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.34	2.28	2.35	2.35		± 0.12	4.67	4.67	4.74	4.74		± 0.15	5.20	5.20	5.30	5.30		± 0.20			
HB HGB	g/dl	5.8	5.8	5.9	5.9		± 0.4	13.6	13.5	13.6	13.6		± 0.5	16.7	16.7	16.6	16.6		± 0.6			
	g/l	58	58	59	59		± 4	136	135	136	136		± 5	167	167	166	166		± 6			
	mmol/l	3.60	3.60	3.66	3.66		± 0.25	8.45	8.38	8.45	8.45		± 0.31	10.37	10.37	10.31	10.31		± 0.37			
HT HCT	%	17.6	17.8	18.6	18.6		± 1.5	39.7	40.2	41.2	41.2		± 2.0	48.4	48.4	49.3	49.3		± 2.5			
	l/l	0.176	0.178	0.186	0.186		± 0.015	0.397	0.402	0.412	0.412		± 0.020	0.484	0.484	0.493	0.493		± 0.025			
VGM MCV	µm ³ ; fl	75	78	79	79		± 5	85	86	87	87		± 4	93	93	93	93		± 4			
TGMH MCH	pg	24.8	25.4	25.1	25.1		± 2.0	29.1	28.9	28.7	28.7		± 2.0	32.1	32.1	31.3	31.3		± 2.5			
	fmol	1.54	1.58	1.56	1.56		± 0.12	1.81	1.80	1.78	1.78		± 0.12	1.99	1.99	1.95	1.95		± 0.16			
CCMH MCHC	g/dl	33.0	32.6	31.8	31.8		± 3.0	34.3	33.6	33.0	33.0		± 3.0	34.5	34.5	33.7	33.7		± 3.0			
	g/l	330	326	318	318		± 30	343	336	330	330		± 30	345	345	337	337		± 30			
	mmol/l	20.52	20.25	19.74	19.74		± 1.86	21.28	20.87	20.48	20.48		± 1.86	21.44	21.44	20.91	20.91		± 1.86			
IDR RDW	%	13.5	12.9	15.2	15.2		± 4.0	12.2	12.3	14.5	14.5		± 4.0	11.6	11.7	13.6	13.6		± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	68	70	70		± 20	260	255	260	260		± 30	495	495	495	495		± 50			
VPM MPV	µm ³ ; fl	9.4	9.4	9.4	9.4		± 2.0	8.1	8.2	8.2	8.2		± 2.0	8.0	8.0	8.3	8.3		± 2.0			
NEUT	#	1.50	1.49	1.55	1.58		± 0.30	4.66	4.59	4.89	4.98		± 0.80	13.47	13.09	14.20	14.30		± 1.90			
	%	60.0	59.5	59.7	60.8		± 10.0	63.0	62.0	62.7	63.8		± 10.0	75.7	74.4	77.2	77.7		± 10.0			
LYMPHO	#	0.72	0.73	0.75	0.72		± 0.30	2.09	2.17	2.30	2.22		± 0.70	2.44	2.55	2.67	2.39		± 1.50			
	%	28.7	29.2	29.0	27.5		± 8.0	28.3	29.3	29.5	28.4		± 8.0	13.7	14.5	14.5	13.0		± 8.0			
MONO	#	0.05	0.05	0.05	0.05		± 0.05	0.14	0.14	0.14	0.14		± 0.14	0.33	0.33	0.33	0.33		± 0.33			
	%	1.8	1.8	1.8	1.8		± 1.8	1.8	1.8	1.8	1.8		± 1.8	1.8	1.8	1.8	1.8		± 1.8			
EOS	#	0.17	0.17	0.17	0.17		± 0.17	0.27	0.27	0.27	0.27		± 0.27	0.83	0.88	0.83	0.98		± 0.83			
	%	6.5	6.5	6.5	6.5		± 6.5	3.5	3.5	3.5	3.5		± 3.5	4.5	5.0	4.5	5.3		± 4.5			
BASO	#	0.08	0.08	0.08	0.09		± 0.08	0.25	0.25	0.20	0.20		± 0.20	0.77	0.76	0.37	0.40		± 0.37			
	%	3.0	3.0	3.0	3.4		± 3.0	3.4	3.4	2.5	2.5		± 2.5	4.3	4.3	2.0	2.2		± 2.0			

Ref: FORM-0821 Rev.12 FRONT / RECTO 9930052-B