

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT

PX 013

Unprotect.AssaySheet_PX

CONTROL

(Exp.) 2013-03-05
 (YYYY - MM - DD)

Rev 1

		ABX Lysebio														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA				
		60 60C+ ES60	80 XL80	MS60			60 60C+ ES60	80 XL80	MS60			60 60C+ ES60	80 XL80	MS60		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.2	7.2	7.3		± 1.0	17.6	17.4	17.4		± 2.2
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.25	2.20	2.25		± 0.12	4.70	4.70	4.70		± 0.15	5.14	5.18	5.15		± 0.20
HB HGB	g/dl	6.2	6.2	6.2		± 0.4	13.7	13.8	13.8		± 0.5	16.2	16.2	16.3		± 0.6
	g/l	62	62	62		± 4	137	138	138		± 5	162	162	163		± 6
	mmol/l	3.85	3.85	3.85		± 0.25	8.51	8.57	8.57		± 0.31	10.06	10.06	10.12		± 0.37
HT HCT	%	17.6	17.8	17.1		± 1.5	38.1	38.1	36.7		± 2.0	44.7	45.1	43.8		± 2.5
	l/l	0.176	0.178	0.171		± 0.015	0.381	0.381	0.367		± 0.020	0.447	0.451	0.438		± 0.025
VGM MCV	µm ³ ; fl	78	81	76		± 5.0	81	81	78		± 5.0	87	87	85		± 5.0
TGMH MCH	pg	27.6	28.2	27.6		± 2.0	29.1	29.4	29.4		± 2.0	31.5	31.3	31.7		± 2.5
	fmol	1.71	1.75	1.71		± 0.12	1.81	1.82	1.82		± 0.12	1.96	1.94	1.97		± 0.16
CCMH MCHC	g/dl	35.3	34.8	36.3		± 3.0	36.0	36.2	37.6		± 3.0	36.2	35.9	37.2		± 3.0
	g/l	353	348	363		± 30	360	362	376		± 30	362	359	372		± 30
	mmol/l	21.94	21.61	22.52		± 1.86	22.35	22.51	23.38		± 1.86	22.50	22.32	23.12		± 1.86
IDR RDW	%	12.5	13.2	12.2		± 4.0	12.4	13.0	12.0		± 4.0	12.3	13.0	12.0		± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75	75		± 20	255	255	265		± 30	505	520	522		± 50
VPM MPV	µm ³ ; fl	9.3	9.6	9.4		± 2.0	9.1	9.4	9.1		± 2.0	8.7	9.0	8.7		± 2.0
NEUT	#	1.32	1.32	1.30		± 0.30	3.78	3.71	3.80		± 0.80	12.06	11.57	11.92		± 1.90
	%	55.0	55.0	54.0		± 10.0	52.5	51.5	52.0		± 10.0	68.5	66.5	68.5		± 10.0
LYMPHO	#	0.74	0.74	0.78		± 0.30	2.45	2.48	2.56		± 0.70	3.08	3.13	3.08		± 1.50
	%	31.0	31.0	32.5		± 8.0	34.0	34.5	35.0		± 8.0	17.5	18.0	17.7		± 8.0
MONO	#	0.13	0.13	0.13		± 0.13	0.38	0.38	0.38		± 0.38	0.70	0.78	0.78		± 0.70
	%	5.0	5.0	5.0		± 5.0	5.0	5.0	5.0		± 5.0	4.0	4.5	4.5		± 4.0
EOS	#	0.13	0.14	0.13		± 0.13	0.36	0.43	0.33		± 0.30	0.97	1.13	0.84		± 0.74
	%	5.5	6.0	5.5		± 5.5	5.0	6.0	4.5		± 4.0	5.5	6.5	4.8		± 4.0
BASO	#	0.08	0.07	0.07		± 0.07	0.25	0.23	0.26		± 0.23	0.79	0.78	0.78		± 0.46
	%	3.5	3.0	3.0		± 2.5	3.5	3.0	3.5		± 3.0	4.5	4.5	4.5		± 2.5

9930061-A
BACK / VERSO
Ref: TEMP-0821 Rev.18

HORIBA ABX SAS
 Parc Euromédecine
 34184 Montpellier cedex 4 - France
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT PX 013
 Rev 1

CONTROL

(Exp.) 2013-03-05
 (YYYY-MM-DD)

		ABX Lysebio														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA				
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.5	7.5	7.5		± 1.0	18.5	18.5	18.5		± 2.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.26	2.26	2.26		± 0.12	4.74	4.74	4.74		± 0.15	5.25	5.25	5.25		± 0.20
HB HGB	g/dl	6.3	6.3	6.3		± 0.4	13.8	13.8	13.8		± 0.5	16.2	16.2	16.2		± 0.6
	g/l	63	63	63		± 4	138	138	138		± 5	162	162	162		± 6
	mmol/l	3.91	3.91	3.91		± 0.25	8.57	8.57	8.57		± 0.31	10.06	10.06	10.06		± 0.37
HT HCT	%	18.5	18.5	18.5		± 1.5	38.9	38.9	38.9		± 2.0	46.2	46.2	46.2		± 2.5
	l/l	0.185	0.185	0.185		± 0.015	0.389	0.389	0.389		± 0.020	0.462	0.462	0.462		± 0.025
VGM MCV	µm ³ ; fl	82	82	82		± 5.0	82	82	82		± 5.0	88	88	88		± 5.0
TGMH MCH	pg	27.9	27.9	27.9		± 2.0	29.1	29.1	29.1		± 2.0	30.9	30.9	30.9		± 2.5
	fmol	1.73	1.73	1.73		± 0.12	1.81	1.81	1.81		± 0.12	1.92	1.92	1.92		± 0.16
CCMH MCHC	g/dl	34.0	34.0	34.0		± 3.0	35.5	35.5	35.5		± 3.0	35.1	35.1	35.1		± 3.0
	g/l	340	340	340		± 30	355	355	355		± 30	351	351	351		± 30
	mmol/l	21.11	21.11	21.11		± 1.86	22.05	22.05	22.05		± 1.86	21.78	21.78	21.78		± 1.86
IDR RDW	%	15.1	15.1	15.1		± 4.0	15.9	15.9	15.9		± 4.0	15.0	15.0	15.0		± 4.0
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	75	75		± 20	262	262	262		± 30	515	515	515		± 50
VPM MPV	µm ³ ; fl	9.9	9.9	9.9		± 2.0	9.5	9.5	9.5		± 2.0	9.1	9.1	9.1		± 2.0
NEUT	#	1.39	1.40	1.40		± 0.30	4.05	4.28	4.28		± 0.80	13.00	13.00	13.00		± 1.90
	%	55.5	56.0	56.0		± 10.0	54.0	57.0	57.0		± 10.0	70.0	70.0	70.0		± 10.0
LYMPHO	#	0.78	0.73	0.73		± 0.30	2.55	2.25	2.25		± 0.70	3.15	2.96	2.96		± 1.50
	%	31.0	29.0	29.0		± 8.0	34.0	30.0	30.0		± 8.0	17.0	16.0	16.0		± 8.0
MONO	#	0.13	0.15	0.15		± 0.13	0.38	0.41	0.41		± 0.38	1.20	1.30	1.30		± 0.70
	%	5.0	6.0	6.0		± 5.0	5.0	5.5	5.5		± 5.0	6.5	7.0	7.0		± 4.0
EOS	#	0.15	0.15	0.15		± 0.13	0.30	0.30	0.30		± 0.30	0.74	0.74	0.74		± 0.74
	%	6.0	6.0	6.0		± 5.5	4.0	4.0	4.0		± 4.0	4.0	4.0	4.0		± 4.0
BASO	#	0.07	0.08	0.08		± 0.07	0.23	0.26	0.26		± 0.23	0.46	0.56	0.56		± 0.46
	%	2.5	3.0	3.0		± 2.5	3.0	3.5	3.5		± 3.0	2.5	3.0	3.0		± 2.5

9930066-A
 Ref: TEMP-0821 Rev.18 BACK / VERSO