


LOT PX 402
Rev 2


CONTROL

 (Exp.) 2017-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE		
			CONTROL					L	CONTROL					N	CONTROL						H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA							
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4	2.4	2.5	± 0.4	7.6	7.6	7.6	7.6	7.6	± 1.0	18.0	17.7	17.9	17.7	18.2	± 2.2		
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.35	2.32	2.35	2.32	2.35	± 0.16	4.54	4.54	4.54	4.54	4.52	± 0.20	5.05	5.12	5.10	5.12	5.10	± 0.25		
HB	HGB	g/dl	6.8	6.8	6.8	6.8	6.8	± 0.4	13.5	13.6	13.5	13.6	13.4	± 0.5	16.4	16.5	16.3	16.5	16.3	± 0.6		
		g/l	68	68	68	68	68	± 4	135	136	135	136	134	± 5	164	165	163	165	163	± 6		
		mmol/l	4.22	4.22	4.22	4.22	4.22	± 0.25	8.38	8.45	8.38	8.45	8.32	± 0.31	10.18	10.25	10.12	10.25	10.12	± 0.37		
HT	HCT	%	19.0	19.3	19.0	19.3	18.9	± 1.5	36.8	36.8	36.3	36.8	36.6	± 2.0	44.4	45.1	43.9	45.1	44.6	± 2.5		
		l/l	0.190	0.193	0.190	0.193	0.189	± 0.015	0.368	0.368	0.363	0.368	0.366	± 0.020	0.444	0.451	0.439	0.451	0.446	± 0.025		
VGM	MCV	µm ³ ; fl	81	83	81	83	80.5	± 5	81	81	80	81	81.0	± 5	88	88	86	88	87.5	± 5		
TGMH	MCH	pg	28.9	29.3	28.9	29.3	28.9	± 2.0	29.7	30.0	29.7	30.0	29.6	± 2.0	32.5	32.2	32.0	32.2	32.0	± 2.5		
		fmol	1.80	1.82	1.80	1.82	1.80	± 0.12	1.85	1.86	1.85	1.86	1.84	± 0.12	2.02	2.00	1.98	2.00	1.98	± 0.16		
CCMH	MCHC	g/dl	35.7	35.3	35.7	35.3	35.9	± 3.0	36.7	37.0	37.2	37.0	36.6	± 3.0	36.9	36.6	37.2	36.6	36.5	± 3.0		
		g/l	357	353	357	353	359	± 30	367	370	372	370	366	± 30	369	366	372	366	365	± 30		
		mmol/l	22.18	21.93	22.18	21.93	22.32	± 1.86	22.80	22.97	23.08	22.97	22.73	± 1.86	22.92	22.74	23.08	22.74	22.68	± 1.86		
IDR	RDW	%	13.4	13.2	12.7	13.2	12.0	± 4.0	13.4	13.8	12.6	13.8	13.0	± 4.0	12.7	13.3	12.0	13.3	12.4	± 4.0		
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	68	68	68	70	± 20	248	243	248	243	241	± 30	490	495	510	495	485	± 50		
VPM	MPV	µm ³ ; fl	8.9	9.2	9.2	9.2	8.7	± 2.0	9.2	9.4	9.2	9.4	8.8	± 2.0	8.8	9.0	8.9	9.0	8.4	± 2.0		
NEUT	#		1.38	1.41	1.38	1.41	1.47	± 0.35	4.40	4.45	4.37	4.45	4.45	± 0.90	12.96	12.96	13.07	12.96	13.41	± 1.90		
		%	57.5	58.7	57.5	58.7	58.9	± 10.0	57.9	58.5	57.5	58.5	58.6	± 10.0	72.0	73.2	73.0	73.2	73.7	± 10.0		
LYMPHO	#		0.69	0.67	0.70	0.67	0.72	± 0.33	2.28	2.26	2.39	2.26	2.33	± 0.70	2.52	2.37	2.51	2.37	2.60	± 1.50		
		%	28.7	27.8	29.0	27.8	28.6	± 12.0	30.0	29.7	31.5	29.7	30.7	± 8.0	14.0	13.4	14.0	13.4	14.3	± 8.0		
MONO	#		0.11	0.09	0.10	0.09	0.11	± 0.09	0.34	0.30	0.27	0.30	0.29	± 0.27	0.81	0.80	0.64	0.80	0.66	± 0.64		
		%	4.5	3.7	4.2	3.7	4.2	± 3.7	4.5	4.0	3.5	4.0	3.8	± 3.5	4.5	4.5	3.6	4.5	3.6	± 3.6		
EOS	#		0.14	0.16	0.14	0.16	0.12	± 0.12	0.29	0.32	0.30	0.32	0.27	± 0.27	0.90	0.80	0.90	0.80	0.76	± 0.76		
		%	6.0	6.5	6.0	6.5	4.9	± 4.9	3.8	4.2	4.0	4.2	3.5	± 3.5	5.0	4.5	5.0	4.5	4.2	± 4.2		
BASO	#		0.08	0.08	0.08	0.08	0.09	± 0.08	0.29	0.27	0.27	0.27	0.26	± 0.26	0.81	0.78	0.79	0.78	0.76	± 0.76		
		%	3.3	3.3	3.3	3.3	3.4	± 3.3	3.8	3.6	3.5	3.6	3.4	± 3.4	4.5	4.4	4.4	4.4	4.2	± 4.2		

LOT PX 402
Rev 2

CONTROL

 (Exp.) 2017-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	
			120	DX120	DX NEXUS			120	DX120	DX NEXUS			120	DX120	DX NEXUS			
120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS			120 RETIC	DF120	DF NEXUS							
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.8	7.8	7.8		± 1.0	18.6	18.6	18.6		± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.35	2.35		± 0.16	4.54	4.54	4.54		± 0.20	5.16	5.16	5.16		± 0.25	
		g/dl	6.9	6.9	6.9		± 0.4	13.5	13.5	13.5		± 0.5	16.3	16.3	16.3		± 0.6	
HB	HGB	g/l	69	69	69		± 4	135	135	135		± 5	163	163	163		± 6	
		mmol/l	4.28	4.28	4.28		± 0.25	8.38	8.38	8.38		± 0.31	10.12	10.12	10.12		± 0.37	
HT	HCT	%	19.5	19.5	19.5		± 1.5	37.2	37.2	37.2		± 2.0	45.4	45.4	45.4		± 2.5	
		l/l	0.195	0.195	0.195		± 0.015	0.372	0.372	0.372		± 0.020	0.454	0.454	0.454		± 0.025	
VGM	MCV	µm ³ ; fl	83	83	83		± 5	82	82	82		± 5	88	88	88		± 5	
TGMH	MCH	pg	29.4	29.4	29.4		± 2.0	29.7	29.7	29.7		± 2.0	31.6	31.6	31.6		± 2.5	
		fmol	1.82	1.82	1.82		± 0.12	1.85	1.85	1.85		± 0.12	1.96	1.96	1.96		± 0.16	
		g/dl	35.4	35.4	35.4		± 3.0	36.3	36.3	36.3		± 3.0	35.9	35.9	35.9		± 3.0	
CCMH	MCHC	g/l	354	354	354		± 30	363	363	363		± 30	359	359	359		± 30	
		mmol/l	21.97	21.97	21.97		± 1.86	22.52	22.52	22.52		± 1.86	22.29	22.29	22.29		± 1.86	
IDR	RDW	%	15.0	15.0	15.0		± 4.0	16.8	16.8	16.8		± 4.0	14.7	14.7	14.7		± 4.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	72	72		± 20	250	250	250		± 30	490	490	490		± 50	
VPM	MPV	µm ³ ; fl	9.5	9.5	9.5		± 2.0	9.6	9.6	9.6		± 2.0	9.2	9.2	9.2		± 2.0	
		#	1.41	1.46	1.46		± 0.35	4.56	4.68	4.68		± 0.90	13.90	14.00	14.00		± 1.90	
		%	58.8	60.9	60.9		± 10.0	58.5	60.0	60.0		± 10.0	74.5	75.4	75.4		± 10.0	
		#	0.68	0.62	0.62		± 0.33	2.33	2.18	2.18		± 0.70	2.38	2.21	2.21		± 1.50	
		%	28.3	25.7	25.7		± 12.0	29.9	27.9	27.9		± 8.0	12.8	11.9	11.9		± 8.0	
		#	0.10	0.10	0.10		± 0.10	0.35	0.37	0.37		± 0.35	0.93	0.93	0.93		± 0.93	
		%	4.1	4.2	4.2		± 4.1	4.5	4.7	4.7		± 4.5	5.0	5.0	5.0		± 5.0	
		#	0.15	0.16	0.16		± 0.15	0.32	0.34	0.34		± 0.32	0.97	0.97	0.97		± 0.97	
		%	6.3	6.7	6.7		± 6.3	4.1	4.4	4.4		± 4.1	5.2	5.2	5.2		± 5.2	
		#	0.06	0.06	0.06		± 0.06	0.23	0.23	0.23		± 0.23	0.47	0.47	0.47		± 0.47	
		%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5	

1300027264-A
FRONT / RECTO
Ref: TEMP-0821 Rev.39