


**LOT** PX 412

**CONTROL**

 (Exp.) **2018-09-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				TOLERANCES TOLERANCE		PENTRA				TOLERANCES TOLERANCE		PENTRA				TOLERANCES TOLERANCE	
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC		DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120		DX NEXUS DF NEXUS
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5		± 0.4	7.5	7.5	7.5			± 1.0	17.0	17.0	17.0		± 2.2		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.32	2.32	2.32		± 0.16	4.63	4.63	4.63			± 0.20	5.13	5.13	5.13		± 0.25		
HB HGB	g/dl	6.8	6.8	6.8		± 0.4	13.9	13.9	13.9			± 0.5	15.8	15.8	15.8		± 0.6		
	g/l	68	68	68		± 4	139	139	139			± 5	158	158	158		± 6		
	mmol/l	4.22	4.22	4.22		± 0.25	8.63	8.63	8.63			± 0.31	9.81	9.81	9.81		± 0.37		
HT HCT	%	19.5	19.5	19.5		± 1.5	38.4	38.4	38.4			± 2.0	44.6	44.6	44.6		± 2.5		
	l/l	0.195	0.195	0.195		± 0.015	0.384	0.384	0.384			± 0.020	0.446	0.446	0.446		± 0.025		
VGM MCV	µm <sup>3</sup> ; fl	84	84	84		± 5	83	83	83			± 5	87	87	87		± 5		
TGMH MCH	pg	29.3	29.3	29.3		± 2.0	30.0	30.0	30.0			± 2.0	30.8	30.8	30.8		± 2.5		
	fmol	1.82	1.82	1.82		± 0.12	1.86	1.86	1.86			± 0.12	1.91	1.91	1.91		± 0.16		
CCMH MCHC	g/dl	34.9	34.9	34.9		± 3.0	36.2	36.2	36.2			± 3.0	35.4	35.4	35.4		± 3.0		
	g/l	349	349	349		± 30	362	362	362			± 30	354	354	354		± 30		
	mmol/l	21.67	21.67	21.67		± 1.86	22.46	22.46	22.46			± 1.86	21.98	21.98	21.98		± 1.86		
IDR RDW	%	15.4	15.4	15.4		± 4.0	15.7	15.7	15.7			± 4.0	15.0	15.0	15.0		± 4.0		
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	70	70		± 20	251	251	251			± 30	486	486	486		± 50		
VPM MPV	µm <sup>3</sup> ; fl	9.7	9.7	9.7		± 2.0	9.5	9.5	9.5			± 2.0	9.3	9.3	9.3		± 2.0		
NEUT	#	1.43	1.50	1.50		± 0.35	4.28	4.43	4.43			± 0.90	12.00	12.30	12.30		± 1.90		
	%	57.1	59.9	59.9		± 10.0	57.0	59.0	59.0			± 10.0	70.5	72.2	72.2		± 10.0		
LYMPHO	#	0.74	0.68	0.68		± 0.33	2.43	2.27	2.27			± 0.70	2.82	2.62	2.62		± 1.50		
	%	29.7	27.0	27.0		± 12.0	32.4	30.2	30.2			± 8.0	16.6	15.4	15.4		± 8.0		
MONO	#	0.10	0.10	0.10		± 0.10	0.29	0.31	0.31			± 0.29	0.71	0.71	0.71		± 0.71		
	%	3.9	4.0	4.0		± 3.9	3.9	4.1	4.1			± 3.9	4.2	4.2	4.2		± 4.2		
EOS	#	0.16	0.17	0.17		± 0.16	0.20	0.22	0.22			± 0.20	0.80	0.80	0.80		± 0.80		
	%	6.3	6.7	6.7		± 6.3	2.7	2.9	2.9			± 2.7	4.7	4.7	4.7		± 4.7		
BASO	#	0.08	0.06	0.06		± 0.06	0.30	0.29	0.29			± 0.29	0.68	0.60	0.60		± 0.60		
	%	3.0	2.4	2.4		± 2.4	4.0	3.8	3.8			± 3.8	4.0	3.5	3.5		± 3.5		

Ref: TEMP-0821 Rev.42 BACK / VERSO 1300040596