


**LOT** PX 403  
Rev 2


**CONTROL**

 (Exp.) 2017-03-05  
(YYYY - MM - DD)

		ABX Lysebio																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA							PENTRA							PENTRA						
		60 60C+ ES60	80 XL80	MS60	XLR	MSCR			60 60C+ ES60	80 XL80	MS60	XLR	MSCR			60 60C+ ES60	80 XL80	MS60	XLR	MSCR		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.5	2.4	2.5	2.5	± 0.4	7.5	7.5	7.4	7.5	7.6	± 1.0	18.6	18.2	18.2	18.2	18.7	± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.34	2.37	2.34	2.34	± 0.16	4.60	4.62	4.59	4.62	4.57	± 0.20	5.10	5.13	5.10	5.13	5.10	± 0.25			
HB HGB	g/dl	6.9	6.8	6.8	6.8	6.8	± 0.4	13.8	13.8	13.7	13.8	13.7	± 0.5	16.2	16.3	16.3	16.3	16.2	± 0.6			
	g/l	69	68	68	68	68	± 4	138	138	137	138	137	± 5	162	163	163	163	162	± 6			
	mmol/l	4.28	4.22	4.22	4.22	4.22	± 0.25	8.57	8.57	8.51	8.57	8.51	± 0.31	10.06	10.12	10.12	10.12	10.06	± 0.37			
HT HCT	%	19.0	19.4	19.0	19.4	18.8	± 1.5	37.3	37.4	37.2	37.4	37.0	± 2.0	44.4	44.6	43.9	44.6	43.9	± 2.5			
	l/l	0.190	0.194	0.190	0.194	0.188	± 0.015	0.373	0.374	0.372	0.374	0.370	± 0.020	0.444	0.446	0.439	0.446	0.439	± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	80	83	80	83	80.5	± 5	81	81	81	81	81.0	± 5	87	87	86	87	86.0	± 5			
TGMH MCH	pg	29.0	29.1	28.7	29.1	29.1	± 2.0	30.0	29.9	29.8	29.9	30.0	± 2.0	31.8	31.8	32.0	31.8	31.8	± 2.5			
	fmol	1.80	1.80	1.78	1.80	1.80	± 0.12	1.86	1.85	1.85	1.85	1.86	± 0.12	1.97	1.97	1.98	1.97	1.97	± 0.16			
CCMH MCHC	g/dl	36.2	35.0	35.9	35.0	36.1	± 3.0	37.0	36.9	36.8	36.9	37.0	± 3.0	36.5	36.5	37.2	36.5	36.9	± 3.0			
	g/l	362	350	359	350	361	± 30	370	369	368	369	370	± 30	365	365	372	365	369	± 30			
	mmol/l	22.50	21.74	22.27	21.74	22.42	± 1.86	23.00	22.90	22.88	22.90	22.98	± 1.86	22.67	22.68	23.08	22.68	22.94	± 1.86			
IDR RDW	%	12.5	12.8	12.5	12.8	11.8	± 4.0	13.1	13.2	12.5	13.2	12.5	± 4.0	12.7	13.2	12.0	13.2	12.3	± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	67	70	67	68	± 20	242	237	245	237	232	± 30	480	480	495	480	465	± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.6	10.0	9.7	10.0	9.3	± 2.0	9.5	9.8	9.6	9.8	9.0	± 2.0	9.5	9.8	9.6	9.8	9.1	± 2.0			
NEUT	#	1.42	1.47	1.39	1.47	1.50	± 0.35	4.13	4.16	4.09	4.16	4.20	± 0.90	12.91	12.92	12.78	12.92	13.46	± 1.90			
	%	59.0	58.9	57.8	58.9	59.8	± 10.0	55.1	55.5	55.3	55.5	55.3	± 10.0	69.4	71.0	70.2	71.0	72.0	± 10.0			
LYMPHO	#	0.69	0.73	0.71	0.73	0.73	± 0.33	2.54	2.50	2.53	2.50	2.60	± 0.70	2.88	2.69	2.80	2.69	2.90	± 1.50			
	%	28.8	29.0	29.5	29.0	29.3	± 12.0	33.8	33.3	34.2	33.3	34.2	± 8.0	15.5	14.8	15.4	14.8	15.5	± 8.0			
MONO	#	0.07	0.08	0.09	0.08	0.08	± 0.07	0.32	0.31	0.26	0.31	0.30	± 0.26	0.95	0.87	0.87	0.87	0.75	± 0.75			
	%	3.0	3.1	3.8	3.1	3.0	± 3.0	4.2	4.1	3.5	4.1	4.0	± 3.5	5.1	4.8	4.8	4.8	4.0	± 4.0			
EOS	#	0.14	0.14	0.13	0.14	0.11	± 0.11	0.26	0.27	0.26	0.27	0.23	± 0.23	1.04	0.93	0.95	0.93	0.80	± 0.80			
	%	5.9	5.7	5.5	5.7	4.4	± 4.4	3.4	3.6	3.5	3.6	3.0	± 3.0	5.6	5.1	5.2	5.1	4.3	± 4.3			
BASO	#	0.08	0.08	0.08	0.08	0.09	± 0.08	0.26	0.26	0.26	0.26	0.27	± 0.26	0.82	0.78	0.80	0.78	0.79	± 0.78			
	%	3.3	3.3	3.4	3.3	3.5	± 3.3	3.5	3.5	3.5	3.5	3.5	± 3.5	4.4	4.3	4.4	4.3	4.2	± 4.2			

**LOT** PX 403  
Rev 2

**CONTROL**

 (Exp.) 2017-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE		
			PENTRA					PENTRA					PENTRA							
			120	DX120	DX NEXUS			120	DX120	DX NEXUS			120	DX120	DX NEXUS					
120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS		120 RETIC	DF120	DF NEXUS										
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5			± 0.4	7.7	7.7	7.7			± 1.0	19.3	19.3	19.3			± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.37	2.37	2.37			± 0.16	4.62	4.62	4.62			± 0.20	5.21	5.21	5.21			± 0.25
		g/dl	6.9	6.9	6.9			± 0.4	13.7	13.7	13.7			± 0.5	16.1	16.1	16.1			± 0.6
HB	HGB	g/l	69	69	69			± 4	137	137	137			± 5	161	161	161			± 6
		mmol/l	4.28	4.28	4.28			± 0.25	8.51	8.51	8.51			± 0.31	10.00	10.00	10.00			± 0.37
HT	HCT	%	19.7	19.7	19.7			± 1.5	37.9	37.9	37.9			± 2.0	45.3	45.3	45.3			± 2.5
		l/l	0.197	0.197	0.197			± 0.015	0.379	0.379	0.379			± 0.020	0.453	0.453	0.453			± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	83	83	83			± 5	82	82	82			± 5	87	87	87			± 5
TGMH	MCH	pg	29.1	29.1	29.1			± 2.0	29.7	29.7	29.7			± 2.0	30.9	30.9	30.9			± 2.5
		fmol	1.81	1.81	1.81			± 0.12	1.84	1.84	1.84			± 0.12	1.92	1.92	1.92			± 0.16
		g/dl	35.1	35.1	35.1			± 3.0	36.2	36.2	36.2			± 3.0	35.5	35.5	35.5			± 3.0
CCMH	MCHC	g/l	351	351	351			± 30	362	362	362			± 30	355	355	355			± 30
		mmol/l	21.78	21.78	21.78			± 1.86	22.46	22.46	22.46			± 1.86	22.06	22.06	22.06			± 1.86
IDR	RDW	%	15.0	15.0	15.0			± 4.0	16.0	16.0	16.0			± 4.0	14.9	14.9	14.9			± 4.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	72	72			± 20	244	244	244			± 30	485	485	485			± 50
VPM	MPV	µm <sup>3</sup> ; fl	10.2	10.2	10.2			± 2.0	10.0	10.0	10.0			± 2.0	10.0	10.0	10.0			± 2.0
		#	1.49	1.53	1.53			± 0.35	4.24	4.37	4.37			± 0.90	13.70	13.90	13.90			± 1.90
NEUT		%	59.4	61.3	61.3			± 10.0	55.1	56.7	56.7			± 10.0	71.2	72.2	72.2			± 10.0
		#	0.68	0.62	0.62			± 0.33	2.49	2.33	2.33			± 0.70	2.68	2.47	2.47			± 1.50
LYMPHO		%	27.1	24.7	24.7			± 12.0	32.3	30.2	30.2			± 8.0	13.9	12.8	12.8			± 8.0
		#	0.12	0.12	0.12			± 0.12	0.49	0.50	0.50			± 0.49	1.37	1.37	1.37			± 1.37
MONO		%	4.8	4.9	4.9			± 4.8	6.3	6.5	6.5			± 6.3	7.1	7.1	7.1			± 7.1
		#	0.16	0.17	0.17			± 0.16	0.25	0.28	0.28			± 0.25	1.02	1.04	1.04			± 1.02
EOS		%	6.2	6.6	6.6			± 6.2	3.3	3.6	3.6			± 3.3	5.3	5.4	5.4			± 5.3
		#	0.06	0.06	0.06			± 0.06	0.23	0.23	0.23			± 0.23	0.48	0.48	0.48			± 0.48
BASO		%	2.5	2.5	2.5			± 2.5	3.0	3.0	3.0			± 3.0	2.5	2.5	2.5			± 2.5

1300027264-A  
FRONT / RECTO  
Ref: TEMP-0821 Rev.39