


LOT PX 413

CONTROL

 (Exp.) 2018-11-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
			CONTROL				L	CONTROL				N	CONTROL				H		
			PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.4	2.4	2.4		± 0.4	7.4	7.4	7.4		± 1.0	18.6	18.6	18.6		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38	2.38	2.38		± 0.16	4.64	4.64	4.64		± 0.20	5.25	5.25	5.25		± 0.25			
HB HGB	g/dl	7.0	7.0	7.0		± 0.4	14.0	14.0	14.0		± 0.5	16.5	16.5	16.5		± 0.6			
	g/l	70	70	70		± 4	140	140	140		± 5	165	165	165		± 6			
	mmol/l	4.35	4.35	4.35		± 0.25	8.69	8.69	8.69		± 0.31	10.25	10.25	10.25		± 0.37			
HT HCT	%	20.0	20.0	20.0		± 1.5	39.0	39.0	39.0		± 2.0	45.7	45.7	45.7		± 2.5			
	l/l	0.200	0.200	0.200		± 0.015	0.390	0.390	0.390		± 0.020	0.457	0.457	0.457		± 0.025			
VGM MCV	µm ³ ; fl	84	84	84		± 5	84	84	84		± 5	87	87	87		± 5			
TGMH MCH	pg	29.4	29.4	29.4		± 2.0	30.2	30.2	30.2		± 2.0	31.4	31.4	31.4		± 2.5			
	fmol	1.83	1.83	1.83		± 0.12	1.87	1.87	1.87		± 0.12	1.95	1.95	1.95		± 0.16			
CCMH MCHC	g/dl	35.0	35.0	35.0		± 3.0	35.9	35.9	35.9		± 3.0	36.1	36.1	36.1		± 3.0			
	g/l	350	350	350		± 30	359	359	359		± 30	361	361	361		± 30			
	mmol/l	21.74	21.74	21.74		± 1.86	22.31	22.31	22.31		± 1.86	22.43	22.43	22.43		± 1.86			
IDR RDW	%	15.5	15.5	15.5		± 4.0	16.0	16.0	16.0		± 4.0	14.8	14.8	14.8		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	75	75		± 20	252	252	252		± 30	490	490	490		± 50			
VPM MPV	µm ³ ; fl	9.0	9.0	9.0		± 2.0	8.8	8.8	8.8		± 2.0	8.8	8.8	8.8		± 2.0			
NEUT	#	1.46	1.49	1.49		± 0.35	4.39	4.40	4.40		± 0.90	13.70	13.70	13.70		± 1.90			
	%	60.8	62.1	62.1		± 10.0	59.3	59.4	59.4		± 10.0	73.6	73.6	73.6		± 10.0			
LYMPHO	#	0.62	0.56	0.56		± 0.33	2.23	2.09	2.09		± 0.70	2.72	2.51	2.51		± 1.50			
	%	25.8	23.4	23.4		± 12.0	30.1	28.2	28.2		± 8.0	14.6	13.5	13.5		± 8.0			
MONO	#	0.10	0.10	0.10		± 0.10	0.29	0.30	0.30		± 0.29	0.86	0.86	0.86		± 0.86			
	%	4.2	4.3	4.3		± 4.2	3.9	4.0	4.0		± 3.9	4.6	4.6	4.6		± 4.6			
EOS	#	0.16	0.17	0.17		± 0.16	0.27	0.30	0.30		± 0.27	0.87	0.89	0.89		± 0.87			
	%	6.7	7.1	7.1		± 6.7	3.7	4.0	4.0		± 3.7	4.7	4.8	4.8		± 4.7			
BASO	#	0.06	0.07	0.07		± 0.06	0.22	0.33	0.33		± 0.22	0.47	0.65	0.65		± 0.47			
	%	2.5	3.1	3.1		± 2.5	3.0	4.4	4.4		± 3.0	2.5	3.5	3.5		± 2.5			

Ref: TEMP-0821 Rev.42 BACK / VERSO 1300040596