


LOT PX 404
Rev 1


CONTROL

 (Exp.) 2017-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	TOLERANCES TOLERANCE		
			CONTROL					L	CONTROL					N	CONTROL						H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA							
			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP			
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	2.5	2.6	± 0.4	7.3	7.3	7.4	7.3	7.5	± 1.0	17.4	17.1	17.4	17.1	17.7	± 2.2		
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.34	2.30	2.35	2.30	2.35	± 0.16	4.58	4.60	4.60	4.60	4.62	± 0.20	5.15	5.19	5.20	5.19	5.22	± 0.25		
HB	HGB	g/dl	6.6	6.6	6.6	6.6	6.6	± 0.4	13.5	13.6	13.6	13.6	13.5	± 0.5	16.4	16.5	16.5	16.5	16.4	± 0.6		
		g/l	66	66	66	66	66	± 4	135	136	136	136	135	± 5	164	165	165	165	164	± 6		
		mmol/l	4.10	4.10	4.10	4.10	4.10	± 0.25	8.38	8.45	8.45	8.45	8.38	± 0.31	10.18	10.25	10.25	10.25	10.18	± 0.37		
HT	HCT	%	18.7	19.1	18.8	19.1	18.7	± 1.5	37.1	37.7	36.8	37.3	37.4	± 2.0	44.8	45.2	44.7	44.6	44.9	± 2.5		
		l/l	0.187	0.191	0.188	0.191	0.187	± 0.015	0.371	0.377	0.368	0.373	0.374	± 0.020	0.448	0.452	0.447	0.446	0.449	± 0.025		
VGM	MCV	µm ³ ; fl	80	83	80	83	79.5	± 5	81	82	80	81	81.0	± 5	87	87	86	86	86.0	± 5		
TGMH	MCH	pg	28.2	28.7	28.1	28.7	28.1	± 2.0	29.5	29.6	29.6	29.6	29.2	± 2.0	31.8	31.8	31.7	31.8	31.4	± 2.5		
		fmol	1.75	1.78	1.74	1.78	1.74	± 0.12	1.83	1.84	1.84	1.84	1.81	± 0.12	1.98	1.97	1.97	1.97	1.95	± 0.16		
CCMH	MCHC	g/dl	35.3	34.6	35.1	34.6	35.3	± 3.0	36.4	36.1	37.0	36.5	36.1	± 3.0	36.6	36.5	36.9	37.0	36.5	± 3.0		
		g/l	353	346	351	346	353	± 30	364	361	370	365	361	± 30	366	365	369	370	365	± 30		
		mmol/l	21.89	21.47	21.80	21.47	21.94	± 1.86	22.60	22.39	22.95	22.67	22.40	± 1.86	22.73	22.69	22.91	22.96	22.69	± 1.86		
IDR	RDW	%	12.8	13.0	12.7	13.0	12.3	± 4.0	12.5	13.0	12.6	13.0	12.5	± 4.0	12.6	13.0	12.3	13.0	12.8	± 4.0		
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	74	72	73	72	73	± 20	250	247	253	247	243	± 30	495	495	520	495	490	± 50		
VPM	MPV	µm ³ ; fl	8.8	9.0	8.9	9.0	8.3	± 2.0	9.5	9.8	9.6	9.8	9.1	± 2.0	9.1	9.3	9.1	9.3	8.7	± 2.0		
NEUT	#	#	1.39	1.37	1.41	1.37	1.46	± 0.35	4.07	4.07	4.08	4.07	4.19	± 0.90	12.21	11.83	12.23	11.83	12.53	± 1.90		
		%	55.5	54.9	56.4	54.9	56.0	± 10.0	55.8	55.7	55.2	55.7	55.8	± 10.0	70.2	69.2	70.3	69.2	70.8	± 10.0		
LYMPHO	#	#	0.79	0.80	0.79	0.80	0.84	± 0.33	2.39	2.39	2.52	2.39	2.55	± 0.70	3.03	2.98	3.10	2.98	3.20	± 1.50		
		%	31.5	31.9	31.5	31.9	32.2	± 12.0	32.8	32.7	34.0	32.7	34.0	± 8.0	17.4	17.4	17.8	17.4	18.1	± 8.0		
MONO	#	#	0.10	0.11	0.09	0.11	0.10	± 0.09	0.30	0.31	0.26	0.31	0.27	± 0.26	0.70	0.80	0.61	0.80	0.60	± 0.60		
		%	4.0	4.3	3.4	4.3	3.7	± 3.4	4.1	4.3	3.5	4.3	3.6	± 3.5	4.0	4.7	3.5	4.7	3.4	± 3.4		
EOS	#	#	0.14	0.14	0.14	0.14	0.12	± 0.12	0.27	0.28	0.28	0.28	0.25	± 0.25	0.73	0.75	0.71	0.75	0.64	± 0.64		
		%	5.6	5.7	5.4	5.7	4.7	± 4.7	3.7	3.8	3.8	3.8	3.3	± 3.3	4.2	4.4	4.1	4.4	3.6	± 3.6		
BASO	#	#	0.09	0.08	0.08	0.08	0.09	± 0.08	0.26	0.26	0.26	0.26	0.25	± 0.25	0.73	0.74	0.75	0.74	0.73	± 0.73		
		%	3.4	3.2	3.3	3.2	3.4	± 3.2	3.6	3.5	3.5	3.5	3.3	± 3.3	4.2	4.3	4.3	4.3	4.1	± 4.1		

LOT PX 404
Rev 1

CONTROL

 (Exp.) 2017-05-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA				PENTRA							
			120	DX120	DX NEXUS		120	DX120	DX NEXUS		120	DX120	DX NEXUS					
120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS	120 RETIC	DF120	DF NEXUS										
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.6	2.6	2.6		± 0.4	7.7	7.7	7.7		± 1.0	18.5	18.5	18.5		± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.35	2.35		± 0.16	4.63	4.63	4.63		± 0.20	5.28	5.28	5.28		± 0.25	
HB	HGB	g/dl	6.7	6.7	6.7		± 0.4	13.5	13.5	13.5		± 0.5	16.3	16.3	16.3		± 0.6	
		g/l	67	67	67		± 4	135	135	135		± 5	163	163	163		± 6	
		mmol/l	4.16	4.16	4.16		± 0.25	8.38	8.38	8.38		± 0.31	10.12	10.12	10.12		± 0.37	
HT	HCT	%	19.5	19.5	19.5		± 1.5	38.0	38.0	38.0		± 2.0	46.5	46.5	46.5		± 2.5	
		l/l	0.195	0.195	0.195		± 0.015	0.380	0.380	0.380		± 0.020	0.465	0.465	0.465		± 0.025	
VGM	MCV	µm ³ ; fl	83	83	83		± 5	82	82	82		± 5	88	88	88		± 5	
TGMH	MCH	pg	28.5	28.5	28.5		± 2.0	29.2	29.2	29.2		± 2.0	30.9	30.9	30.9		± 2.5	
		fmol	1.77	1.77	1.77		± 0.12	1.81	1.81	1.81		± 0.12	1.92	1.92	1.92		± 0.16	
CCMH	MCHC	g/dl	34.4	34.4	34.4		± 3.0	35.6	35.6	35.6		± 3.0	35.1	35.1	35.1		± 3.0	
		g/l	344	344	344		± 30	356	356	356		± 30	351	351	351		± 30	
		mmol/l	21.33	21.33	21.33		± 1.86	22.08	22.08	22.08		± 1.86	21.79	21.79	21.79		± 1.86	
IDR	RDW	%	14.8	14.8	14.8		± 4.0	16.2	16.2	16.2		± 4.0	15.4	15.4	15.4		± 4.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	72	72		± 20	252	252	252		± 30	495	495	495		± 50	
VPM	MPV	µm ³ ; fl	9.1	9.1	9.1		± 2.0	9.7	9.7	9.7		± 2.0	9.1	9.1	9.1		± 2.0	
NEUT	#	#	1.43	1.49	1.49		± 0.35	4.24	4.37	4.37		± 0.90	12.90	13.20	13.20		± 1.90	
		%	54.9	57.3	57.3		± 10.0	55.1	56.7	56.7		± 10.0	69.6	71.0	71.0		± 10.0	
LYMPHO	#	#	0.82	0.75	0.75		± 0.33	2.45	2.29	2.29		± 0.70	3.15	2.90	2.90		± 1.50	
		%	31.6	28.7	28.7		± 12.0	31.8	29.7	29.7		± 8.0	17.0	15.7	15.7		± 8.0	
MONO	#	#	0.14	0.14	0.14		± 0.14	0.47	0.49	0.49		± 0.47	1.17	1.15	1.15		± 1.15	
		%	5.3	5.4	5.4		± 5.3	6.1	6.3	6.3		± 6.1	6.3	6.2	6.2		± 6.2	
EOS	#	#	0.15	0.16	0.16		± 0.15	0.31	0.33	0.33		± 0.31	0.85	0.85	0.85		± 0.85	
		%	5.7	6.1	6.1		± 5.7	4.0	4.3	4.3		± 4.0	4.6	4.6	4.6		± 4.6	
BASO	#	#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.46	0.46	0.46		± 0.46	
		%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5	

1300032435-A
FRONT / RECTO
Ref: TEMP-0821 Rev.40