



(Exp.) **2019-01-05**
 (YYYY-MM-DD)

LOT		PX 414		CONTROL																		
		Rev 2		ABX Lysebio																		
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L		TOLERANCES TOLERANCE	CONTROL				N		TOLERANCES TOLERANCE	CONTROL				H		TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA												
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC		DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120		DX NEXUS DF NEXUS						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5			± 0.4	8.1	8.1	8.1			± 1.0	19.2	19.2	19.2			± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41	2.41			± 0.16	4.70	4.70	4.70			± 0.20	5.24	5.24	5.24			± 0.25			
HB HGB	g/dl	7.1	7.1	7.1			± 0.4	13.9	13.9	13.9			± 0.5	16.2	16.2	16.2			± 0.6			
	g/l	71	71	71			± 4	139	139	139			± 5	162	162	162			± 6			
	mmol/l	4.41	4.41	4.41			± 0.25	8.63	8.63	8.63			± 0.31	10.06	10.06	10.06			± 0.37			
HT HCT	%	20.2	20.2	20.2			± 1.5	38.5	38.5	38.5			± 2.0	45.6	45.6	45.6			± 2.5			
	l/l	0.202	0.202	0.202			± 0.015	0.385	0.385	0.385			± 0.020	0.456	0.456	0.456			± 0.025			
VGM MCV	µm ³ ; fl	84	84	84			± 5	82	82	82			± 5	87	87	87			± 5			
TGMH MCH	pg	29.5	29.5	29.5			± 2.0	29.6	29.6	29.6			± 2.0	30.9	30.9	30.9			± 2.5			
	fmol	1.83	1.83	1.83			± 0.12	1.84	1.84	1.84			± 0.12	1.92	1.92	1.92			± 0.16			
CCMH MCHC	g/dl	35.1	35.1	35.1			± 3.0	36.1	36.1	36.1			± 3.0	35.5	35.5	35.5			± 3.0			
	g/l	351	351	351			± 30	361	361	361			± 30	355	355	355			± 30			
	mmol/l	21.78	21.78	21.78			± 1.86	22.40	22.40	22.40			± 1.86	22.07	22.07	22.07			± 1.86			
IDR RDW	%	15.5	15.5	15.5			± 4.0	16.2	16.2	16.2			± 4.0	15.2	15.2	15.2			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	73	73	73			± 20	252	252	252			± 30	490	490	490			± 50			
VPM MPV	µm ³ ; fl	10.3	10.3	10.3			± 2.0	10.3	10.3	10.3			± 2.0	9.8	9.8	9.8			± 2.0			
NEUT	#	1.41	1.45	1.45			± 0.35	4.63	4.71	4.71			± 0.90	14.20	14.20	14.20			± 1.90			
	%	56.2	58.1	58.1			± 10.0	57.1	58.2	58.2			± 10.0	73.7	73.5	73.5			± 10.0			
LYMPHO	#	0.73	0.68	0.68			± 0.33	2.62	2.44	2.44			± 0.70	2.86	2.63	2.63			± 1.50			
	%	29.2	27.2	27.2			± 12.0	32.3	30.1	30.1			± 8.0	14.9	13.7	13.7			± 8.0			
MONO	#	0.12	0.11	0.11			± 0.11	0.33	0.33	0.33			± 0.33	0.75	0.75	0.75			± 0.75			
	%	4.9	4.5	4.5			± 4.5	4.1	4.1	4.1			± 4.1	3.9	3.9	3.9			± 3.9			
EOS	#	0.18	0.20	0.20			± 0.18	0.28	0.33	0.33			± 0.28	0.96	1.08	1.08			± 0.96			
	%	7.2	7.8	7.8			± 7.2	3.5	4.1	4.1			± 3.5	5.0	5.6	5.6			± 5.0			
BASO	#	0.06	0.06	0.06			± 0.06	0.24	0.28	0.28			± 0.24	0.48	0.63	0.63			± 0.48			
	%	2.5	2.4	2.4			± 2.4	3.0	3.5	3.5			± 3.0	2.5	3.3	3.3			± 2.5			

1300040596
 Ref: TEMP-0821 Rev.42 BACK / VERSO