


LOT PX 411

CONTROL

 (Exp.) 2018-07-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
		CONTROL				L	CONTROL				N	CONTROL				H		
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.8	7.8	7.8			± 1.0	18.5	18.5	18.5		± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.38	2.38	2.38		± 0.16	4.68	4.68	4.68			± 0.20	5.27	5.27	5.27		± 0.25	
HB HGB	g/dl	6.9	6.9	6.9		± 0.4	13.9	13.9	13.9			± 0.5	16.2	16.2	16.2		± 0.6	
	g/l	69	69	69		± 4	139	139	139			± 5	162	162	162		± 6	
	mmol/l	4.28	4.28	4.28		± 0.25	8.63	8.63	8.63			± 0.31	10.06	10.06	10.06		± 0.37	
HT HCT	%	19.8	19.8	19.8		± 1.5	38.8	38.8	38.8			± 2.0	45.8	45.8	45.8		± 2.5	
	l/l	0.198	0.198	0.198		± 0.015	0.388	0.388	0.388			± 0.020	0.458	0.458	0.458		± 0.025	
VGM MCV	µm ³ ; fl	83	83	83		± 5	83	83	83			± 5	87	87	87		± 5	
TGMH MCH	pg	29.0	29.0	29.0		± 2.0	29.7	29.7	29.7			± 2.0	30.7	30.7	30.7		± 2.5	
	fmol	1.80	1.80	1.80		± 0.12	1.84	1.84	1.84			± 0.12	1.91	1.91	1.91		± 0.16	
CCMH MCHC	g/dl	34.9	34.9	34.9		± 3.0	35.8	35.8	35.8			± 3.0	35.3	35.3	35.3		± 3.0	
	g/l	349	349	349		± 30	358	358	358			± 30	353	353	353		± 30	
	mmol/l	21.69	21.69	21.69		± 1.86	22.22	22.22	22.22			± 1.86	21.94	21.94	21.94		± 1.86	
IDR RDW	%	15.5	15.5	15.5		± 4.0	16.0	16.0	16.0			± 4.0	14.8	14.8	14.8		± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	65	65	65		± 20	250	250	250			± 30	495	495	495		± 50	
VPM MPV	µm ³ ; fl	10.6	10.6	10.6		± 2.0	9.7	9.7	9.7			± 2.0	9.6	9.6	9.6		± 2.0	
NEUT	#	1.36	1.42	1.42		± 0.35	4.52	4.64	4.64			± 0.90	13.30	13.50	13.50		± 1.90	
	%	54.4	56.6	56.6		± 10.0	58.0	59.5	59.5			± 10.0	71.9	73.0	73.0		± 10.0	
LYMPHO	#	0.75	0.68	0.68		± 0.33	2.35	2.20	2.20			± 0.70	2.72	2.52	2.52		± 1.50	
	%	30.0	27.3	27.3		± 12.0	30.1	28.2	28.2			± 8.0	14.7	13.6	13.6		± 8.0	
MONO	#	0.15	0.15	0.15		± 0.15	0.42	0.44	0.44			± 0.42	1.11	1.09	1.09		± 1.09	
	%	6.0	6.1	6.1		± 6.0	5.4	5.6	5.6			± 5.4	6.0	5.9	5.9		± 5.9	
EOS	#	0.17	0.18	0.18		± 0.17	0.20	0.21	0.21			± 0.20	0.63	0.65	0.65		± 0.63	
	%	6.6	7.0	7.0		± 6.6	2.5	2.7	2.7			± 2.5	3.4	3.5	3.5		± 3.4	
BASO	#	0.08	0.08	0.08		± 0.08	0.31	0.31	0.31			± 0.31	0.74	0.74	0.74		± 0.74	
	%	3.0	3.0	3.0		± 3.0	4.0	4.0	4.0			± 4.0	4.0	4.0	4.0		± 4.0	

1300040596
BACK / VERSO
Ref. TEMP-0821 Rev.42