

ABX Difftrol



LOT PX 415
Rev 2

CONTROL

(Exp.) 2019-03-05
(YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC		DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120		DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS	
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.7	7.7	7.7			± 1.0	19.1	19.1	19.1		± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41	2.41		± 0.16	4.63	4.63	4.63			± 0.20	5.23	5.23	5.23		± 0.25		
HB HGB	g/dl	7.0	7.0	7.0		± 0.4	13.6	13.6	13.6			± 0.5	15.9	15.9	15.9		± 0.6		
	g/l	70	70	70		± 4	136	136	136			± 5	159	159	159		± 6		
HT HCT	mmol/l	4.35	4.35	4.35		± 0.25	8.45	8.45	8.45			± 0.31	9.87	9.87	9.87		± 0.37		
	%	20.2	20.2	20.2		± 1.5	38.0	38.0	38.0			± 2.0	45.5	45.5	45.5		± 2.5		
VGM MCV	l/l	0.202	0.202	0.202		± 0.015	0.380	0.380	0.380			± 0.020	0.455	0.455	0.455		± 0.025		
	µm ³ ·fl	84	84	84		± 5	82	82	82			± 5	87	87	87		± 5		
TGMH MCH	pg	29.0	29.0	29.0		± 2.0	29.4	29.4	29.4			± 2.0	30.4	30.4	30.4		± 2.5		
	fmoi	1.80	1.80	1.80		± 0.12	1.82	1.82	1.82			± 0.12	1.89	1.89	1.89		± 0.16		
CCMH MCHC	g/dl	34.6	34.6	34.6		± 3.0	35.8	35.8	35.8			± 3.0	34.9	34.9	34.9		± 3.0		
	g/l	346	346	346		± 30	358	358	358			± 30	349	349	349		± 30		
	mmol/l	21.47	21.47	21.47		± 1.86	22.25	22.25	22.25			± 1.86	21.70	21.70	21.70		± 1.86		
IDR RDW	%	15.2	15.2	15.2		± 4.0	16.0	16.0	16.0			± 4.0	15.0	15.0	15.0		± 4.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	70		± 20	255	255	255			± 30	500	500	500		± 50		
VPM MPV	µm ³ ·fl	9.0	9.0	9.0		± 2.0	9.3	9.3	9.3			± 2.0	9.6	9.6	9.6		± 2.0		
NEUT	#	1.43	1.48	1.48		± 0.35	4.50	4.62	4.62			± 0.90	13.90	14.10	14.10		± 1.90		
	%	57.0	59.0	59.0		± 10.0	58.4	60.0	60.0			± 10.0	73.0	74.0	74.0		± 10.0		
LYMPHO	#	0.72	0.66	0.66		± 0.33	2.39	2.24	2.24			± 0.70	2.69	2.52	2.52		± 1.50		
	%	28.7	26.2	26.2		± 12.0	31.0	29.1	29.1			± 8.0	14.1	13.2	13.2		± 8.0		
MONO	#	0.09	0.09	0.09		± 0.09	0.28	0.29	0.29			± 0.28	0.82	0.80	0.80		± 0.80		
	%	3.5	3.5	3.5		± 3.5	3.7	3.8	3.8			± 3.7	4.3	4.2	4.2		± 4.2		
EOS	#	0.20	0.21	0.21		± 0.20	0.26	0.28	0.28			± 0.26	0.97	0.97	0.97		± 0.97		
	%	7.8	8.3	8.3		± 7.8	3.4	3.6	3.6			± 3.4	5.1	5.1	5.1		± 5.1		
BASO	#	0.08	0.08	0.08		± 0.08	0.27	0.27	0.27			± 0.27	0.67	0.67	0.67		± 0.67		
	%	3.0	3.0	3.0		± 3.0	3.5	3.5	3.5			± 3.5	3.5	3.5	3.5		± 3.5		

Ref: TEMP-0821 Rev.42 BACK / VERSO 1300040596