


LOT PX 425

CONTROL

 (Exp.) 2020-11-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H				
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5	± 0.4	7.9	7.9	7.9	± 1.0	19.3	19.3	19.3	± 2.2					
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.29	2.29	2.29	± 0.16	4.60	4.60	4.60	± 0.20	5.17	5.17	5.17	± 0.25					
HB HGB	g/dl	6.6	6.6	6.6	± 0.4	13.9	13.9	13.9	± 0.5	16.1	16.1	16.1	± 0.6					
	g/l	66	66	66	± 4	139	139	139	± 5	161	161	161	± 6					
	mmol/l	4.10	4.10	4.10	± 0.25	8.63	8.63	8.63	± 0.31	10.00	10.00	10.00	± 0.37					
HT HCT	%	18.8	18.8	18.8	± 1.5	38.2	38.2	38.2	± 2.0	45.0	45.0	45.0	± 2.5					
	l/l	0.188	0.188	0.188	± 0.015	0.382	0.382	0.382	± 0.020	0.450	0.450	0.450	± 0.025					
VGM MCV	µm ³ ; fl	82	82	82	± 5	83	83	83	± 5	87	87	87	± 5					
TGMH MCH	pg	28.8	28.8	28.8	± 2.0	30.2	30.2	30.2	± 2.0	31.1	31.1	31.1	± 2.5					
	fmol	1.79	1.79	1.79	± 0.12	1.88	1.88	1.88	± 0.12	1.93	1.93	1.93	± 0.16					
CCMH MCHC	g/dl	35.1	35.1	35.1	± 3.0	36.4	36.4	36.4	± 3.0	35.8	35.8	35.8	± 3.0					
	g/l	351	351	351	± 30	364	364	364	± 30	358	358	358	± 30					
	mmol/l	21.83	21.83	21.83	± 1.86	22.61	22.61	22.61	± 1.86	22.23	22.23	22.23	± 1.86					
IDR RDW	%	15.5	15.5	15.5	± 4.0	15.0	15.0	15.0	± 4.0	14.7	14.7	14.7	± 4.0					
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	70	70	70	± 20	262	262	262	± 30	490	490	490	± 50					
VPM MPV	µm ³ ; fl	9.2	9.2	9.2	± 2.0	9.5	9.5	9.5	± 2.0	9.2	9.2	9.2	± 2.0					
NEUT	#	1.40	1.43	1.43	± 0.35	4.58	4.58	4.58	± 0.90	13.50	13.60	13.60	± 1.90					
	%	56.0	57.3	57.3	± 10.0	58.0	58.0	58.0	± 10.0	70.0	70.5	70.5	± 10.0					
LYMPHO	#	0.73	0.68	0.68	± 0.33	2.35	2.29	2.29	± 0.70	3.11	2.91	2.91	± 1.50					
	%	29.0	27.3	27.3	± 12.0	29.7	29.0	29.0	± 8.0	16.1	15.1	15.1	± 8.0					
MONO	#	0.11	0.12	0.12	± 0.11	0.42	0.47	0.47	± 0.42	1.16	1.22	1.22	± 1.16					
	%	4.5	4.7	4.7	± 4.5	5.3	6.0	6.0	± 5.3	6.0	6.3	6.3	± 6.0					
EOS	#	0.19	0.19	0.19	± 0.19	0.30	0.30	0.30	± 0.30	0.93	0.97	0.97	± 0.93					
	%	7.5	7.7	7.7	± 7.5	3.8	3.8	3.8	± 3.8	4.8	5.0	5.0	± 4.8					
BASO	#	0.08	0.08	0.08	± 0.08	0.25	0.25	0.25	± 0.25	0.60	0.60	0.60	± 0.60					
	%	3.0	3.0	3.0	± 3.0	3.2	3.2	3.2	± 3.2	3.1	3.1	3.1	± 3.1					

Ref. TEMP-0821 Rev.44 BACK / VERSO 1300040596