

 (Exp.) **2021-01-05**
(YYYY-MM-DD)

LOT PX 426

CONTROL

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				120 RETIC		PENTRA				120 RETIC		PENTRA					
		120	DX120	DX NEXUS	DF NEXUS		120	DX120	DX NEXUS	DF NEXUS	120		DX120	DX NEXUS	DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8	2.8		± 0.4	8.2	8.2	8.2		± 1.0	19.4	19.4	19.4		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.27	2.27	2.27		± 0.16	4.53	4.53	4.53		± 0.20	5.08	5.08	5.08		± 0.25			
HB HGB	g/dl	6.5	6.5	6.5		± 0.4	13.6	13.6	13.6		± 0.5	16.2	16.2	16.2		± 0.6			
	g/l	65	65	65		± 4	136	136	136		± 5	162	162	162		± 6			
	mmol/l	4.04	4.04	4.04		± 0.25	8.45	8.45	8.45		± 0.31	10.06	10.06	10.06		± 0.37			
HT HCT	%	18.6	18.6	18.6		± 1.5	37.6	37.6	37.6		± 2.0	45.2	45.2	45.2		± 2.5			
	l/l	0.186	0.186	0.186		± 0.015	0.376	0.376	0.376		± 0.020	0.452	0.452	0.452		± 0.025			
VGM MCV	µm ³ ; fl	82	82	82		± 5	83	83	83		± 5	89	89	89		± 5			
TGMH MCH	pg	28.6	28.6	28.6		± 2.0	30.0	30.0	30.0		± 2.0	31.9	31.9	31.9		± 2.5			
	fmol	1.78	1.78	1.78		± 0.12	1.86	1.86	1.86		± 0.12	1.98	1.98	1.98		± 0.16			
CCMH MCHC	g/dl	34.9	34.9	34.9		± 3.0	36.2	36.2	36.2		± 3.0	35.8	35.8	35.8		± 3.0			
	g/l	349	349	349		± 30	362	362	362		± 30	358	358	358		± 30			
	mmol/l	21.69	21.69	21.69		± 1.86	22.46	22.46	22.46		± 1.86	22.25	22.25	22.25		± 1.86			
IDR RDW	%	15.5	15.5	15.5		± 4.0	15.0	15.0	15.0		± 4.0	14.2	14.2	14.2		± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	69	69	69		± 20	248	248	248		± 30	487	487	487		± 50			
VPM MPV	µm ³ ; fl	9.7	9.7	9.7		± 2.0	8.8	8.8	8.8		± 2.0	8.6	8.6	8.6		± 2.0			
NEUT	#	1.46	1.51	1.51		± 0.35	4.34	4.44	4.44		± 0.90	13.40	13.60	13.60		± 1.90			
	%	52.1	54.0	54.0		± 10.0	52.9	54.1	54.1		± 10.0	68.8	69.9	69.9		± 10.0			
LYMPHO	#	0.80	0.72	0.72		± 0.33	2.38	2.22	2.22		± 0.70	2.83	2.62	2.62		± 1.50			
	%	28.4	25.8	25.8		± 12.0	29.0	27.1	27.1		± 8.0	14.6	13.5	13.5		± 8.0			
MONO	#	0.25	0.25	0.25		± 0.25	0.83	0.85	0.85		± 0.83	1.94	1.94	1.94		± 1.94			
	%	8.9	9.1	9.1		± 8.9	10.1	10.4	10.4		± 10.1	10.0	10.0	10.0		± 10.0			
EOS	#	0.21	0.22	0.22		± 0.21	0.41	0.44	0.44		± 0.41	0.72	0.72	0.72		± 0.72			
	%	7.5	8.0	8.0		± 7.5	5.0	5.4	5.4		± 5.0	3.7	3.7	3.7		± 3.7			
BASO	#	0.09	0.09	0.09		± 0.09	0.25	0.25	0.25		± 0.25	0.56	0.56	0.56		± 0.56			
	%	3.1	3.1	3.1		± 3.1	3.0	3.0	3.0		± 3.0	2.9	2.9	2.9		± 2.9			

Ref. TEMP-0821 Rev.44 BACK / VERSO 1300040596