


LOT PX 427

CONTROL

 (Exp.) 2021-03-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				TOLERANCES TOLERANCE		PENTRA				TOLERANCES TOLERANCE		PENTRA				TOLERANCES TOLERANCE	
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC		DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120		DX NEXUS DF NEXUS
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.7	2.7	2.7		± 0.4	8.1	8.1	8.1			± 1.0	19.5	19.5	19.5		± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.34	2.34	2.34		± 0.16	4.54	4.54	4.54			± 0.20	5.20	5.20	5.20		± 0.25		
HB HGB	g/dl	6.7	6.7	6.7		± 0.4	13.6	13.6	13.6			± 0.5	16.7	16.7	16.7		± 0.6		
	g/l	67	67	67		± 4	136	136	136			± 5	167	167	167		± 6		
	mmol/l	4.16	4.16	4.16		± 0.25	8.45	8.45	8.45			± 0.31	10.37	10.37	10.37		± 0.37		
HT HCT	%	19.4	19.4	19.4		± 1.5	38.1	38.1	38.1			± 2.0	46.8	46.8	46.8		± 2.5		
	l/l	0.194	0.194	0.194		± 0.015	0.381	0.381	0.381			± 0.020	0.468	0.468	0.468		± 0.025		
VGM MCV	µm ³ ; fl	83	83	83		± 5	84	84	84			± 5	90	90	90		± 5		
TGMH MCH	pg	28.6	28.6	28.6		± 2.0	30.0	30.0	30.0			± 2.0	32.1	32.1	32.1		± 2.5		
	fmol	1.78	1.78	1.78		± 0.12	1.86	1.86	1.86			± 0.12	1.99	1.99	1.99		± 0.16		
CCMH MCHC	g/dl	34.5	34.5	34.5		± 3.0	35.7	35.7	35.7			± 3.0	35.7	35.7	35.7		± 3.0		
	g/l	345	345	345		± 30	357	357	357			± 30	357	357	357		± 30		
	mmol/l	21.42	21.42	21.42		± 1.86	22.15	22.15	22.15			± 1.86	22.16	22.16	22.16		± 1.86		
IDR RDW	%	14.0	14.0	14.0		± 4.0	16.5	16.5	16.5			± 4.0	14.5	14.5	14.5		± 4.0		
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72	72		± 20	255	255	255			± 30	485	485	485		± 50		
VPM MPV	µm ³ ; fl	8.8	8.8	8.8		± 2.0	8.5	8.5	8.5			± 2.0	9.1	9.1	9.1		± 2.0		
NEUT	#	1.40	1.46	1.46		± 0.35	4.10	4.23	4.23			± 0.90	13.10	13.30	13.30		± 1.90		
	%	51.8	54.0	54.0		± 10.0	50.6	52.2	52.2			± 10.0	66.8	68.0	68.0		± 10.0		
LYMPHO	#	0.83	0.76	0.76		± 0.33	2.69	2.51	2.51			± 0.70	3.02	2.81	2.81		± 1.50		
	%	30.9	28.1	28.1		± 12.0	33.2	31.0	31.0			± 8.0	15.5	14.4	14.4		± 8.0		
MONO	#	0.23	0.23	0.23		± 0.23	0.82	0.84	0.84			± 0.82	1.93	1.91	1.91		± 1.91		
	%	8.5	8.7	8.7		± 8.5	10.1	10.4	10.4			± 10.1	9.9	9.8	9.8		± 9.8		
EOS	#	0.17	0.18	0.18		± 0.17	0.26	0.28	0.28			± 0.26	0.98	0.98	0.98		± 0.98		
	%	6.2	6.6	6.6		± 6.2	3.2	3.5	3.5			± 3.2	5.0	5.0	5.0		± 5.0		
BASO	#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23			± 0.23	0.55	0.55	0.55		± 0.55		
	%	2.6	2.6	2.6		± 2.6	2.9	2.9	2.9			± 2.9	2.8	2.8	2.8		± 2.8		

1300040596
BACK / VERSO
Ref. TEMP-0821 Rev.44