


LOT PX 418

CONTROL

 (Exp.) **2019-09-05**
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																TOLERANCES TOLERANCE
		CONTROL				L	CONTROL				N	CONTROL				H		
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE		
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.5	2.5	2.5		± 0.4	7.7	7.7	7.7		± 1.0	19.1	19.1	19.1		± 2.2		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.33	2.33		± 0.16	4.61	4.61	4.61		± 0.20	5.26	5.26	5.26		± 0.25		
HB HGB	g/dl	6.9	6.9	6.9		± 0.4	13.9	13.9	13.9		± 0.5	16.6	16.6	16.6		± 0.6		
	g/l	69	69	69		± 4	139	139	139		± 5	166	166	166		± 6		
	mmol/l	4.28	4.28	4.28		± 0.25	8.63	8.63	8.63		± 0.31	10.31	10.31	10.31		± 0.37		
HT HCT	%	19.8	19.8	19.8		± 1.5	38.7	38.7	38.7		± 2.0	46.8	46.8	46.8		± 2.5		
	l/l	0.198	0.198	0.198		± 0.015	0.387	0.387	0.387		± 0.020	0.468	0.468	0.468		± 0.025		
VGM MCV	µm ³ ; fl	85	85	85		± 5	84	84	84		± 5	89	89	89		± 5		
TGMH MCH	pg	29.6	29.6	29.6		± 2.0	30.2	30.2	30.2		± 2.0	31.6	31.6	31.6		± 2.5		
	fmol	1.84	1.84	1.84		± 0.12	1.87	1.87	1.87		± 0.12	1.96	1.96	1.96		± 0.16		
CCMH MCHC	g/dl	34.8	34.8	34.8		± 3.0	35.9	35.9	35.9		± 3.0	35.5	35.5	35.5		± 3.0		
	g/l	348	348	348		± 30	359	359	359		± 30	355	355	355		± 30		
	mmol/l	21.64	21.64	21.64		± 1.86	22.29	22.29	22.29		± 1.86	22.02	22.02	22.02		± 1.86		
IDR RDW	%	15.0	15.0	15.0		± 4.0	16.0	16.0	16.0		± 4.0	15.5	15.5	15.5		± 4.0		
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	72	72		± 20	255	255	255		± 30	510	510	510		± 50		
VPM MPV	µm ³ ; fl	9.8	9.8	9.8		± 2.0	9.5	9.5	9.5		± 2.0	9.5	9.5	9.5		± 2.0		
NEUT	#	1.50	1.54	1.54		± 0.35	4.28	4.41	4.41		± 0.90	13.40	13.60	13.60		± 1.90		
	%	59.8	61.7	61.7		± 10.0	55.6	57.3	57.3		± 10.0	70.1	71.3	71.3		± 10.0		
LYMPHO	#	0.64	0.58	0.58		± 0.33	2.38	2.22	2.22		± 0.70	2.87	2.65	2.65		± 1.50		
	%	25.7	23.3	23.3		± 12.0	30.9	28.8	28.8		± 8.0	15.0	13.9	13.9		± 8.0		
MONO	#	0.13	0.14	0.14		± 0.13	0.48	0.49	0.49		± 0.48	1.22	1.20	1.20		± 1.20		
	%	5.3	5.4	5.4		± 5.3	6.2	6.4	6.4		± 6.2	6.4	6.3	6.3		± 6.3		
EOS	#	0.16	0.17	0.17		± 0.16	0.25	0.26	0.26		± 0.25	0.88	0.88	0.88		± 0.88		
	%	6.4	6.8	6.8		± 6.4	3.2	3.4	3.4		± 3.2	4.6	4.6	4.6		± 4.6		
BASO	#	0.07	0.07	0.07		± 0.07	0.32	0.32	0.32		± 0.32	0.74	0.74	0.74		± 0.74		
	%	2.8	2.8	2.8		± 2.8	4.1	4.1	4.1		± 4.1	3.9	3.9	3.9		± 3.9		

Ref: TEMP-0821 Rev.43 BACK / VERSO 1300040596