


**LOT** PX 421

**CONTROL**

 (Exp.) 2020-03-05  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC		DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.7	2.7	2.7		± 0.4	8.0	8.0	8.0		± 1.0	19.2	19.2	19.2		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.41	2.41	2.41		± 0.16	4.48	4.48	4.48		± 0.20	5.19	5.19	5.19		± 0.25			
HB HGB	g/dl	7.0	7.0	7.0		± 0.4	13.2	13.2	13.2		± 0.5	16.4	16.4	16.4		± 0.6			
	g/l	70	70	70		± 4	132	132	132		± 5	164	164	164		± 6			
	mmol/l	4.35	4.35	4.35		± 0.25	8.20	8.20	8.20		± 0.31	10.18	10.18	10.18		± 0.37			
HT HCT	%	20.0	20.0	20.0		± 1.5	36.3	36.3	36.3		± 2.0	45.2	45.2	45.2		± 2.5			
	l/l	0.200	0.200	0.200		± 0.015	0.363	0.363	0.363		± 0.020	0.452	0.452	0.452		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	83	83	83		± 5	81	81	81		± 5	87	87	87		± 5			
TGMH MCH	pg	29.0	29.0	29.0		± 2.0	29.5	29.5	29.5		± 2.0	31.6	31.6	31.6		± 2.5			
	fmol	1.80	1.80	1.80		± 0.12	1.83	1.83	1.83		± 0.12	1.96	1.96	1.96		± 0.16			
CCMH MCHC	g/dl	35.0	35.0	35.0		± 3.0	36.4	36.4	36.4		± 3.0	36.3	36.3	36.3		± 3.0			
	g/l	350	350	350		± 30	364	364	364		± 30	363	363	363		± 30			
	mmol/l	21.73	21.73	21.73		± 1.86	22.59	22.59	22.59		± 1.86	22.56	22.56	22.56		± 1.86			
IDR RDW	%	14.5	14.5	14.5		± 4.0	15.5	15.5	15.5		± 4.0	14.5	14.5	14.5		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	73	73		± 20	245	245	245		± 30	500	500	500		± 50			
VPM MPV	µm <sup>3</sup> ; fl	10.1	10.1	10.1		± 2.0	9.8	9.8	9.8		± 2.0	9.2	9.2	9.2		± 2.0			
NEUT	#	1.47	1.53	1.53		± 0.35	4.78	4.90	4.90		± 0.90	13.50	13.70	13.70		± 1.90			
	%	54.6	56.7	56.7		± 10.0	59.8	61.2	61.2		± 10.0	70.1	71.4	71.4		± 10.0			
LYMPHO	#	0.81	0.74	0.74		± 0.33	2.35	2.20	2.20		± 0.70	3.00	2.76	2.76		± 1.50			
	%	30.0	27.3	27.3		± 12.0	29.4	27.5	27.5		± 8.0	15.6	14.4	14.4		± 8.0			
MONO	#	0.14	0.14	0.14		± 0.14	0.33	0.34	0.34		± 0.33	0.96	0.94	0.94		± 0.94			
	%	5.2	5.3	5.3		± 5.2	4.1	4.3	4.3		± 4.1	5.0	4.9	4.9		± 4.9			
EOS	#	0.21	0.22	0.22		± 0.21	0.30	0.33	0.33		± 0.30	1.13	1.13	1.13		± 1.13			
	%	7.7	8.2	8.2		± 7.7	3.8	4.1	4.1		± 3.8	5.9	5.9	5.9		± 5.9			
BASO	#	0.07	0.07	0.07		± 0.07	0.23	0.23	0.23		± 0.23	0.65	0.65	0.65		± 0.65			
	%	2.5	2.5	2.5		± 2.5	2.9	2.9	2.9		± 2.9	3.4	3.4	3.4		± 3.4			

Ref. TEMP-0821 Rev.43 BACK / VERSO 1300040596