

**LOT PX 073**

**CONTROL**

(Exp.) **2013-09-05**  
 (YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA						PENTRA						PENTRA					
			60 60C+ ES60	80 XL80	MS60	XL80	60 60C+ ES60	80 XL80	MS60	XL80	60 60C+ ES60	80 XL80	MS60	XL80	60 60C+ ES60	80 XL80	MS60	XL80		
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.5	2.4		± 0.4	7.3	7.3	7.3	7.3		± 1.0	17.8	17.5	17.9	17.5		± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.42	2.37	2.39	2.37		± 0.12	4.58	4.58	4.58	4.58		± 0.15	5.10	5.11	5.08	5.11		± 0.20
HB	HGB	g/dl	6.8	6.8	6.8	6.8		± 0.4	13.3	13.3	13.3	13.3		± 0.5	16.1	16.1	16.1	16.1		± 0.6
		g/l	68	68	68	68		± 4	133	133	133	133		± 5	161	161	161	161		± 6
		mmol/l	4.22	4.22	4.22	4.22		± 0.25	8.26	8.26	8.26	8.26		± 0.31	10.00	10.00	10.00	10.00		± 0.37
HT	HCT	%	19.1	19.7	18.6	19.7		± 1.5	36.2	36.6	35.7	36.6		± 2.0	44.4	44.5	43.2	44.5		± 2.5
		l/l	0.191	0.197	0.186	0.197		± 0.015	0.362	0.366	0.357	0.366		± 0.020	0.444	0.445	0.432	0.445		± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	79	83	78	83		± 5	79	80	78	80		± 5	87	87	85	87		± 5
TGMH	MCH	pg	28.1	28.7	28.5	28.7		± 2.0	29.0	29.0	29.0	29.0		± 2.0	31.6	31.5	31.7	31.5		± 2.5
		fmol	1.74	1.78	1.77	1.78		± 0.12	1.80	1.80	1.80	1.80		± 0.12	1.96	1.96	1.97	1.96		± 0.16
CCMH	MCHC	g/dl	35.6	34.6	36.5	34.6		± 3.0	36.8	36.3	37.2	36.3		± 3.0	36.3	36.2	37.3	36.2		± 3.0
		g/l	356	346	365	346		± 30	368	363	372	363		± 30	363	362	373	362		± 30
		mmol/l	22.09	21.47	22.65	21.47		± 1.86	22.83	22.54	23.12	22.54		± 1.86	22.53	22.49	23.15	22.49		± 1.86
IDR	RDW	%	12.7	13.7	12.5	13.7		± 4.0	12.5	13.6	12.5	13.6		± 4.0	12.3	13.4	11.7	13.4		± 4.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	72	70	72		± 20	250	245	245	245		± 30	510	505	510	505		± 50
VPM	MPV	µm <sup>3</sup> ; fl	9.2	9.4	9.6	9.4		± 2.0	9.4	9.6	9.6	9.6		± 2.0	9.3	9.6	9.5	9.6		± 2.0
NEUT	#		1.36	1.36	1.38	1.36		± 0.30	4.23	4.25	4.20	4.25		± 0.80	12.82	12.60	12.89	12.60		± 1.90
		%	56.7	56.7	55.0	56.7		± 10.0	57.9	58.2	57.5	58.2		± 10.0	72.0	72.0	72.0	72.0		± 10.0
LYMPHO	#		0.73	0.72	0.79	0.72		± 0.25	2.28	2.21	2.26	2.21		± 0.65	2.58	2.54	2.69	2.54		± 1.50
		%	30.5	30.0	31.7	30.0		± 8.0	31.3	30.3	31.0	30.3		± 8.0	14.5	14.5	15.0	14.5		± 8.0
MONO	#		0.08	0.08	0.08	0.08		± 0.08	0.30	0.30	0.33	0.30		± 0.30	0.71	0.62	0.62	0.62		± 0.62
		%	3.5	3.3	3.3	3.3		± 3.3	4.0	4.0	4.5	4.0		± 4.0	4.0	3.5	3.5	3.5		± 3.5
EOS	#		0.14	0.16	0.16	0.16		± 0.14	0.27	0.29	0.27	0.29		± 0.27	0.89	0.96	0.90	0.96		± 0.83
		%	6.0	6.5	6.5	6.5		± 5.5	3.5	4.0	3.5	4.0		± 3.5	5.0	5.5	5.0	5.5		± 4.5
BASO	#		0.08	0.08	0.09	0.08		± 0.08	0.24	0.26	0.26	0.26		± 0.23	0.80	0.79	0.81	0.79		± 0.46
		%	3.3	3.5	3.5	3.5		± 3.0	3.3	3.5	3.5	3.5		± 3.0	4.5	4.5	4.5	4.5		± 2.5

**LOT** PX 073

**CONTROL**

(Exp.) 2013-09-05  
 (YYYY-MM-DD)

Rev 2

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS				120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.5	2.5	2.5		± 0.4	7.8	7.8	7.8		± 1.0	18.5	18.5	18.5		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.42	2.42	2.42		± 0.12	4.62	4.62	4.62		± 0.15	5.20	5.20	5.20		± 0.20			
HB HGB	g/dl	6.9	6.9	6.9		± 0.4	13.2	13.2	13.2		± 0.5	16.0	16.0	16.0		± 0.6			
	g/l	69	69	69		± 4	132	132	132		± 5	160	160	160		± 6			
	mmol/l	4.28	4.28	4.28		± 0.25	8.20	8.20	8.20		± 0.31	9.94	9.94	9.94		± 0.37			
HT HCT	%	19.8	19.8	19.8		± 1.5	37.4	37.4	37.4		± 2.0	45.2	45.2	45.2		± 2.5			
	l/l	0.198	0.198	0.198		± 0.015	0.374	0.374	0.374		± 0.020	0.452	0.452	0.452		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	82	82	82		± 5	81	81	81		± 5	87	87	87		± 5			
TGMH MCH	pg	28.5	28.5	28.5		± 2.0	28.6	28.6	28.6		± 2.0	30.8	30.8	30.8		± 2.5			
	fmol	1.77	1.77	1.77		± 0.12	1.77	1.77	1.77		± 0.12	1.91	1.91	1.91		± 0.16			
CCMH MCHC	g/dl	34.8	34.8	34.8		± 3.0	35.3	35.3	35.3		± 3.0	35.4	35.4	35.4		± 3.0			
	g/l	348	348	348		± 30	353	353	353		± 30	354	354	354		± 30			
	mmol/l	21.59	21.59	21.59		± 1.86	21.90	21.90	21.90		± 1.86	21.96	21.96	21.96		± 1.86			
IDR RDW	%	16.2	16.2	16.2		± 4.0	16.8	16.8	16.8		± 4.0	16.0	16.0	16.0		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	78	78		± 20	250	250	250		± 30	510	510	510		± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.4	9.4	9.4		± 2.0	9.5	9.5	9.5		± 2.0	9.5	9.5	9.5		± 2.0			
NEUT	#	1.41	1.48	1.48		± 0.30	4.33	4.67	4.67		± 0.80	13.20	13.30	13.30		± 1.90			
	%	56.3	59.0	59.0		± 10.0	55.5	59.9	59.9		± 10.0	71.5	72.1	72.1		± 10.0			
LYMPHO	#	0.76	0.69	0.69		± 0.25	2.48	2.14	2.14		± 0.65	2.72	2.60	2.60		± 1.50			
	%	30.2	27.6	27.6		± 8.0	31.8	27.5	27.5		± 8.0	14.7	14.0	14.0		± 8.0			
MONO	#	0.13	0.10	0.10		± 0.08	0.48	0.42	0.42		± 0.30	1.26	1.18	1.18		± 0.62			
	%	5.0	3.9	3.9		± 3.3	6.2	5.4	5.4		± 4.0	6.8	6.4	6.4		± 3.5			
EOS	#	0.14	0.16	0.16		± 0.14	0.27	0.31	0.31		± 0.27	0.83	0.93	0.93		± 0.83			
	%	5.5	6.5	6.5		± 5.5	3.5	4.0	4.0		± 3.5	4.5	5.0	5.0		± 4.5			
BASO	#	0.08	0.08	0.08		± 0.08	0.23	0.25	0.25		± 0.23	0.46	0.46	0.46		± 0.46			
	%	3.0	3.0	3.0		± 3.0	3.0	3.2	3.2		± 3.0	2.5	2.5	2.5		± 2.5			

Ref: TEMP-0821 Rev.33 BACK / VERSO 9930080-B