


LOT PX 055
Rev 2


CONTROL

 (Exp.) 2015-07-05
(YYYY - MM - DD)

		ABX Lysebio																				
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL					L	TOLERANCES TOLERANCE	CONTROL					N	TOLERANCES TOLERANCE	CONTROL					H	TOLERANCES TOLERANCE
		PENTRA							PENTRA							PENTRA						
		60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP			60 60C+ ES60	80 XL80	MS60	XLR	MSCRIP		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.4	2.4	2.4	2.4	± 0.4	7.4	7.5	7.5	7.5	7.4	± 1.0	17.8	17.6	18.0	17.6	17.7	± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.50	2.44	2.46	2.44	2.43	± 0.16	4.68	4.70	4.70	4.70	4.64	± 0.20	5.20	5.25	5.24	5.25	5.13	± 0.25			
HB HGB	g/dl	7.1	7.1	7.1	7.1	7.0	± 0.4	14.1	14.2	14.1	14.2	14.0	± 0.5	16.6	16.7	16.6	16.7	16.5	± 0.6			
	g/l	71	71	71	71	70	± 4	141	142	141	142	140	± 5	166	167	166	167	165	± 6			
	mmol/l	4.41	4.41	4.41	4.41	4.35	± 0.25	8.76	8.82	8.76	8.82	8.69	± 0.31	10.31	10.37	10.31	10.37	10.25	± 0.37			
HT HCT	%	20.0	20.3	19.7	20.3	19.3	± 1.5	38.4	38.5	37.6	38.5	37.6	± 2.0	45.8	46.2	45.1	46.2	44.6	± 2.5			
	l/l	0.200	0.203	0.197	0.203	0.193	± 0.015	0.384	0.385	0.376	0.385	0.376	± 0.020	0.458	0.462	0.451	0.462	0.446	± 0.025			
VGM MCV	µm ³ ; fl	80	83	80	83	79.5	± 5	82	82	80	82	81.0	± 5	88	88	86	88	87.0	± 5			
TGMH MCH	pg	28.4	29.1	28.9	29.1	28.8	± 2.0	30.1	30.2	30.0	30.2	30.2	± 2.0	31.9	31.8	31.7	31.8	32.2	± 2.5			
	fmol	1.76	1.81	1.79	1.81	1.79	± 0.12	1.87	1.88	1.86	1.88	1.87	± 0.12	1.98	1.98	1.97	1.98	2.00	± 0.16			
CCMH MCHC	g/dl	35.5	35.1	36.1	35.1	36.2	± 3.0	36.7	36.8	37.5	36.8	37.2	± 3.0	36.3	36.1	36.8	36.1	37.0	± 3.0			
	g/l	355	351	361	351	362	± 30	367	368	375	368	372	± 30	363	361	368	361	370	± 30			
	mmol/l	22.05	21.77	22.40	21.77	22.50	± 1.86	22.82	22.88	23.29	22.88	23.13	± 1.86	22.53	22.45	22.88	22.45	22.96	± 1.86			
IDR RDW	%	12.5	13.7	12.5	13.7	10.7	± 4.0	13.0	14.3	12.5	14.3	10.9	± 4.0	12.7	13.4	11.7	13.4	10.5	± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	68	68	68	70	± 20	262	258	262	258	255	± 30	485	482	495	482	480	± 50			
VPM MPV	µm ³ ; fl	9.3	9.3	9.3	9.3	8.9	± 2.0	9.0	9.2	9.0	9.2	8.5	± 2.0	8.8	9.0	8.9	9.0	8.4	± 2.0			
NEUT	#	1.31	1.42	1.44	1.42	1.52	± 0.35	4.17	4.29	4.35	4.29	4.37	± 0.90	12.37	12.27	12.78	12.27	12.83	± 1.90			
	%	57.0	59.2	60.0	59.2	63.5	± 10.0	56.3	57.2	58.0	57.2	59.0	± 10.0	69.5	69.7	71.0	69.7	72.5	± 10.0			
LYMPHO	#	0.68	0.66	0.66	0.66	0.60	± 0.33	2.46	2.44	2.44	2.44	2.29	± 0.70	2.90	2.90	2.88	2.90	2.60	± 1.50			
	%	29.5	27.5	27.5	27.5	25.0	± 12.0	33.2	32.5	32.5	32.5	31.0	± 8.0	16.3	16.5	16.0	16.5	14.7	± 8.0			
MONO	#	0.09	0.07	0.07	0.07	0.07	± 0.07	0.26	0.23	0.23	0.23	0.22	± 0.22	0.75	0.62	0.63	0.62	0.53	± 0.53			
	%	4.0	3.0	3.0	3.0	3.0	± 3.0	3.5	3.0	3.0	3.0	3.0	± 3.0	4.2	3.5	3.5	3.5	3.0	± 3.0			
EOS	#	0.14	0.16	0.16	0.16	0.12	± 0.12	0.26	0.29	0.26	0.29	0.26	± 0.26	0.98	1.02	0.99	1.02	0.97	± 0.97			
	%	6.0	6.8	6.5	6.8	5.0	± 5.0	3.5	3.8	3.5	3.8	3.5	± 3.5	5.5	5.8	5.5	5.8	5.5	± 5.5			
BASO	#	0.08	0.08	0.07	0.08	0.08	± 0.07	0.26	0.26	0.23	0.26	0.26	± 0.23	0.80	0.79	0.72	0.79	0.76	± 0.72			
	%	3.5	3.5	3.0	3.5	3.5	± 3.0	3.5	3.5	3.0	3.5	3.5	± 3.0	4.5	4.5	4.0	4.5	4.3	± 4.0			

LOT PX 055
Rev 2

CONTROL

 (Exp.) 2015-07-05
(YYYY - MM - DD)

		ABX Lysebio																			
		CONTROL				L		CONTROL				N		CONTROL				H		TOLERANCES	
PARAMETRES PARAMETERS	UNITES UNITS	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE					
		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS							
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.3	2.3	2.3		± 0.4	7.6	7.6	7.6		± 1.0	18.2	18.2	18.2		± 2.2					
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.48	2.48	2.48		± 0.16	4.70	4.70	4.70		± 0.20	5.27	5.27	5.27		± 0.25					
	g/dl	7.2	7.2	7.2		± 0.4	14.1	14.1	14.1		± 0.5	16.5	16.5	16.5		± 0.6					
HB HGB	g/l	72	72	72		± 4	141	141	141		± 5	165	165	165		± 6					
	mmol/l	4.47	4.47	4.47		± 0.25	8.76	8.76	8.76		± 0.31	10.25	10.25	10.25		± 0.37					
HT HCT	%	20.6	20.6	20.6		± 1.5	39.0	39.0	39.0		± 2.0	46.9	46.9	46.9		± 2.5					
	l/l	0.206	0.206	0.206		± 0.015	0.390	0.390	0.390		± 0.020	0.469	0.469	0.469		± 0.025					
VGM MCV	µm ³ ·fl	83	83	83		± 5	83	83	83		± 5	89	89	89		± 5					
TGMH MCH	pg	29.0	29.0	29.0		± 2.0	30.0	30.0	30.0		± 2.0	31.3	31.3	31.3		± 2.5					
	fmol	1.80	1.80	1.80		± 0.12	1.86	1.86	1.86		± 0.12	1.94	1.94	1.94		± 0.16					
	g/dl	35.0	35.0	35.0		± 3.0	36.1	36.1	36.1		± 3.0	35.2	35.2	35.2		± 3.0					
CCMH MCHC	g/l	350	350	350		± 30	361	361	361		± 30	352	352	352		± 30					
	mmol/l	21.72	21.72	21.72		± 1.86	22.45	22.45	22.45		± 1.86	21.85	21.85	21.85		± 1.86					
IDR RDW	%	14.5	14.5	14.5		± 4.0	16.0	16.0	16.0		± 4.0	14.3	14.3	14.3		± 4.0					
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	72	72		± 20	265	265	265		± 30	495	495	495		± 50					
VPM MPV	µm ³ ·fl	9.3	9.3	9.3		± 2.0	8.9	8.9	8.9		± 2.0	8.7	8.7	8.7		± 2.0					
NEUT	#	1.37	1.42	1.42		± 0.35	4.31	4.46	4.46		± 0.90	12.70	13.00	13.00		± 1.90					
	%	59.6	61.8	61.8		± 10.0	56.7	58.7	58.7		± 10.0	69.5	71.2	71.2		± 10.0					
LYMPHO	#	0.62	0.56	0.56		± 0.33	2.48	2.30	2.30		± 0.70	3.19	2.88	2.88		± 1.50					
	%	27.1	24.4	24.4		± 12.0	32.6	30.2	30.2		± 8.0	17.5	15.8	15.8		± 8.0					
MONO	#	0.11	0.12	0.12		± 0.11	0.40	0.42	0.42		± 0.40	1.20	1.20	1.20		± 1.20					
	%	4.9	5.0	5.0		± 4.9	5.3	5.5	5.5		± 5.3	6.6	6.6	6.6		± 6.6					
EOS	#	0.14	0.14	0.14		± 0.14	0.18	0.20	0.20		± 0.18	0.71	0.71	0.71		± 0.71					
	%	5.9	6.3	6.3		± 5.9	2.4	2.6	2.6		± 2.4	3.9	3.9	3.9		± 3.9					
BASO	#	0.06	0.06	0.06		± 0.06	0.23	0.23	0.23		± 0.23	0.46	0.46	0.46		± 0.46					
	%	2.5	2.5	2.5		± 2.5	3.0	3.0	3.0		± 3.0	2.5	2.5	2.5		± 2.5					

9930096-A FRONT / RECTO Ref: TEMP-0821 Rev.36