


LOT PX 420

CONTROL

 (Exp.) 2020-01-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE
			CONTROL			L	CONTROL			N	CONTROL			H				
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE				
			120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS		120 120 RETIC	DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.8	2.8	2.8	± 0.4	8.1	8.1	8.1	± 1.0	20.2	20.2	20.2	± 2.2					
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.48	2.48	2.48	± 0.16	4.50	4.50	4.50	± 0.20	5.19	5.19	5.19	± 0.25					
HB HGB	g/dl	7.3	7.3	7.3	± 0.4	13.4	13.4	13.4	± 0.5	16.0	16.0	16.0	± 0.6					
	g/l	73	73	73	± 4	134	134	134	± 5	160	160	160	± 6					
	mmol/l	4.53	4.53	4.53	± 0.25	8.32	8.32	8.32	± 0.31	9.94	9.94	9.94	± 0.37					
HT HCT	%	21.1	21.1	21.1	± 1.5	37.4	37.4	37.4	± 2.0	44.6	44.6	44.6	± 2.5					
	l/l	0.211	0.211	0.211	± 0.015	0.374	0.374	0.374	± 0.020	0.446	0.446	0.446	± 0.025					
VGM MCV	µm ³ ; fl	85	85	85	± 5	83	83	83	± 5	86	86	86	± 5					
TGMH MCH	pg	29.4	29.4	29.4	± 2.0	29.8	29.8	29.8	± 2.0	30.8	30.8	30.8	± 2.5					
	fmol	1.83	1.83	1.83	± 0.12	1.85	1.85	1.85	± 0.12	1.91	1.91	1.91	± 0.16					
CCMH MCHC	g/dl	34.6	34.6	34.6	± 3.0	35.9	35.9	35.9	± 3.0	35.8	35.8	35.8	± 3.0					
	g/l	346	346	346	± 30	359	359	359	± 30	358	358	358	± 30					
	mmol/l	21.51	21.51	21.51	± 1.86	22.28	22.28	22.28	± 1.86	22.26	22.26	22.26	± 1.86					
IDR RDW	%	14.5	14.5	14.5	± 4.0	16.0	16.0	16.0	± 4.0	15.0	15.0	15.0	± 4.0					
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70	70	± 20	250	250	250	± 30	515	515	515	± 50					
VPM MPV	µm ³ ; fl	9.3	9.3	9.3	± 2.0	9.2	9.2	9.2	± 2.0	10.3	10.3	10.3	± 2.0					
NEUT	#	1.52	1.57	1.57	± 0.35	4.50	4.63	4.63	± 0.90	13.90	14.10	14.10	± 1.90					
	%	54.2	56.1	56.1	± 10.0	55.6	57.1	57.1	± 10.0	68.7	69.9	69.9	± 10.0					
LYMPHO	#	0.79	0.72	0.72	± 0.33	2.46	2.30	2.30	± 0.70	3.17	2.93	2.93	± 1.50					
	%	28.2	25.7	25.7	± 12.0	30.4	28.4	28.4	± 8.0	15.7	14.5	14.5	± 8.0					
MONO	#	0.20	0.21	0.21	± 0.20	0.52	0.53	0.53	± 0.52	1.33	1.33	1.33	± 1.33					
	%	7.3	7.4	7.4	± 7.3	6.4	6.6	6.6	± 6.4	6.6	6.6	6.6	± 6.6					
EOS	#	0.22	0.23	0.23	± 0.22	0.33	0.36	0.36	± 0.33	1.21	1.21	1.21	± 1.21					
	%	7.8	8.3	8.3	± 7.8	4.1	4.5	4.5	± 4.1	6.0	6.0	6.0	± 6.0					
BASO	#	0.07	0.07	0.07	± 0.07	0.28	0.28	0.28	± 0.28	0.61	0.61	0.61	± 0.61					
	%	2.5	2.5	2.5	± 2.5	3.5	3.4	3.4	± 3.4	3.0	3.0	3.0	± 3.0					

Ref: TEMP-0821 Rev.43 BACK / VERSO 1300040596